A machine learning approach to predict instrument bending in stereotactic neurosurgery

Alejandro Granados¹, Matteo Mancini¹, Sjoerd B. Vos^{1,3}, Oeslle Lucena⁶,
Vejay Vakharia^{1,2,3}, Roman Rodionov^{2,3}, Anna Miserocchi²,
Andrew W. McEvoy², John S. Duncan^{2,3}, Rachel Sparks¹, and
Sébastien Ourselin^{1,4,5}

¹ Wellcome/EPSRC Centre for Interventional & Surgical Sciences, UCL, UK

² National Hospital of Neurology and Neurosurgery, London, UK

³ Department of Clinical and Experimental Epilepsy, Institute of Neurology,
National Hospital for Neurology & Neurosurgery, UK

⁴ Dementia Research Centre, UCL Inst Neurology, London, UK

⁵ School of Biomedical Engineering and Imaging Sciences,
St Thomas Hospital, Kings College London, London, United Kingdom

⁶ Medical Imaging Computing Lab, State University of Campinas, Brazil
`{alejandro.granados}@ucl.ac.uk`

Abstract. The accurate implantation of stereo-electroencephalography (SEEG) electrodes is crucial for localising the seizure onset zone in patients with refractory epilepsy. Electrode placement may differ from planning due to instrument deflection during surgical insertion. We present a regression-based model to predict instrument bending using image features extracted from structural and diffusion images. We compare three machine learning approaches: *Random Forest*, *Feed-Forward Neural Network* and *Long Short-Term Memory* on accuracy in predicting global instrument bending in the context of SEEG implantation. We segment electrodes from post-implantation CT scans and interpolate position at 1 mm intervals along the trajectory. Electrodes are modelled as elastic rods to quantify 3 degree-of-freedom (DOF) bending using Darboux vectors. We train our models to predict instrument bending from image features. We then iteratively infer instrument positions from the predicted bending. In 32 SEEG post-implantation cases we were able to predict trajectory position with a MAE of 0.49 mm using RF. Comparatively a FFNN had MAE of 0.71 mm and LSTM had a MAE of 0.93 mm.

Keywords: machine learning, instrument bending, neurosurgery, trajectory prediction, surgical planning

1 Introduction

Minimally-invasive surgical interventions use thin, tubular, and flexible instruments inserted through the skull to target small regions of interest with the aim of acquiring data (e.g. electroencephalography, biopsy) or delivering therapy (e.g. injection, stimulation, ablation). The accurate placement of these instruments

is important for patient safety, accurate diagnosis, and treatment efficacy [1, 11, 22]. Preoperative trajectory planning aims to minimise target point errors whilst avoiding critical structures [9, 18]. However, accurate surgical implantation may be difficult to achieve as instruments may deflect during insertion. Deflection may be caused by instrument design (mechanical properties, tip shape), tissue properties (stiffness, inhomogeneity, anisotropy), insertion forces (depth, velocity, steering) and physiological processes [11, 17]. The prediction of instrument trajectories inserted into deformable tissue is a challenging problem and an active area of research [5, 9, 17].

Needle-tissue interaction models. Modelling of trajectory deflection is mostly based on mechanical modelling of tissue-needle interaction. These methods require accurate extraction of the mechanical properties of tissue and instrument together with insertion measurements (e.g. forces, velocities) to accurately predict instrument deflection. Typically such methods use approximations or simplifications derived from hand-crafted modelling techniques to account for unknown parameters. Different instrument models have been proposed including cantilever beams [17], Finite Element Methods (FEM) [5, 9], Timoshenko formulation, and Cosserat rods [19, 12]. A mechanical model proposed by [17] predicts needle deflection during needle-tissue interaction by modelling forces at the tip (dependent on tissue stiffness) and friction forces along the needle shaft. A simulation-based approach for haptic radiofrequency ablation proposed by [9] iteratively computes instrument deflection during insertion using a FEM.

Machine learning models. Several approaches use data-driven machine learning models to predict trajectories (e.g. vehicles, rigged skeleton models) based on spatio-temporal data. These model-free approaches require large amounts of data that must be general enough to predict unseen cases. While, the application of machine learning to predict instrument bending has to the best of our knowledge not yet been presented, these approaches represent a promising avenue to learn features of bending from real world data (i.e. previous examples of instrument trajectories and medical images). Machine learning techniques have long been applied to prediction of spatio-temporal data, most recently with Deep Learning techniques [7]. Recurrent Neural Networks (RNNs), in particular Long Short-Term Memory (LSTM) and Gated Recurrent Units (GRUs), have been applied to predict trajectories by taking into account data at previous timesteps and learning in sequences. An example of this is the prediction of the direction of white matter fibre streamlines from diffusion weighted imaging (DWI) [15].

Our contributions. We present a novel machine learning regression approach to predict global instrument bending from image features. Instruments are modelled as elastic rods with 3DOF bending and twisting. This method is applied to predicted intra-cranial electrode bending using features extracted from structural T1-weighted (T1-w) and diffusion MRI (dMRI).

Fig. 1. *Top:* Imaging modalities: a) post-implantation CT scan, b) MRI, c) probabilities from segmentation, d) parcellation, e) dMRI. *Bottom:* f) SEEG electrode segmentation and interaction of electrodes with surface models including: g) scalp, h) cortex (in translucent pink), white matter (in white) and i) deep grey matter (in blue).

2 Methods

Image acquisition and preprocessing. Prior to electrode implantation, MRI data was acquired either on a 3T GE Signa HDx, consisting of a 3D-T1-w and single-shell dMRI scan (as in [20]), or on a 3T GE MR750, using a T1-w MPRAGE and a multi-shell dMRI scan (as in [13]). dMRI was corrected for susceptibility-induced and eddy-current induced distortions using FSLs topup and eddy, respectively. The T1-w MRI was used to compute a brain parcellation and segmentation probabilities using geodesic information flows (GIF) [2] (Fig. 1). Smoothed 3D polygon meshes of the scalp [4] and superficial grey, white, and deep grey matter were generated from the parcellation. dMRI was modelled as a multi-tissue constrained-spherical deconvolution (MT-CSD) [10] with the Dhollander algorithm [3] implemented in MRtrix3 [21]. We characterise fibre direction and density using the ‘fixel’ framework [16]. Rigid registration of the CT image to the T1-w MRI was performed by minimising normalised mutual information (NMI) [14]. Similarly, dMRI images were registered by minimising the NMI between the T1-w MRI and b0 dMRI scan. A CT image was acquired immediately after electrode implantation.

Instrument segmentation and interpolation. Automatic segmentation of SEEG electrode contacts was performed [8] (Fig. 1). We interpolated instrument position from the electrode contact positions at 1 mm intervals using a shape-preserving piecewise cubic Hermite interpolation (PCHIP) [6]. For each position, the following set of features were calculated (Fig. 2, top).

- **Structural:** voxel probability of being background, cerebral spin fluid, grey matter, white matter, deep grey matter and brain stem; and the brain region.

Fig. 2. *Top:* Interpolation points (red circles) along trajectory with structural, white matter fibre tracks-related and trajectory-related features. *Bottom:* Elastic rod model where interpolated points are particles. Material frames and their rate of change (Darboux vectors) are computed to characterise local/global bending.

- **White matter fibre tracks:** number and direction of crossing fibres, fibre density and angle of the fibre with respect to the trajectory.
- **Trajectory:** number of regions traversed, length of region at position, angle with respect to surfaces scalp, superficial grey, white, and deep grey matter, distance to grey matter, distance to cortex, distance to previous and next region, and distance from scalp to cortex and from cortex to white matter.

Instrument bending model. Instruments were modelled as Cosserat elastic rods in the Position-based Dynamics library. Particles were placed at the interpolated instrument position. Ghosts particles were placed orthogonally half-way between each particle. A material frame is computed between particles with a unit vector aligned tangentially to the direction of the rod and two orthonormal vectors. The rate of change between two consecutive material frames was computed as a Darboux vector [12] (Fig. 2, bottom), where the x , y and z components indicate lateral bending, upward/downward bending and twisting, respectively. For each particle two Darboux vectors were computed: a) *local bending* (between material frames of neighbouring particles) and *global bending* (between the material frame corresponding to the first particle and the current particle).

Training model. An Ordinary Least Squares multiple regression was executed (Statsmodels library) to determine which features have a statistically significant effect on bending. Three regression models were built and evaluated: a) Random Forest (RF), b) Feed-Forward Neural Network (FFNN) and c) LSTM (see Table 1). Global bending at the electrode tip was used as the predicted output of the regression model. Local bending and previous global bending are used as input features.

Table 1. Training model architecture

Model	ML Library	Topology	Details	Activation	Noise
RF	scikit-learn	trees: 200 max. tree depth: 50			No
FFNN	Keras with Tensorflow as a backend	layers: 3 Multiple Layer Perceptron hidden units: equal to number of features	- L2 normalisation (hyperparameter of $1e-4$) - normal function to initialise weights	- hidden layer: scaled exponential linear unit (SELU) - output layer: linear	Gaussian ($\mu = 0, \sigma = 0.0025$)
LSTM		layers: - Input layer - 3 LSTM layers (200, 100, 50 units) - Dropout (0.2) between LSTM layers - Dense	sequences of data (window length of 5 interpolated points)	- output layer: linear	No

3 Experimental design and validation.

We trained our models from 32 post-implantation SEEG cases (296 electrodes, 13107 interpolation points) comprising two different surgical approaches, placing a guiding stylet close to or far from target point. Through multiple linear regression, we found a significant effect on instrument bending of the features listed in Table 2 ($R^2 = 0.313$). The data from these features (27 in total) was normalised. We created dichotomous variables for the treating categorical variables.

Table 2. Features with significant effect ($p < 0.05$) on instrument bending

Positively correlated	Coefficient	Negatively correlated	Coefficient
displacement w.r.t. planned trajectory	0.2753	x-component of previous bending	-0.0533
z-component of previous bending	0.1315	y-component of previous bending	-0.0472
angle w.r.t. white matter	0.0379	number of fibre tracks	-0.0157
distance to previous region	0.0346	length of the current traversed region	-0.0089
distance of cortex traversed	0.0343	using stylet	-0.0048
probabilities of segmentation of deep grey matter	0.0328		
probabilities of segmentation of white matter	0.0320		
probabilities of segmentation of grey matter	0.0279		
MRI voxel intensity	0.0240		
angle w.r.t. cortex	0.0215		
closest distance to grey matter	0.0198		
regions traversed	0.0135		
target point region categories (with the exception of central and insula)			

A 5-fold cross validation was performed for each model (i.e. RF, FFNN and LSTM). For each fold, we split the data into three sets. A *validation set* contains a trajectory randomly selected from 15 SEEG cases (i.e. 15 trajectories in total). The remaining data is split into a *train set* containing 80% and a *test set* containing 20%. We then conducted two experiments using these three sets: (a) model validation and (b) trajectory accuracy. In experiment (a), we trained our regressor models using the train set and evaluated the inference scores against the test set. The models were trained and average prediction scores across folds are reported in Table 3.

Table 3. Performance of bending prediction of a) RF, FFNN and LSTM models

Model	K-Fold μ (σ)		
	MSE	MAE	R^2
RF	0.0004 (0.0001)	0.0077 (0.0007)	0.8533 (0.0260)
FFNN	0.0004 (0.0001)	0.0079 (0.0007)	0.8540 (0.0230)
LSTM	0.0007 (0.0002)	0.0126 (0.0015)	0.6583 (0.1043)

In experiment (b), we selected 30 trajectories from the validation set across folds (15,4,4,4,3 respectively). Using the trained models of the respective folds, the insertion of an electrode was simulated iteratively starting with the first six interpolated points (sequence length for LSTM) at 1 mm intervals from the electrode bolt. Details of the simulation loop can be found in Algorithm 1.

Algorithm 1 Simulation of electrode implantation based on predicted bending

```

1: Load images and 3D meshes
2: do
3:   Generate trajectory features: The Medical Imaging Interaction Toolkit (MITK)
     The Visualisation Toolkit (VTK)
4:     - load position of particles along trajectory
5:     - generate features from imaging along inserted trajectory and save
6:   Predict bending: Keras Python deep learning library / Tensorflow ML framework
7:     - load image features along trajectory of electrode
8:     - normalise and categorise features
9:     - generate sequences (only for LSTM)
10:    - load trained model for the corresponding fold
11:    - predict global bending (Darboux vector)
12:  Compute position of electrode tip: Position-based dynamics library
13:    - create an elastic rod consisting of three particles (and two ghost particles)
      in the direction of the trajectory
14:    - fix two of the particles (and one ghost particle)
15:    - apply constraints (distance, perpendicular bisector, ghost point
      edge distance, and Darboux vector)
16:    - set predicted bending as a resting Darboux vector
17:    - compute position of particle (tip of electrode) and append to trajectory
18: while current insertion depth < planned insertion depth

```

We then evaluated the accuracy of the trajectory bending prediction (Fig. 3). Distances between true and predicted trajectories were measured. RF had the lowest MAE ($\mu = 0.49, \sigma = 0.43$), followed by FFNN ($\mu = 0.71, \sigma = 0.41$) and by LSTM ($\mu = 0.93, \sigma = 0.65$). Under the current training paradigm, we observed that LSTMs tend to accumulate error (given the current accuracy of the 5-fold cross validation in Table 3). Further analysis and method development is required to keep predicted electrode bending within realistic values.

4 Conclusion

We present a novel machine learning regression model to predict local instrument bending from image features. The instrument is modelled as an elastic rod, with 3DOF, that traverses soft tissue, characterised as inhomogeneous and

Fig. 3. Prediction accuracy. *Top:* Mean average error (MAE) of predicted trajectories (RF, FFNN, LSTM) with respect to True trajectory. X-axis indicates the MAE of True trajectory with respect to Plan (rigid) trajectory. Size of circle indicates standard deviation (see two examples for reference). *Bottom:* 3D rendering of planned (in white), true (in yellow) and predicted (RF in red, FFNN in green, LSTM in blue) electrode trajectories of two examples highlighted in scatter plot above.

anisotropic. Our method allows for the prediction of orientation and position along the instrument trajectory. The main limitation of our current work is that we study instrument bending in a specific application (SEEG electrode implantation). In future work, we will validate our method in other applications. Additionally, we will integrate our model into a trajectory planning algorithm with the aim of improved pre-operative planning.

Acknowledgements. This publication represents in part independent research commissioned by the Health Innovation Challenge Fund (WT106882), the Wellcome/EPSRC [203145Z/16/Z], and the National Institute for Health Research University College London Hospitals Biomedical Research Centre (NIHR BRC UCLH/UCL High Impact Initiative). We are grateful to the Wolfson Foundation and the Epilepsy Society for supporting the Epilepsy Society MRI scanner. The views expressed in this publication are those of the authors and not necessarily those of the Wellcome Trust or NIHR. Oeslle Lucena thanks FAPESP (2016/18332-8) and FAPESP (2017/23747-5).

References

1. Abolhassani, N., Patel, R., Mehrdad Moallem. Needle insertion into soft tissue: A survey. *Medical Engineering & Physics* 29(4) , 413-431 (2007)
2. Cardoso, M. J., Modat, M., Wolz, R., et al.: Geodesic Information Flows: Spatially-Variant Graphs and Their Application to Segmentation and Fusion. *IEEE Trans. Med. Imag.* 34(9), 1976-1988 (2015)

3. Dhollander, T., Raffelt, D., Connelly, A.: Unsupervised 3-tissue response function estimation from single-shell or multi-shell diffusion MR data without a co-registered T1 image. *ISMRM Workshop on Breaking the Barriers of Diffusion MRI*, 5 (2016)
4. Dogdas, B., Shattuck, D. W., Leahy, R. M.: Segmentation of skull and scalp in 3-D human MRI using mathematical morphology. *Human Brain Mapping*, 26(4), 273–285 (2005)
5. Duriez, C., Guebert, C., Marchal, M., Cotin, S., Grisoni, L.: Interactive simulation of flexible needle insertions based on constraint models. *MICCAI* (2009)
6. Fritsch, F. N., Carlson, R. E.: Monotone Piecewise Cubic Interpolation. *SIAM Journal on Numerical Analysis*, 17(2), 238–246 (1980)
7. Gamboa, J.: Deep Learning for Time-Series Analysis. *arXiv preprint. arXiv:1701.01887* (2017)
8. Granados, A., Vakharia, V., Rodionov, R. et al. *Int J CARS* (2018) 13: 935. <https://doi.org/10.1007/s11548-018-1740-8>
9. Hamze, N., Peterlik, I., Cotin, S., Essert, C.: Preoperative Trajectory Planning for Percutaneous Procedures in Deformable Environments. *Computerized Medical Imaging and Graphics*. 1–15 (2015)
10. Jeurissen, B., Tournier, J. D., Dhollander, T., Connelly, A., Sijbers, J.: Multi-tissue constrained spherical deconvolution for improved analysis of multi-shell diffusion MRI data. *NeuroImage*, 103, 411–426 (2014)
11. de Jong, T.: Needle Deflection in Tissue. Master’s thesis. Delft University of Technology (2015)
12. Kugelstadt, T., Schömer, E.: Position and Orientation Based Cosserat Rods. *Eurographics ACM SIGGRAPH Symposium on Computer Animation*. 1–10 (2016)
13. Mancini, M., Vos, S.B., Vakharia, V., Sparks, R., Trimmell, K., Winstom, G., Duncan, J., Ourselin, S.: Anatomy-constrained automated fibre tract reconstruction for surgery planning: a validation study in language-related white matter tracts. *Proc. Intl. Soc. Mag. Reson. Med.* 26: 075 (2018)
14. Modat M., Cash D.M., Daga P., et al. Global image registration using a symmetric block-matching approach. *J. of Med. Imag.* 2014;1(2):024003.
15. Poulin, P., Cote, M-A., Houde, J-C., et al.: Learn to Track: Deep Learning for Tractography. *MICCAI*, 540-547 (2017)
16. Raffelt, D., Smith, R., Ridgway, G. R., et al.: Connectivity-based fixel enhancement: whole-brain statistical analysis of diffusion MRI measures in the presence of crossing fibres. *NeuroImage*, 117, 40–55 (2015)
17. Roesthuis, R.J., van Veen, Y. R. J., Jahya A., Misra, S.: Mechanics of Needle-Tissue Interaction. *IEEE Intl. Conf. Int. Robots and Systems*, pp. 2557–63 (2011)
18. Sparks, R., Vakharia, V. N., Rodionov, R., et al.: Anatomy-driven multiple trajectory planning (ADMTP) of intracranial electrodes for epilepsy surgery. *Intl. J. Comp. Assisted Radiology and Surgery*. 1–11 (2017)
19. Spillmann, J., Harders, M: Inextensible elastic rods with torsional friction based on Lagrange multipliers. *Comp. Anim. And Virtual Worlds*, 19, 271–281 (2010)
20. Taylor, P. N., Sinha, N., Wang, Y., et al.: The impact of epilepsy surgery on the structural connectome and its relation to outcome. *NeuroImage*, 18, 202-14 (2018)
21. Tournier, J. D., Calamante, F., Connelly, A.: MRtrix: diffusion tractography in crossing fiber regions. *Intl. J. of Imag. Systems and Tech.* 22, 53–66 (2012)
22. Vakharia, V. N., Sparks, R., O’Keeffe, A. G., et al.: Accuracy of intracranial electrode placement for stereoencephalography: A systematic review and meta-analysis. *Epilepsia*, 58(6), 921–932 (2017)