

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR
IV MEDICATION SAFETY

**DEVELOPING STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART
PUMPS IN ADVANCED HEALTHCARE SYSTEMS TO IMPROVE INTRAVENOUS
MEDICATION SAFETY**

Adam Sutherland

Division of Pharmacy & Optometry, Faculty of Biology, Medicine & Health, University
of Manchester, Oxford Road, MANCHESTER M13 9PL, UK

Adam.sutherland@manchester.ac.uk

ORCID: 0000-0002-5198-3727

Dr Matthew D Jones

Department of Pharmacy & Pharmacology, University of Bath, Bath, UK

ORCID: 0000-0002-2617-4098

Dr Moninne Howlett

Children's Health Ireland, Dublin, Ireland

ORCID: 0000-0002-5798-6240

Dr Sara Arenas-Lopez

Evelina London Children's Hospital, Guys & St Thomas's Hospitals NHS Foundation
Trust, London, UK ORCID: 0000-0003-2231-1339

Mr Arif Patel

Trust HQ, East Lancashire Hospitals NHS Trust, Blackburn, UK

Professor Bryony Dean Franklin

Centre for Medication Safety and Service Quality, Imperial College Healthcare NHS
Trust, London, UK / UCL School of Pharmacy, London, UK

Word Count: 3450

ACKNOWLEDGEMENTS

We would like to thank the following for their contribution to and participation in the
symposium

- Dr Paul Dean
Chair, Standards Division, Intensive Care Society
- Paul T. Lee. BA (Hons), MIPEM, RSci
Medical Devices Training Manager
Swansea Bay University Health Board and Chair NAMDET (www.namdet.org)
- Mr Patrick O'Sullivan
Medication Safety Officer
Imperial College Healthcare NHS Trust
- Susan Keeling
Medusa Injectable Medicines Website Co-Ordinator
- Catherine Maddock

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

Acute & Chronic Pain Specialist Nurse Practitioner and Non Medical Prescribing Lead

Hampshire Hospitals NHS Foundation Trust

- Ann Blandford PhD CEng,
Professor of Human–Computer Interaction, University College London
- Dr Virginia Aguado Lorenzo
Safety Injectable Medicines Pharmacist
Guys & St Thomas’s Hospitals NHS Foundation Trust, London
- Dr Malcolm Phillips
Head of Clinical Engineering.
Department of Medical Physics, Royal Infirmary of Edinburgh, NHS Lothian
- Steve Tomlin
Chief Pharmacist, Great Ormond Street Hospital
- Dr Anna Maria Rollin, MBE
Patients Association (UK)
- Dr David Cousins
Safe Medication Practice Consultant
Action against Medical Accidents (AvMA)
- Nanna Christiansen
Associate Chief Pharmacist
Evelina London Children’s Hospital
- Fraser Hanks
Principal Pharmacist, Critical Care
Guys & St Thomas’s Hospital NHS Foundation Trust
- Tina Worth LLM, MBA, PGCE, RGN
Manager, Medical Affairs (Medication Management Solutions)
BD

DECLARATIONS

Author Contributions

All authors contributed to the conception of the strategy. The first draft of the manuscript was written by Adam Sutherland and all authors edited and commented on previous versions. All authors have read and approved the final manuscript.

Funding

The symposium was supported by Becton-Dickinson (UK) Ltd., through the provision of video conferencing facilities and co-ordination of delegates

Competing Interests

AS has received research funding and honoraria from BD (UK) Ltd to study the effectiveness of infusion devices. He is also supported by the National Institute of Health Research (NIHR) Integrated Clinical Academic Programme (ICA-CDRF-2017-

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

03-024). The views expressed are those of the authors and not necessarily those of the NIHR or the Department of Health and Social Care.

MDJ was funded by a National Institute for Health Research (NIHR) Translational Research Fellowship (TRF-2017-10-006). The views expressed are those of the authors and not necessarily those of the NIHR or the Department of Health and Social Care. He has also received honoraria for conference speaking engagements from BBraun Medical Ltd.

SAL has no conflicts of interest to declare

MH has no conflicts of interest to declare

AP has received honoraria for speaking engagements and medical device research from BD(UK) Ltd.

BDF is involved with a separate project that has received funding from BD(UK) Ltd. She is also supported by the National Institute for Health Research (NIHR) Imperial Patient Safety Translational Research Centre. The views expressed are those of the authors and not necessarily those of the NIHR or the Department of Health and Social Care

Ethics Approval Not applicable

Consent to Participate Not applicable

Consent for Publication Not applicable

Availability of data and material Not applicable

Code Availability Not applicable

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

ABSTRACT

Avoidable harm associated with medication is a persistent problem in health systems and the use of pre-programmed infusion devices (“smart pumps”) and data monitoring is seen as a core approach to mitigating and reducing the incidence of these harms. However, smart pumps are costly to procure, configure and maintain (in both human and financial terms) and are often poorly implemented. Variation in the manner in which medicines are prepared and used within complex modern healthcare systems exacerbates these challenges, and a strategic human-centred approach is needed to support their implementation.

A symposium of 36 clinical and academic medication safety experts met virtually to discuss the current “state of the art” and to propose strategic recommendations to support the implementation of medication administration technology to improve medication safety. The recommendations were that health systems: 1) standardise infusion concentrations to facilitate the development of ready-to-administer formulations of frequently used medicines, and support ‘out of the box’ programming of infusion devices; 2) Develop and implement drug libraries using human-centred approaches and the aforementioned standard concentrations, with a theoretical understanding of how devices are used in practice; 3) Develop standardised metrics and outcomes to support the interpretation of data produced by infusion devices; 4) Involve all stakeholders in the development of drug libraries and metrics to ensure broad understanding of the devices, their benefits and limitations; 5) Leverage input into device design, working with manufacturers and users.

Using this strategic approach, it is then possible to envisage and plan real-world implementation studies using a uniform approach to quantify improvements in safety, efficiency and cost effectiveness.

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

KEY POINTS

- 1** Avoidable patient harm associated with medicines is common, and while the harm related to intravenous (IV) medication is unclear, there is a perception that a large proportion of IV medication administrations are associated with an error and the consequences of IV medication errors are much greater than with other medicines.
- 2** The use of technology has been advocated as a key way to improve medication safety and service efficiency, but technology is often implemented poorly, and decision makers and policy makers have a poor understanding of how IV medication is administered in real life. The time to correct this is now, at a time when multiple recommendations are being announced directing healthcare organisations to “change” their approach to medication safety.
- 3** A collaborative systems-focused approach to IV medication safety including national standardisation of infusion practice, centralised smart pump libraries and data management, and end-user involvement in device development are all recommended to foster a system-wide approach to IV medication safety.

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

1 INTRODUCTION Medication safety has been a focus of work for academics and clinicians for at least twenty years, leading to heightened awareness of the causes of adverse medication events, and the development of many interventions. However, there is limited evidence for a corresponding improvement in actual harm medication events. Globally, it is estimated that 25% of avoidable medical harm events are associated with medication errors.[1] More specifically, there is also growing concern that a large proportion of intravenous (IV) medication administrations are associated with an error,[2] although the harm associated with these is unclear.[3] Additionally, in the context of the COVID-19 pandemic, patient safety and supporting healthcare professionals in the provision of safe care have never been higher on the global healthcare agenda.[4]

The World Health Organization established its third global patient safety challenge in 2017 with a call for the reduction in harmful medication errors by 50% within five years.[5] Following this, the NHS England Patient Safety Strategy advocated for the swift adoption of electronic prescribing and medication administration systems across the National Health Service (NHS) at pace as an intervention to improve quality and safety of care. Medication administration systems include (but are not limited to) computerised physician order entry (CPOE), barcode medication administration (BCMA), automated dispensing cabinets and “Smart Pumps.” It has been suggested that by using these systems together, closed loop medication administration processes can lead to reduced medication related harm by cross checking patients, medication and prescription using digital methods.[6]

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

2 CHALLENGES AND CONSIDERATIONS

Four high-level policy directives emerged in the United Kingdom (UK) in 2020, centred on IV medication safety (Table 1) and the interventions that UK health services should adopt. While these were specific to the UK NHS, the principles upon which they were based can be generalised into broader strategies that could be adopted by other developed healthcare services.

Author(s)	Core themes	Recommendations
National Association of Medical Device Educators and Trainers (NAMDET) [7]	The use of Smart Pumps in NHS hospitals should be standard of care	<ul style="list-style-type: none"> • Standardised smart pump data sets • Sharing of event data at a national level
Health Safety Investigation Board (HSIB) [8]	Three non-fatal ten-fold overdoses of fentanyl caused by non-standard practice across three sites in a single organization highlights the need for operational standardisation	<ul style="list-style-type: none"> • A nationally standardised drug library for smart pumps. • The sharing of event log data allowing independent third-party review and system-wide learning.
NHS England and Department of Health and Social Care Aseptic Services Redesign[9] and regional equivalents.[10]	To increase manufacturing capacity in the NHS to: free up nursing capacity for care; support work already delivered in standardisation of injectable medicines. [11–14]	<ul style="list-style-type: none"> • Standardise IV products where possible

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

Global Digital Exemplar (GDE) programme.[15]	Drive to implement sophisticated electronic medicines administration systems in digitally mature organisations	<ul style="list-style-type: none"> • Targeted funding for enhancements and provision of closed loop medication administration systems (e.g. barcode medication verification)
--	--	---

Table 1- UK Policy Directives on Injectable Medication Safety, 2020

Smart pumps are central to the digital safety revolution because of their capacity to cross check and prevent inadvertent dose or concentration deviations. Smart pumps are infusion devices pre-programmed with parameters for drugs, concentrations and dosing limits (referred to as a “drug library”) that are defined locally. They can alert users to programming outside these parameters and potentially prevent adverse medication events. However, the causes and contributory factors to adverse events associated with IV medications are many and multifaceted.[16] There is a complex interplay between devices, the customs, practices and experiences of the people who administer them, and the competing priorities of patients, organisations and other professionals that influence how these medicines are administered.[3,17–19]

None of the documents in Table 1 offer operational guidance, and are written with a generalised focus on the prevailing issues. Thus, there is a need to provide material guidance on how these devices should be introduced and managed, to ensure that the systems within which healthcare staff work are fit for purpose, and developed with an appropriate understanding of how work is delivered in practice while supporting safety across the entire spectrum of care.

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

Standardising medication practices at a systems level requires considerable collaboration between policy makers, organisational managers, practitioners and researchers to develop and implement operational interventions. The increasing reliance on technology without adequate sociotechnical assessment presents a risk of wasted resource due to the implementation of ineffective systems. There is a need for clear strategies to support the NHS (and similar centralised healthcare systems) in implementing recommendations made by policy makersto deliver improvements in medication safety. These strategies need to ensure that the environment and technology in which healthcare practitioners work reflects the work they need to do. This can only be achieved though collaboration with clinical practitioners, patient safety specialists and human factors and ergonomics practitioners.

3 THE EXPERT SYMPOSIUM

In November 2020, 36 academics, clinicians and members of two prominent UK patient advocate groups (the Patients' Association and the Association against Medical Accidents) came together to discuss the current state of IV medication management in the UK and Ireland. The symposium examined practical, real-world experience and knowledge of interventions to mitigate IV medication errors from both the UK and Ireland, and considered how to translate and scale up these interventions for system-wide use. Further, the group considered the development of research partnerships for future study of IV medication safety. The intended

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

outcomes of the symposium were to propose generalisable strategies for IV medication safety and to outline a research strategy to accompany this work.

3.1 Symposium Origins and Organisation

Two systematic reviews published in the last ten years have highlighted both the potential burden of IV medication associated harm [20] and the challenges in implementing technology to mitigate them (specifically smart pumps.)[21] In January of 2019, Becton Dickinson (UK) Limited brought together a group of 19 academics, clinicians, policy makers and patient organisations to consider how health services might overcome some of these barriers. Out of this, three pillars of action were identified:

- 3.1.1 Technology – implement and use smart pumps; design out errors with technology; design technology with future care settings in mind.
- 3.1.2 Health System – Implement meaningful standardisation; reconsider assessment of device safety and efficacy; improve co-ordination of information across organisations; ensure that changes in practice are measured and evaluated robustly.
- 3.1.3 Workforce – strengthen the culture of patient safety; provide staff with requisite support; ensure that professional standards reflect modern practice

In light of this, a follow-up symposium was held in November 2020 to consider how the technology and health systems pillars might be operationalised. This half-day symposium was hosted virtually, and attended by 36 invited experts (including the six authors and all acknowledged participants) to develop a practical strategy for

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

mitigating the barriers and obstacles identified the year previously. BD provided access to a video conferencing platform and assisted in the identification of participants. All participants were invited to the symposium through the corresponding author (AS). All participants gave their time freely and joined voluntarily.

The symposium was run over two sessions. Session 1 was a forum to share recent research in the context of the current understanding of IV medication error. BDF, MJ, MH, SA and AP presented overviews of their work on the causative factors of IV medication error, the science of information presentation, progress and feasibility of national standardisation of infusion concentrations, and the organisational aspects of smart pump introduction. Session 2 consisted of an open-floor discussion on how to scale up the work that has gone before, and how to operationalise recommendations from recent directives as outlined in Table 1 above. There was also consideration of the work needed to build multi-disciplinary partnerships between clinicians, academics, human factors and ergonomics practitioners, patients and industry to ensure that future interventions were evaluated robustly.

3.2 Areas of Discussion at the Symposium

The understanding of IV medication processes is poor, with a high prevalence of operational protocol deviations and low incidence of patient harm.[22] The symposium agreed that work to standardise processes and information may be beneficial. Difficulty finding and understanding information on IV medication administration can be a cause of medication errors.[23] Assessing the performance

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

of such guidelines in the hands of 'real' users and subsequently redesigning them to prevent problems might lead to safer medicines administration.[24] The organisation-wide standardised provision of infusions in ready-to-administer presentations in a large children's hospital led to the almost complete elimination of medication administration errors.[25–27] A similar project in Ireland demonstrated the feasibility of implementing smart pumps and standard concentration infusions in paediatric services using a single national drug library, with evidence of effectiveness.[11,28,29] The implementation of these interventions however have required ongoing sponsorship and co-ordination at a system-level, and cannot be led or managed at a local level. To support this need for high-level support and sponsorship, the unpublished experience from a large multi-site English NHS organisation was presented. Using localised implementation, drug library compliance was poor. When widened to include executive level leadership, robust outcome measures and adequate resource, the proportion of infusions administered via the drug library increased from 50% to over 80%. The ongoing non-compliance was related to physical device limitations and the ongoing burden of both training and maintenance. All symposium participants agreed that smart pump implementation is not something that can be done on a small scale, but needs to be system wide, and executive led.

After panel discussion, the following strategic recommendations were agreed.

3.3 Recommendation 1: Standardise infusion concentrations

There are sufficient data to indicate that ward-prepared infusion solutions are often different to the concentrations prescribed.[30–33] This is not negligence, nor can it be "trained out"[32] but is likely related to humans being human, and the lack of

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

suitable medical devices designed to measure very small volumes. The availability of manufactured ready-to-administer dosage forms, prepared from large volume stock solutions, is likely to be the only way to mitigate these issues. Particularly in paediatric care, there is still a preponderance towards weight based infusions which exacerbate the issues with accuracy above.[34] Variable infusion practices also lead to practice discrepancies and errors in preparation and administration.[19,35–38] We recommend that infusion concentrations should be standardised. The NHS England Aseptic Services Review has standardisation of formulations at its core.[9] There is considerable work already achieved in this sphere with adult intensive care and paediatric solutions already defined.[14,39]

3.4 Recommendation 2: Develop and implement drug libraries using human-centred approaches

While the data exploring the benefits of smart pump use are largely equivocal, this may be at least partly related to low uptake of smart pumps across healthcare systems and poor drug library design.[3,40,41] However it is clear that smart pumps have a potential role in all areas where injectable medicines are used, to provide support for adoption and administration of standard infusions safely and consistently,[29,36] to offer opportunities for system-wide learning, and to protect patients.[9,36] Smart pumps are most likely to prevent the most catastrophic events, with little impact on the day to day discrepancies and deviations seen in large, well-powered observational studies. Two multicenter studies in the US and UK both identified policy violations, including unauthorized administrations and failure to use the drug library where available, as the most commonly observed deviations.[3,42] In a post-hoc analysis of both studies, it was identified that smart pumps would not

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

have prevented the most serious errors, though there may have been a slightly higher error rate in the English centres, where adoption of the technology lagged behind that of the US.[19] Many of the issues with compliance and overrides of smart pump programming may be overcome with a holistic programme of training, review and user support.[19,21,40]

Across health systems, smart pumps will likely become the standard-of-care within five to ten years as providers come to replace obsolete infusion systems.[37] Given the diversity of end-users of these systems, there is a risk that localised implementations will lead to degradation of the benefits of the systems as operators have to adapt to variant practices.[43] The burden and risk associated with the development, validation and implementation of drug libraries in modern complex healthcare systems should not be underestimated and needs to be resourced and overseen at organizational level.[8] We recommend that drug libraries should be centralised across organisations, and constructed based on a standardized nomenclature derived from study and synthesis of existing datasets and practices as has been demonstrated in the UK.[34,47] Furthermore, this standardised nomenclature should ensure interoperability across platforms. However it would be naïve to think that one standard library will be suitable for all users. It is likely that a certain level of localisation will be necessary, which access to a pre-defined library could simplify.

In NHS organisations, oversight for this should lie with Drug & Therapeutics Committees, Medication Safety Officers and Medical Device Safety Officers working in conjunction with Patient Safety Specialists and human factors and ergonomics

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

practitioners. These groups must be aware that engagement with stakeholders involved in the procurement and clinical set up of these systems is critical and be sensitive to the range of differing perspectives . They must also have technological awareness of the capabilities of these systems. There must also be mechanisms in place to ensure that when changes are made to systems, infusion devices and smart pump libraries, that these are properly evaluated to ensure that the intended outcomes are achieved.

Drug libraries and smart pumps themselves may introduce potential for different types of medication error [43] because they are a complex technology that changes the way operators interact with them and the medications that are being administered.[18,44] As few studies have taken a systems-focused approach to the development and implementation of smart pumps and drug libraries[45,46] we recommend the engagement of HF/E specialists to support this work.

In the UK, the HSIB has already recommended that the NHS Injectable Medicines Guide advisory board should "...develop validated drug libraries for smart infusion pumps." [8] These libraries would be applicable across all of the UK. They also observe that a configuration control and management system should be specified. We view the standardisation of drug concentrations and infusion products (including presentation and labelling) as an important potential outcome of this work and thus these two strategies are complementary, requiring collaboration with the same stakeholders.

3.5 Recommendation 3: Develop Standardised Definitions and Metrics

There is great variation in the definitions and metrics that have been used to populate and implement smart pumps, and evaluate the impact and outcomes associated with them. This is part of a wider definitional challenge around medication safety in general [39]. For onward analysis and to support system wide learning, clear, consensus-based definitions for pump-related events, with reference to the limited contextual data held within smart pump systems, are needed to present reliable and robust data.

Formal research methods and iterative improvement approaches provide an opportunity to standardise the definitions and parameters used to evaluate the impact of these interventions. These metrics must define and capture patient outcomes (actual and potential) and economic outcomes (harm caused or avoided, costs, efficiencies and cost effectiveness).

Many organisations do not utilise smart pump data because of human resource limitations and a lack of reliable infrastructure e.g. wireless connectivity of devices and real-time pump downloads.[20] In the UK, there is often a reliance on pump manufacturers to store and process the data. Furthermore, manufacturers do not currently collect or report data from their devices in a standardised way and this makes comparison difficult. Thus, using the parameters and definitions arrived at in section 4.5 should enable this data to be held and managed by health systems, and

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

reviewed regularly at local and national levels by designated groups. There is evidence that regular review of smart pump data informs amendments to drug libraries and datasets and leads to a significant reduction in errors.[51] By curating and managing this data, there is also the potential for it to be useful in the field of artificial intelligence and machine learning for the analysis of large healthcare datasets which to our knowledge has not yet been employed for smart pump data. Given the potential size of these datasets, the strategies above open the door to innovation in this field but the resources required should not be underestimated.

3.6 Recommendation 4: Ensure stakeholder engagement and representation

IV medication administration is an extremely complex system made up of multiple actors. Stakeholders must be identified in order to:

- Develop the components of this IV therapy strategy using all available data and evidence
- Agree and ratify interventions (including smart pumps, standard infusions, barcode systems, electronic patient record alignment with smart pump data sets)
- Support the design and implementation of these interventions in the clinical setting

There is a need to balance the clinical needs of health systems with the safety benefits that these systems may offer, as well as having a comprehensive and affordable package to implement. There should be drug library groups (local,

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

regional, or national) who can oversee the content and nature of these (and in turn, the wider IV system) and provide organisational insight and validation. Having nationally validated libraries and frameworks for implementation will reduce much of the operational burden associated with the introduction of these systems into routine practice. The work involved in the standardisation of infusion concentrations, packaging design, development of drug libraries and proposing standards for implementation is substantial. Adequate resources need to be provided to ensure these strategies can be implemented.

3.7 Recommendation 5: Leverage input into device and systems design

As large organisational purchasers of these technologies, healthcare systems have the opportunity to influence infusion device and management system design with manufacturers. Additionally as part of their commissioning and implementation, organisations should undertake review of their processes and systems within which devices will be used, in order to inform the system-wide adjustments that will be required to accommodate these devices. Furthermore, these strategic recommendations should be used to influence regulatory agencies towards a human-factors informed design approach to device interfaces.[48] It is important for there to be meaningful engagement between manufacturers, patient representatives and practitioners with appropriate support from human factors and ergonomics practitioners. It has been recognised that design disparities between systems can contribute to adverse events and that a more considered approach to the design and implementation of all facets of smart pumps, while ensuring compliance with digital

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

safety standards, is needed. This includes how the devices interface with existing and future electronic patient record systems and how operators interact with the devices, software libraries and nomenclature and presentation of information. [40]

4 FUTURE RESEARCH OPPORTUNITIES

The symposium also considered the necessary research which will underpin and secure the use of these systems. With a national strategy and standardised definitions, there is an opportunity to undertake real-world implementation studies using a uniform approach to quantify improvements in safety and efficiency.

Suggested research questions are:

- What is the best approach to the development and implementation of closed-loop medication administration systems in large complex distributed healthcare systems?
- What is the impact of closed-loop systems incorporating electronic prescribing and medication administration systems, smart pumps and standardised infusion concentrations on medication errors in a real-world context?
- What are the effects on nursing resource utilisation in acute care from the adoption of ready-to-administer IV dosage forms?
- Does the use of user-tested IV administration guidelines improve patient safety and/or nursing resource utilisation in real-world practice?
- What is the utility of machine learning and artificial intelligence to identify error-causing conditions from large smart pump datasets, and to identify unrecognised design issues?

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

In relation to aseptic preparation of products:

- What is the drug stability data available and/or required for different drugs and diluents to present economically feasible products to practitioners?
- What is the safest approach for labelling ready-to-administer products to reduce the likelihood of administration errors?

5 CONCLUSIONS

Healthcare systems are at a turning point in their approach to IV therapy and systems redesign. Now more than ever, there is a clear need for an IV medication safety strategy that is evidence based and rooted in real-world practice. We believe that this paper brings forward a comprehensive and actionable package of measures that is multidisciplinary in nature, and implementable at a strategic and operational level. Previous guidance has been abstract and profession-specific, leaving service leaders, clinicians and researchers at odds or with conflicting goals. We believe that the principles outlined in this paper, though reflective of the UK policy context, can be adopted across the majority of healthcare systems. Technology has been seen as an “easy” solution to systems safety in the past, but this is no longer the case. Current policy in patient safety increasingly requires the use of technology, but with a greater focus on implementation and adaptation.[49] This is a considerable expectation and delivery requires a considered, collaborative and coordinated approach. There is a risk that if left to individual organisations, a confused and

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

variable landscape of implementation will emerge that will detract from the benefits that can be realised.

A strategy of definition and development, that is adequately resourced and lead, to support organisations to implement new technology and systems is recommended.

A programme of research to support rational and rigorous evaluation of this system-wide approach to realise the true patient safety and economic benefits is also required.

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

6 REFERENCES

1. Panagioti M, Khan K, Keers RN, Abuzour A, Phipps D, Kontopantelis E, et al. Prevalence, severity, and nature of preventable patient harm across medical care settings: systematic review and meta-analysis. *BMJ*. 2019;366:l4185. Available from: <http://dx.doi.org/10.1136/bmj.l4185>
2. Sutherland A, Canobbio M, Clarke J, Randall M, Skelland T, Weston E. Incidence and prevalence of intravenous medication errors in the UK: a systematic review. *Eur J Hosp Pharm*. British Medical Journal Publishing Group; 2018;ejhpharm-2018-001624. Available from: <http://ejhp.bmj.com/lookup/doi/10.1136/ejhpharm-2018-001624>
3. Lyons I, Furniss D, Blandford A, Chumbley G, Iacovides I, Wei L, et al. Errors and discrepancies in the administration of intravenous infusions: a mixed methods multihospital observational study. *BMJ Qual Saf*. 2018;27:892–901. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/29627799>
4. Staines A, Amalberti R, Berwick DM, Braithwaite J, Lachman P, Vincent CA. COVID-19: patient safety and quality improvement skills to deploy during the surge. *Int J Qual Heal Care*. Oxford University Press (OUP); 2021;33:1–3. Available from: <https://academic.oup.com/intqhc/advance-article/doi/10.1093/intqhc/mzaa050/5836316>
5. Donaldson LJ, Kelley ET, Dhingra-Kumar N, Kieny M-P, Sheikh A. Medication Without Harm: WHO's Third Global Patient Safety Challenge [Internet]. 2017. Available from: <http://www.who.int/mediacentre/news/>
6. Bhatti S. Adoption of closed loop medicines administration into the NHS - The Pharmaceutical Journal. *Pharm J*. 2019; Available from: <https://pharmaceutical->

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

journal.com/article/opinion/adoption-of-closed-loop-medicines-administration-into-the-nhs

7. Paul Lee, NAMDET. National DERS and “Smart Pump” Position Statement

[Internet]. Swansea; 2020 Nov. Available from:

<https://www.flipsnack.com/6B8ED86D75E/ders-report/full-view.html>

8. Healthcare Safety Investigation Branch. Procurement, usability and adoption of “smart” infusion pumps [Internet]. London; 2020. Available from:

www.hsib.org.uk/tell-us-what-you-think

9. Lord Carter of Coles. Transforming NHS Pharmacy Aseptic Services in England A national report for the Department of Health and. London; 2020.

10. NHS Scotland Shared Services. National Pharmacy Aseptic Dispensing Programme (NPAD) [Internet]. Edinburgh; 2018. Available from:

www.sharedservices.scot.nhs

11. Howlett MM, Butler E, Lavelle KM, Cleary BJ, Breatnach C V. The Impact of Technology on Prescribing Errors in Pediatric Intensive Care: A Before and After Study. *Appl Clin Inform*. Thieme Medical Publishers; 2020;11:323. Available from: [/pmc/articles/PMC7202919/](https://pubmed.ncbi.nlm.nih.gov/37202919/)

12. Plumridge RJ, Sewell GJ. Dose-banding of cytotoxic drugs: A new concept in cancer chemotherapy. *Am J Heal Pharm*. 2001;58:1760–4. Available from:

<http://www.ajhp.org/content/ajhp/58/18/1760.full.pdf>

13. Cree ML, Stocker CF, Tu QM, Scaini LF. Adherence to standard medication infusion concentrations and its impact on paediatric intensive care patient outcomes.

Aust Crit Care. Australian College of Critical Care Nurses Ltd; 2017;3–7. Available

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

from: <http://dx.doi.org/10.1016/j.aucc.2017.07.003>

14. Titiesari YD, Barton G, Borthwick M, Keeling S, Keeling P. Infusion medication concentrations in UK's critical care areas: Are the Intensive Care Society's recommendations being used? *J Intensive Care Soc.* 2017;18:30–5. Available from: <http://inc.sagepub.com/lookup/doi/10.1177/1751143716662664>

15. NHS England » Global Digital Exemplars [Internet]. Available from: <https://www.england.nhs.uk/digitaltechnology/connecteddigitalsystems/exemplars/>

16. Keers RN, Williams SD, Cooke J, Ashcroft DM. Understanding the causes of intravenous medication administration errors in hospitals: a qualitative critical incident study. *BMJ Open.* 2015;5:e005948.

17. Furniss D, Lyons I, Franklin BD, Mayer A, Chumbley G, Wei L, et al. Procedural and documentation variations in intravenous infusion administration: a mixed methods study of policy and practice across 16 hospital trusts in England. *BMC Health Serv Res.* 2018;18:270. Available from: <https://doi.org/10.1186/s12913-018-3025-x>

18. Furniss D, Mayer A, Franklin BD, Blandford A. Exploring structure, agency and performance variability in everyday safety: An ethnographic study of practices around infusion devices using distributed cognition. *Saf Sci.* 2019;118:687–701. Available from: <https://doi.org/10.1016/j.ssci.2019.06.006>

19. Blandford A, Dykes PC, Franklin BD, Furniss D, Galal-Edeen GH, Schnock KO, et al. Intravenous Infusion Administration: A Comparative Study of Practices and Errors Between the United States and England and Their Implications for Patient Safety. *Drug Saf.* Springer International Publishing; 2019;42:1157–65. Available

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

from: <http://link.springer.com/10.1007/s40264-019-00841-2>

20. Sutherland A, Canobbio M, Clarke J, Randall M, Skelland T, Weston E.

Incidence and prevalence of intravenous medication errors in the UK: a systematic review. *Eur J Hosp Pharm*. BMJ Publishing Group; 2020;27:3–8. Available from: <https://ejhp.bmj.com/lookup/doi/10.1136/ejhpharm-2018-001624>

21. Ohashi K, Dalleur O, Dykes PC, Bates DW. Benefits and Risks of Using Smart Pumps to Reduce Medication Error Rates: A Systematic Review. *Drug Saf*. 2014;37:1011–20.

22. Blandford A, Furniss D, Galal-Edeen GH, Chumbley G, Wei L, Mayer A, et al. Intravenous infusion practices across England and their impact on patient safety: a mixed-methods observational study. *Heal Serv Deliv Res*. National Institute for Health Research; 2020;8:1–116.

23. Jones MD, Clarke J, Feather C, Franklin BD, Sinha R, Maconochie I, et al. Use of Pediatric Injectable Medicines Guidelines and Associated Medication Administration Errors: A Human Reliability Analysis. *Ann Pharmacother*. SAGE Publications; 2021;55:1333–40. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/33641479>

24. Jones MD, McGrogan A, Raynor DK, Watson MC, Franklin BD. User-testing guidelines to improve the safety of intravenous medicines administration: a randomised in situ simulation study. *BMJ Qual Saf*. BMJ Publishing Group Ltd; 2021;30:17–26. Available from: <http://dx.doi.org/10.1136/bmjqs-2020-010884>

25. Perkins J, Aguado-Lorenzo V, Arenas-Lopez S. Standard concentration infusions in paediatric intensive care: the clinical approach. *J Pharm Pharmacol*. 2017;69:537–

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

43. Available from: <http://doi.wiley.com/10.1111/jphp.12604>
26. Arenas-López S, Stanley IM, Tunstell P, Aguado-Lorenzo V, Philip J, Perkins J, et al. Safe implementation of standard concentration infusions in paediatric intensive care. *J Pharm Pharmacol*. 2017;69:529–36. Available from: <http://doi.wiley.com/10.1111/jphp.12580>
27. Rashed AN, Whittlesea C, Forbes B, Tomlin S. The feasibility of using dose-banded syringes to improve the safety and availability of patient-controlled opioid analgesic infusions in children. *Eur J Hosp Pharm*. 2014;21:306–8. Available from: <https://ejhp.bmj.com/lookup/doi/10.1136/ejhpharm-2014-000441>
28. Howlett M, Curtin M, Doherty D, Gleeson P, Sheerin M, Breatnach C. Paediatric Standardised Concentration Infusions - a National Solution. *Arch Dis Child*. 2016;101:e2. Available from: <http://search.ebscohost.com/login.aspx?direct=true&db=mdc&AN=27540257&lang=pt-br&site=ehost-live>
29. Howlett MM, Breatnach C V., Brereton E, Cleary BJ. Direct Observational Study of Interfaced Smart-Pumps in Pediatric Intensive Care. *Appl Clin Inform*. Georg Thieme Verlag; 2020;11:659–70. Available from: <http://www.thieme-connect.de/DOI/DOI?10.1055/s-0040-1716527>
30. Aguado-Lorenzo V, Weeks K, Tunstell P, Turnock K, Watts T, Arenas-Lopez S. Accuracy of the concentration of morphine infusions prepared for patients in a neonatal intensive care unit. *Arch Dis Child*. 2013;98:975–9. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/24148890>
31. Parshuram CS, Ng GYT, Ho TKL, Klein J, Moore AM, Bohn D, et al.

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

Discrepancies between ordered and delivered concentrations of opiate infusions in critical care. *Crit Care Med.* 2003;31:2483–7. Available from:

<http://www.ncbi.nlm.nih.gov/pubmed/14530755>

32. Parshuram CS, To T, Seto W, Trope A, Koren G, Laupacis A. Systematic evaluation of errors occurring during the preparation of intravenous medication. *Can Med Assoc J.* 2008;178:42–8.

33. Wheeler DW, Degnan BA, Sehmi JS, Burnstein RM, Menon DK, Gupta AK. Variability in the concentrations of intravenous drug infusions prepared in a critical care unit. *Intensive Care Med.* 2008;34:1441–7. Available from:

<http://www.ncbi.nlm.nih.gov/pubmed/18427780>

34. Oskarsdottir T, Harris D, Sutherland A, Wignell A, Christiansen N. A national scoping survey of standard infusions in paediatric and neonatal intensive care units in the United Kingdom. *J Pharm Pharmacol.* John Wiley & Sons, Ltd (10.1111); 2018;70:1324–31. Available from: <http://doi.wiley.com/10.1111/jphp.12992>

35. Lehmann CU, Kim GR, Gujral R, Veltri M a, Clark JS, Miller MR. Decreasing errors in pediatric continuous intravenous infusions. *Pediatr Crit care Med.* 2006;7:225–30.

36. Larsen GY, Parker HB, Cash J, O'Connell M, Grant MC. Standard drug concentrations and smart-pump technology reduce continuous-medication-infusion errors in pediatric patients. *Pediatrics.* 2005;116:e21-5. Available from:

<http://www.ncbi.nlm.nih.gov/pubmed/15995017>

37. American Society of Hospital Pharmacists, Food and Drug Administration. ASHP IV Adult Continuous Infusion Guidelines. Washington DC; 2016.

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

38. ASHP. STANDARDIZE 4 SAFETY INITIATIVE [Internet]. Washington DC; 2020. Available from: <http://www.ihi.org/communities/blogs/does-standardization-mean-the-end-of-autonomy>.

39. RCPCH, NPPG. Standardising intravenous infusion concentrations in children in the UK | RCPCH [Internet]. 2021. Available from: <https://www.rcpch.ac.uk/resources/standardising-intravenous-infusion-concentrations-children-uk>

40. Schnock KO, Dykes PC, Albert J, Ariosto D, Cameron C, Carroll DL, et al. A Multi-hospital Before–After Observational Study Using a Point-Prevalence Approach with an Infusion Safety Intervention Bundle to Reduce Intravenous Medication Administration Errors. *Drug Saf.* 2018;41:591–602. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/29411338>

41. Rothschild JM, Keohane C a, Cook EF, Orav EJ, Burdick E, Thompson S, et al. A controlled trial of smart infusion pumps to improve medication safety in critically ill patients. *Crit Care Med.* 2005;33:533–40. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/15753744>

42. Schnock KO, Dykes PC, Albert J, Ariosto D, Call R, Cameron C, et al. The frequency of intravenous medication administration errors related to smart infusion pumps: A multihospital observational study. *BMJ Qual Saf.* BMJ Publishing Group; 2017;26:131–40. Available from: <https://qualitysafety.bmj.com/content/26/2/131>

43. Kirkendall ES, Timmons K, Huth H, Walsh K, Melton K. Human-Based Errors Involving Smart Infusion Pumps: A Catalog of Error Types and Prevention Strategies. *Drug Saf.* Springer; 2020;1–15. Available from: <http://link.springer.com/10.1007/s40264-020-00986-5>

STRATEGIC RECOMMENDATIONS FOR IMPLEMENTING SMART PUMPS FOR IV MEDICATION SAFETY

44. Carayon P, Wetterneck TB, Hundt AS, Ozkaynak M, Ram P, DeSilvey J, et al. Observing Nurse Interaction with Infusion Pump Technologies [Internet]. *Adv. Patient Saf. From Res. to Implement.* (Volume 2 Concepts Methodol. Agency for Healthcare Research and Quality (US); 2005. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/21249836>
45. Carayon P, Wetterneck TB, Rivera-Rodriguez AJ, Hundt AS, Hoonakker P, Holden R, et al. Human factors systems approach to healthcare quality and patient safety. *Appl Ergon.* Elsevier Ltd; 2014;45:14–25. Available from: <http://dx.doi.org/10.1016/j.apergo.2013.04.023>
46. Carayon P, Wetterneck TB, Cartmill R, Blosky MA, Brown R, Kim R, et al. Characterising the complexity of medication safety using a human factors approach: an observational study in two intensive care units. *BMJ Qual Saf.* 2014;23:56–65.
47. Keeling P, Scales K, Keeling S, Borthwick M. Towards IV standardization in critical care. *Br J Nurs.* 2010;19:30–3.
48. Rajkomar A, Blandford A. Understanding infusion administration in the ICU through Distributed Cognition. *J Biomed Inform.* Elsevier Inc.; 2012;45:580–90. Available from: <http://dx.doi.org/10.1016/j.jbi.2012.02.003>
49. Department of Health and Social Care. *The Future of Healthcare: Our vision for digital, data and technology in health and care* [Internet]. London; 2018. Available from: <https://www.gov.uk/government/publications/the-future-of-healthcare-our-vision-for-digital-data-and-technology-in-health-and-care/the-future-of-healthcare-our-vision-for-digital-data-and-technology-in-health-and-care>