

Embedded deep learning in ophthalmology: making ophthalmic imaging smarter

Petteri Teikari , Raymond P. Najjar, Leopold Schmetterer and Dan Milea

Abstract: Deep learning has recently gained high interest in ophthalmology due to its ability to detect clinically significant features for diagnosis and prognosis. Despite these significant advances, little is known about the ability of various deep learning systems to be embedded within ophthalmic imaging devices, allowing automated image acquisition. In this work, we will review the existing and future directions for ‘active acquisition’-embedded deep learning, leading to as high-quality images with little intervention by the human operator. In clinical practice, the improved image quality should translate into more robust deep learning-based clinical diagnostics. Embedded deep learning will be enabled by the constantly improving hardware performance with low cost. We will briefly review possible computation methods in larger clinical systems. Briefly, they can be included in a three-layer framework composed of edge, fog, and cloud layers, the former being performed at a device level. Improved edge-layer performance via ‘active acquisition’ serves as an automatic data curation operator translating to better quality data in electronic health records, as well as on the cloud layer, for improved deep learning-based clinical data mining.

Keywords: artificial intelligence, deep learning, embedded devices, medical devices, ophthalmic devices, ophthalmology

Received: 17 October 2018; revised manuscript accepted: 20 December 2018.

Introduction

Recent years have seen an explosion in the use of deep learning algorithms for medical imaging,^{1–4} including ophthalmology.^{5–9} Deep learning has been very efficient in detecting clinically significant features for ophthalmic diagnosis^{9,10} and prognosis.^{11,12} Recently, Google Brain demonstrated how one can, surprisingly, predict subject’s cardiovascular risk, age, and sex from a fundus image,¹³ a task impossible for an expert clinician.

Research effort has so far focused on the development of post hoc deep learning algorithms for already acquired data sets.^{9,10} There is, however, growing interest for embedding deep learning at the medical device level itself for real-time image quality optimization, with little or no operator expertise. Most of the clinically available fundus cameras and optical coherence tomography (OCT) devices require the involvement of a skilled operator in order to achieve satisfactory

image quality, for clinical diagnosis. Ophthalmic images display inherent quality variability due to both technical limitations of the imaging devices and individual ocular characteristics. Recent studies in hospital settings have shown that 38% of nonmydriatic fundus images for diabetic screening,¹⁴ and 42–43% of spectral domain (SD)-OCTs acquired for patients with multiple sclerosis¹⁵ did not have acceptable image quality for clinical evaluation.

Desktop retinal cameras have been increasingly replaced by portable fundus cameras in standalone format^{16–18} or as smartphone add-ons,¹⁹ making the retinal imaging less expensive and accessible to various populations. The main drawback of the current generation portable fundus camera is the lower image quality. Some imaging manufacturers have started to include image quality assessment algorithms to provide a feedback for the operator to either re-acquire the image or accept it.²⁰ To the best of

Ther Adv Ophthalmol

1–21

DOI: 10.1177/
2515841419827172

© The Author(s), 2019.
Article reuse guidelines:
sagepub.com/journals-
permissions

Correspondence to:

Petteri Teikari
Visual Neurosciences
Group, Singapore Eye
Research Institute,
Academia, 20 College
Road, Discovery Tower
Level 6, Singapore 169856.
petteri.teikari@gmail.com

Petteri Teikari
Visual Neurosciences
Group, Singapore Eye
Research Institute,
Singapore
Advanced Ocular Imaging,
Lee Kong Chian School
of Medicine, Nanyang
Technological University,
Singapore

Raymond P. Najjar
Visual Neurosciences
Group, Singapore Eye
Research Institute,
Singapore
Ophthalmology and
Visual Sciences Academic
Clinical Program, Duke-
NUS Medical School,
National University of
Singapore, Singapore

Leopold Schmetterer
Visual Neurosciences
Group, Singapore Eye
Research Institute,
Singapore
Advanced Ocular Imaging,
Lee Kong Chian School
of Medicine, Nanyang
Technological University,
Singapore
Center for Medical
Physics and Biomedical
Engineering, Medical
University of Vienna,
Vienna, Austria
Christian Doppler
Laboratory for Ocular
and Dermal Effects
of Thiomers, Medical
University of Vienna,
Vienna, Austria

Dan Milea
Visual Neurosciences
Group, Singapore Eye
Research Institute,
Singapore
Ophthalmology and
Visual Sciences Academic
Clinical Program, Duke-
NUS Medical School,
National University of
Singapore, Singapore
Neuro-Ophthalmology
Department, Singapore
National Eye Centre,
Singapore

our knowledge, no current commercial system is automatically reconstructing ‘the best possible image’ from multiframe image acquisitions.

Embedding of more advanced algorithms and high computation power at the camera level can be referred to as ‘smart camera architectures’,²¹ with or without the use of deep learning. For example, Google launched its Clips camera, and Amazon Web Services (AWS) its DeepLens camera which are capable of running deep learning models within the camera itself without relying on external processing. Verily, the life sciences research organization of Alphabet Inc., partnered with Nikon and Optos to integrate deep learning algorithms for fundus imaging and diabetic retinopathy screening (<https://verily.com/projects/interventions/retinal-imaging/>). Similar implementation of ‘intelligence’ at the device level is happening in various other medical fields,²² including portable medical ultrasound imaging, with more of the traditional signal processing being accelerated graphics processing units (GPUs),²³ with the deep learning integrated at the device level.²⁴

There are various ways of distributing the signal processing from data acquisition to clinical diagnostics. For example, the use of fundus cameras in remote locations with no Internet access requires all the computations to be performed within the device itself, a system which has been implemented by SocialEyes, for retinal screening on GPU-accelerated tablets.²⁵ This computing paradigm, known as *edge computing*,²⁶ is based on locally performed computations, on the ‘edge’,^{27,28} as opposed to cloud computing in which the fundus image is transmitted over the Internet to a remote *cloud* GPU server, allowing subsequent image classification. In some situations, when there is a need for multilayer computational load distribution, additional nodes are inserted between the edge device and the cloud – a computation paradigm known as *mist*²⁹ or *fog computing*.³⁰ This situation applies typically to Internet-of-things (IoT) medical sensors, which often have very little computational capability.³¹

The main aim of the current review is to summarize the current knowledge related to device-level (*edge* computing) deep learning. We will refer to this as ‘active acquisition’, for improved ophthalmic diagnosis via optimization of image quality (Figure 1). We will also overview various possibilities of computing platforms integrate into the

typical clinical workflow with a focus on standard retinal imaging techniques (i.e. fundus photography and OCT).

Embedded ophthalmic devices

Emerging intelligent retinal imaging

The increased prevalence of ophthalmic conditions affecting the retinas and optic nerves of vulnerable populations prompts higher access to ophthalmic care both in developed³³ and developing countries.³⁴ This translates into an increased need of more efficient screening, diagnosis, and disease management technology, operated with no or little training in clinical settings or even at home.¹⁶ Although paraprofessionals with technical training are currently able to acquire fundus images, a third of these images may not be of satisfactory quality, being nongradable,³⁵ due to reduced transparency of the ocular media.

Acquisition of such images may be even more difficult in nonophthalmic settings, such as emergency departments.³⁶ Recent attempts have aimed to automate retinal imaging processing using a clinical robotic platform InTouch Lite (InTouch Technologies, Inc., Santa Barbara, CA, USA)³⁷ or by integrating a motor to the fundus camera for automated pupil tracking (Nexy; Next Sight, Prodenone, Italy).³⁸ These approaches have not been validated clinically and are based on relatively slow motors, possibly not adapted to clinically challenging situations. Automated acquisition becomes even more important with the recent surge of many smartphone-based fundus imagers.³⁹ Due to the pervasiveness of smartphones, this approach would represent a perfect tool for non-eye specialists.⁴⁰

Similar to fundus imaging, OCT systems are getting more portable and inexpensive and would benefit from easier and robust image acquisition.^{17,18,41} Kim and colleagues⁴¹ developed a low-cost experimental OCT system at a cost of US\$ 7200 using a microelectromechanical system (MEMS) mirror⁴² with a tunable variable focus liquid lens to simplify the design of scanning optics, with inexpensive Arduino Uno microcontroller⁴³ and GPU-accelerated mini PC handling the image processing. The increased computing power from GPUs enables some of the hardware design compromises to be offset through computational techniques.^{44,45} For example, Tang and colleagues⁴⁶ employed three GPU units for




Figure 1. Comparison between traditional passive acquisition and intelligent active acquisition approaches for fundus imaging. (Top-left) In passive acquisition, the healthcare professional manually aligns the camera and decides the best moment for image acquisition. This acquisition has to be often repeated, especially if the patient is not compliant, if the pupils are not dilated, or if there are media opacities, that is, cornea scar, cataract, and so on. (Top-right) In an ‘intelligent’ active acquisition process, the device is able vary imaging parameters and iterates automatically frames until the deep learning is been able to reconstruct an image of satisfactory quality. (Bottom) This intelligent acquisition serves as automated data curation operator for diagnostic deep learning networks (C)^{9,10} leading to improved deep leading to better class separation (healthy D vs disease E). In traditional passive acquisition, the image quality is less consistent leading to many false positives [patient from disease population B (cyan) is classified as healthy A (red)] and negatives [patient from healthy population A (red) is classified as disease B (cyan)]. The gray line represents the decision boundary of the classifier,³² and each point represents one patient.

real-time computational adaptive optics (AO) system, and recently Maloca and colleagues⁴⁷ employed GPUs for volumetric OCT in virtual reality environment for enhanced visualization in medical education.

Active data acquisition

The computationally heavier algorithms made possible by the increased hardware performance can be roughly divided into two categories: (1) ‘passive’ single-frame processing and (2) ‘active’ multiframe processing. In our nomenclature, the ‘passive’ techniques refer to the standard way of acquiring ophthalmic images in which an operator takes an image, which is subsequently subjected to various image enhancement algorithms either before being analyzed by clinician or graded automatically by an algorithm.⁴⁸ In ‘active’ image

acquisition, multiple frames of the same structure are obtained either with automatic reconstruction or with interactive operator-assisted reconstruction of the image. In this review, we will focus on the ‘active’ paradigm, where clinically meaningful images would be reconstructed automatically from multiple acquisitions with varying image quality.

One example for the active acquisition in retinal imaging is the ‘Lucky imaging’ approach,^{49,50} in which multiple frames are acquired in quick succession assuming that at least some of the frames are of good quality. In magnetic resonance imaging (MRI), a ‘prospective gating scheme’ is proposed for acquiring because motion-free image acquisition is possible between the cardiovascular and respiration artifacts, iterating the imaging until satisfactory result is achieved.⁵¹ For

three-dimensional (3D) computed tomography (CT), an active reinforcement learning-based algorithm was used to detect missing anatomical structures from incomplete volume data⁵² and trying to re-acquire the missing parts instead of relying just on postacquisition inpainting.⁵³ In other words, the active acquisition paradigms have some level of knowledge of acquisition completeness or uncertainty based on ideal images, for example, via ‘active learning’ framework⁵⁴ or via recently proposed Generative Query Networks (GQNs).⁵⁵

To implement active data acquisition on an ophthalmic imaging device, we need to define a *loss function* (error term for the deep learning network to minimize) to quantify the ‘goodness’ of the image either directly from the image or using some auxiliary sensors and actuators, to drive the automatic reconstruction process. For example, eye movement artifacts during acquisition of OCT can significantly degrade the image quality,⁵⁶ and we would like to quantify the retinal motion either from the acquired frames itself⁵⁷ or using auxiliary sensors such as digital micromirror device (DMD).⁵⁸ The latter approach has also been applied for correction of light scatter by opaque media.⁵⁹ Due to the scanning nature of OCT, one can re-acquire the same retinal volume and merge only the subvolumes that were sampled without artifacts.^{60,61}

Deep learning-based retinal image processing

Traditional single-frame OCT signal processing pipelines have employed GPUs allowing real-time signal processing.^{62,63} GPUs have been increasingly used in medical image processing even before the recent popularity of deep learning.⁶⁴ The GPUs are becoming essentially obligatory with contemporary high-speed OCT systems.⁶⁵ The traditional image restoration pipelines employ the intrinsic characteristics of the image in tasks such as denoising⁶⁶ and deblurring⁶⁷ without considering image statistics of a larger data set.

Traditionally, these multiframe reconstruction algorithms have been applied after the acquisition without real-time consideration of the image quality of the individual frames. Retinal multiframe acquisition such as fundus videography can exploit the redundant information across the consecutive frames and improve the image degradation model over single-frame acquisition.^{68,69} Köhler and

colleagues⁷⁰ demonstrated how a multiframe super-resolution framework can be used to reconstruct a single high-resolution image from sequential low-resolution video frames. Stankiewicz and colleagues⁷¹ implemented a similar framework for reconstructing super-resolved volumetric OCT stacks from several low-quality volumetric OCT scans. Neither of these approaches, however, applied the reconstruction in real time.

In practice, all of the traditional image processing algorithms can be updated for deep learning framework (Figure 2). The ‘passive’ approaches using input–output pairs to learn image processing operators range from updating individual processing blocks,⁷⁴ to joint optimization of multiple processing blocks,^{75,76} or training an end-to-end network such as DeepISP (ISP, Image Signal Processor) to handle image pipeline from raw image toward the final edited image.⁷⁷ The DeepISP network was developed as offline algorithm,⁷⁷ with no real-time optimization of camera parameters during acquisition. Sitzmann and colleagues⁷⁸ extended the idea even further by jointly optimizing the imaging optics and the image processing for extended depth of field and super-resolution.

With deep learning, many deep image restoration networks have been proposed to replace traditional algorithms. These networks are typically trained with input *versus* synthetic corruption image pairs, with the goodness of the restoration measured as the network’s capability to correct this synthetic degradation. Plötz and Roth⁷⁹ demonstrated that the synthetic degradation model had significant limitation, and traditional state-of-the-art denoising algorithm BM3D⁸⁰ was still shown to outperform many deep denoising networks, when the synthetic noise was replaced with real photographic noise. This highlights the need of creating multiframe database of multiple modalities from multiple device manufacturers for realistic evaluation of image restoration networks in general, as was done by Mayer and colleagues⁸¹ by providing a freely available multiframe OCT data set obtained from *ex vivo* pig eyes.

Image restoration. Most of the literature on multiframe-based deep learning has focused on super-resolution and denoising. Super-resolution algorithms aim to improve the spatial resolution of the reconstructed image beyond what could be obtained from a single input frame.⁸² Tao and colleagues⁸³ implemented a deep learning ‘subpixel




Figure 2. Typical image processing operators used in retinal image processing that are illustrated with 2D fundus images for simplicity. (a) Multiple frames are acquired in a quick succession, which are then registered (aligned) with semantic segmentation for clinically meaningful structures such as vasculature (in blue) and optic disc (in green). (b) Region-of-interest (ROI) zoom on optic disc of the registered image. The image is denoised with shape priors from the semantic segmentation to help the denoising to keep sharp edges. The noise residual is normalized for visualization showing some removal of structural information. The denoised image is decomposed⁷² into *base* that contain the texture-free structure (edge-aware smoothing) and the *detail* that contains the residual texture without the vasculature and optic disc. (c) An example of how the decomposed parts can be edited 'layer-wise'⁷³ and combined to *detail enhanced* image, in order to allow for optimized visualization of the features of interest.

motion compensation' network for video input capable of learning the inter-frame alignment (i.e. image registration) and motion compensation needed for video super-resolution. In retinal imaging, especially with OCT, typical problems for efficient super-resolution are the retinal motion, lateral resolution limits set by the optical media, and image noise. Wang and colleagues⁸⁴ demonstrated using photographic video that motion compensation can be learned from the data, simplifying data set acquisition for retinal deep learning training.

Deblurring (or deconvolution), close to denoising, allows the computational removal of static and movement blur from acquired images. In most cases, the exact blurring point spread function (PSF) is not known and has to be estimated (blind deconvolution) from an acquired image⁸⁵ or sequential images.⁸⁶ In retinal imaging, the most common source for image deblurring is retinal motion,⁵⁶ scattering caused by ocular media opacities,⁸⁷ and optical aberrations caused by the optical characteristics of the human eye itself.⁸⁸ This estimation problem falls under the umbrella term *inverse problems* that have been solved with deep learning recently.⁸⁹

Physical estimation and correction of the image degradation. Efficient PSF estimation retinal imaging can be augmented with auxiliary sensors trying to measure the factors causing retina to move during acquisition. Retinal vessel pulsations due to pressure fluctuations during the cardiac cycle can impact the quality. Gating allows imaging during diastole, when pressure remains almost stable.⁹⁰ Optical methods exist for measuring retinal movement directly using, for example, DMDs⁵⁸ and AO systems measuring the dynamic wavefront aberrations as caused, for instance, by tear film fluctuations.⁸⁸

All these existing physical methods can be combined with deep learning, providing the measured movements as intermediate targets for the network to optimize.⁹¹ Examples of such approaches are the works by Bollepalli and colleagues,⁹² who provided training of the network for robust heart-beat detection, and Li and colleagues,⁹³ who have estimated the blur PSF of light scattered through a glass diffuser simulating the degradation caused by cataract for retinal imaging.

Fei and colleagues⁹⁴ used pairs of uncorrected and AO-corrected scanning laser ophthalmoscopic

(AOSLO) images for learning a 'digital AO' correction. This type of AO-driven network training in practice might be very useful, providing a cost-effective version of super-resolution imaging. For example, Jian and colleagues⁹⁵ proposed to replace deformable mirrors with waveform-correcting lens lowering the cost and simplifying the optical design,⁹⁵ Carpentras and Moser⁹⁶ demonstrated a see-through scanning ophthalmoscope without AO correction, and very recently a hand-held AOSLO imager based on the use of miniature MEMS mirrors was demonstrated by DuBose and colleagues.⁹⁷

In practice, all the discussed hardware and software corrections are not applied simultaneously, that is, joint image restoration with image classification.⁷⁵ Thus, the aim of these operations is to achieve image restoration without loss of clinical information.

High-dynamic-range ophthalmic imaging. In ophthalmic applications requiring absolute or relative pixel intensity values for quantitative analysis, as in fundus densitometry,⁹⁸ or Purkinje imaging for crystalline lens absorption measurements,⁹⁹ it is desirable to extend the intensity dynamic range from multiple differently exposed frames using an approach called high-dynamic-range (HDR) imaging.¹⁰⁰ OCT modalities requiring phase information, such as motion measurement, can benefit from higher bit depths.¹⁰¹ Even in simple fundus photography, the boundaries between optic disc and cup can sometimes be hard to delineate in some cases due to overexposed optic disc compared with surrounding tissue, illustrated by Köhler and colleagues⁷⁰ in their multiframe reconstruction pipeline. Recent feasibility study by Ittarat and colleagues¹⁰² showed that HDR acquisition with tone mapping¹⁰⁰ of fundus images, visualized on standard displays, increased the sensitivity but reduced specificity for glaucoma detection in glaucoma experts. In multimodal or multispectral acquisition, visible light range acquisition can be enhanced by high-intensity near-infrared (NIR) strobe¹⁰³ if the visible light spectral bands do not provide sufficient illumination for motion-free exposure. The vasculature can be imaged clearly with NIR strobe for estimating the motion blur between successive visible light frames.¹⁰⁴

Customized spectral filter arrays. Another operation handled by the ISP is demosaicing¹⁰⁵ which involves interpolation of the color channels. Most

color RGB (red-green-blue) cameras, including fundus cameras, include sensors with a filter grid called Bayer array that is composed of a 2×2 pixel grid with two green, one blue, and one red filter. In fundus imaging, the red channel has very little contrast, and hypothetically custom demosaicing algorithms for fundus ISPs may allow for better visualization of clinically relevant ocular structures. Furthermore, the network training could be supervised by custom illumination based on light-emitting diodes (LEDs) for pathology-specific imaging. Bartczak and colleagues¹⁰⁶ showed that with pathology-optimized illumination, the contrast of diabetic lesions is enhanced by 30–70% compared with traditional red-free illumination imaging.

Recently, commercial sensors with more than three color channels have been released, Omnivision (Santa Clara, CA, USA) OV4682, for example, replaced one green filter of the Bayer array with an NIR filter. In practice, one could acquire continuous fundus video without pupil constriction using just the NIR channel for the video illumination and capturing fundus snapshot simultaneously with a flash of visible light in addition to the NIR.

The number of spectral bands on the filter array of the sensor was extended up 32 bands by Imec (Leuven, Belgium). This enables snapshot multispectral fundus imaging for retinal oximetry.¹⁰⁷ These additional spectral bands or custom illuminants could also be used to aid the image processing itself before clinical diagnostics.¹⁰⁸ For example, segmenting the macular region becomes easier with a spectral band around blue 460 nm, as the macular pigment absorbs strongly at that wavelength and appears darker than its background on this band.¹⁰⁹

Depth-resolved fundus photography. Traditionally, depth-resolved fundus photography has been done via stereo illumination of the posterior pole that involves either dual path optics increasing the design complexity or operator skill to take a picture with just one camera.¹¹⁰ There are alternatives for depth-resolved fundus camera in a compact form factor such as plenoptic fundus imaging that was shown to provide higher degree of stereopsis than traditional stereo fundus photography using an off-the-shelf Lytro Illum (acquired by Google, Mountain View, CA, USA) consumer light field camera.¹¹¹ Plenoptic cameras, however, trade spatial resolution for angular

resolution, for example, Lytro Illum has over 40 million pixels, but the final fundus spatial resolution consists of 635×433 pixels. Simpler optical arrangement for depth imaging with no spatial resolution trade-off is possible with depth-from-focus algorithms¹¹² that can reconstruct depth map from a sequence of images of different focus distances (z -stack). This rapid switching of focus distances can be achieved in practice, for example, using variable-focus liquid lenses, as demonstrated for retinal OCT imaging by Cua and colleagues.¹¹³

Compressed sensing. Especially with OCT imaging, and scanning-based imaging techniques in general, there is a possibility to use compressed sensing to speed up the acquisition and reduce the data rate.¹¹⁴ Compressed sensing is based on the assumption that the sampled signal is sparse in some domain, and thus it can be undersampled and reconstructed to have a matching resolution for the dense grid. Most of the work on combined compressed sensing and deep learning has been on MRI brain scans.¹¹⁵ OCT angiography (OCTA) is a special variant of OCT imaging that acquires volumetric images of the retinal and chorioidal vasculature through motion contrast imaging. OCTA acquisition is very sensitive to motion and would benefit from sparse sampling with optimized scan pattern.¹¹⁶

Defining cost functions. The design of proper cost function used to define suboptimal parts of an image is not trivial at all. Early retinal processing work by Köhler and colleagues¹¹⁷ used the retinal vessel contrast as a proxy measure for image quality, which was implemented later as fast real-time algorithm by Bendaoudi and colleagues.¹¹⁸ Saha and colleagues¹¹⁹ developed a structure-agnostic data-driven deep learning network for flagging fundus images either as acceptable for diabetic retinopathy screening or as to be recaptured. In practice, however, the cost function used for deep learning training can be defined in multiple ways as reviewed by Zhao and colleagues.¹²⁰ They compared different loss functions for image restoration and showed that the most commonly used ℓ_2 norm (squared error or ridge regression) was clearly outperformed in terms of perceptual quality by the multiscale structural similarity index (MS-SSIM).¹²¹ This was shown to improve even slightly when the authors combined MS-SSIM with ℓ_1 norm (absolute deviation, lasso regression). One could hypothesize that a data-driven quality indicator that reflects the diagnostic differentiation




Figure 3. High-level schematic representation of an adaptive optics retinal imaging system. The wavefront from (a) retina is distorted mainly by (b) the cornea and crystalline lens, which is corrected in our example by (c) lens-based actuator designed for compact imaging systems.⁹⁵ (d) The imaging optical system⁸⁸ is illustrated with a single lens for simplicity. The corrected wavefront on (e) the image sensor is a (h) sharper version of the image that would be of (f) lower quality without (c) the waveform correction. The ‘digital adaptive optics’ (g) *universal function approximator* maps the distorted image (f) to corrected image (h), and the network (g) is the network that was trained with the image pairs (uncorrected and corrected). For simplicity, we have omitted the wavefront sensor from the schematic and estimated the distortion in a *sensorless* fashion.⁸⁸ Images (f) and (h) are courtesy of Professor Stephen A. Burns (School of Optometry, Indiana University) from AOSLO off-axis illumination scheme for retinal vasculature imaging.¹²⁴

capability of the image accompanied with perceptual quality would be optimal particularly for fundus images.

Physics-based ground truths. The unrealistic performance of image restoration networks with synthetic noise and the lack of proper real noise benchmark data sets are major limitations at the moment. Plötz and Roth⁷⁹ created their noise benchmark test by varying the ISO setting of the camera and taking the lowest ISO setting as the ground truth ‘noise-free’ image. In retinal imaging, construction of good-quality ground truth requires some special effort. Mayer and colleagues⁸¹ acquired multiple OCT frames of *ex vivo* pig eyes to avoid motion artifacts between acquisitions for speckle denoising.

In humans, commercially available laser speckle reducers can be used to acquire image pairs with two different levels of speckle noise^{122,123} (Figure 3). Similar pair for deblurring network training

could be acquired with and without AO correction¹²⁵ (see Figure 3). In phase-sensitive OCT application such as elastography, angiography, and vibrometry, a dual beam setup could be used with a highly phase-stable laser as the ground truth and ‘ordinary’ laser as the input to be enhanced.¹²⁶

Emerging multimodal techniques, such as combined OCT and SLO,¹²⁷ and OCT with photoacoustic microscopy (PAM), optical Doppler tomography (ODT),¹²⁸ and fluorescence microscopy,¹²⁹ enable interesting joint training from complementary modalities with each of them having different strengths. For example, in practice, the lower quality but inexpensive modality could be computationally enhanced.¹³⁰

Inter-vendor differences could be further addressed by repeating each measurement with different OCT machines as taken into account with clinical diagnosis network by De Fauw and

colleagues.⁹ All these hardware-driven signal restorations could be further combined with existing traditional filters and the filter output could be used as targets for so-called ‘copycat’ filters that can estimate existing filters.¹³¹

Quantifying uncertainty. Within the automatic ‘active acquisition’ scheme, it is important to be able to localize the quality problems in an image or in a volume.^{132,133} Leibig and colleagues¹³⁴ investigated the commonly used Monte Carlo dropout method¹³² for estimating the uncertainty in fundus images for diabetic retinopathy screening and its effect on clinical referral decision quality. The Monte Carlo dropout method improved the identification of substandard images that were either unusable or had large uncertainty on the model classification boundaries. Such an approach should allow rapid identification of patients with suboptimal fundus images for further clinical evaluation by an ophthalmologist.

Similar approach was taken per-patch uncertainty estimation in 3D super-resolution¹³⁵ and in voxel-wise segmentation uncertainty.¹³⁶ Cobb and colleagues¹³⁷ demonstrated an interesting extension to this termed ‘loss-calibrated approximate inference’ that allowed the incorporation of *utility function* to the network. This utility function was used to model the asymmetric clinical implications between prediction of *false negatives* and *false positives*.

The financial and quality-of-life cost of an uncertain patch in an image leading to *false-negative* decision might be a lot larger than *false-positive* that might just lead to an additional checkup by an ophthalmologist. The same utility function could be expanded to cover disease prevalence¹³⁸; enabling end-to-end screening performance to be modeled for diseases such as glaucoma with low prevalence needs very high performance in order to be cost-efficient to screen.¹³⁹

The regional uncertainty can then be exploited during active acquisition by guiding the acquisition iteration to only that area containing the uncertainty. For example, some CMOS sensors (e.g. Sony IMX250) allow readout from only a part of the image, faster than one could do for the full frame. One scenario for smarter fundus imaging could, for example, involve initial imaging with the whole field of view (FOV) of the device, followed by multiframe acquisition of only the

optic disc area to ensure that the cup and disc are well distinguishable, and that the depth information is of good quality (Figure 4). Similar active acquisition paradigm is in use, for example, in drone-based operator-free photogrammetry. In that application, the drone can autonomously reconstruct a 3D building model from multiple views recognizing where it has not scanned yet and fly to that location to scan more.¹⁴¹

Distributing the computational load

In typical postacquisition disease classification studies with deep learning,¹⁰ the network training has been done on large GPU clusters either locally or using cloud-based GPU servers. However, when embedding deep learning within devices, different design trade-offs need to be taken into account. Both in hospital and remote healthcare settings, proper Internet connection might be lacking due to technical infrastructure or institutional policy limitations. Often, the latency requirements are very different for real-time processing of signals making the use of cloud services impossible.¹⁴² For example, a lag due to poor Internet connection is unacceptable at intensive care units (ICUs) as those seconds can affect human lives, and the computing hardware needs to be placed next to the sensing device.¹⁴³

Edge computing

In recent years, the concept of *edge computing* (Figure 5) has emerged as a complementary or alternative to the cloud computing, in which computations are done centrally, that is, away from the ‘edge’. The main driving factor for edge computing is the various IoT applications¹⁴⁵ or Internet of Medical Things (IoMT).¹⁴⁶ Gartner analyst Thomas Bittman has predicted that the market for processing at the edge will expand to similar or increased levels than the current cloud processing.¹⁴⁷ Another market research study by Grand View Research, Inc.¹⁴⁸ projected edge computing segment for healthcare and life sciences to exceed US\$ 326 million by 2025. Specifically, the edge computing is seen as the key enabler of wearables to become a reliable tool for long-term health monitoring.^{149,150}

Fog computing

In many cases, an intermediate layer called *fog* or *mist computing layer* (Figure 5) is introduced between the edge device and the cloud layer to




Figure 4. (a) Example of re-acquisition using a region of interest (ROI) defined from the initial acquisition (the full frame). The ROI has 9% of the pixels of the full frame making the ROI acquisition a lot faster if the image sensor allows ROI-based readout. (b) Multiframe ROI re-acquisition is illustrated with three low-dynamic range (8-bit LDR) with simulated low-quality camera intensity compression. The underexposed frame (b, left) exposes optic disc correctly with less details visible on darker regions of the image as illustrated by the clipped dark values in histogram (c, left, *clipped values at 0*), whereas the overexposed frame (c, right) exposes dark vasculature with detail while overexposing (c, right, *clipped values at 255*) the bright regions such as the optic disc. The normal exposure frame (b, center) is a compromise (c, center) between these two extreme exposures. (d) When the three LDR frames are combined together using a exposure fusion technique¹⁴⁰ into a high-dynamic range (HDR) image, all the relevant clinical features are exposed to correct possibly improving diagnostics.¹⁰²

distribute the computing load.^{31,151–153} At simplest level, this three-layer architecture could constitute of simple low-power IoT sensor (*edge device*) with some computing power.¹⁵⁴ This IoT device could be, for example, an inertial measurement unit (IMU)-based actigraph that sends data real time to user's smartphone (*fog device*) which contains more computing power than the *edge device* for gesture recognition.¹⁵⁵ The gesture recognition model could be used to detect the falls in elderly or send corrective feedback back to *edge*

device which could also contain some actuators or a display. An example of such actuator could be a tactile buzzer for neurorehabilitation applications¹⁵⁶ or a motorized stage for aligning a fundus camera relative to the patient's eye.¹⁵⁷ The smartphone subsequently sends the relevant data to the cloud for analyzing long-term patterns at both individual and population levels.^{16,158} Alternatively, the sensor itself could do some data cleaning and have the fog node to handle the sensor fusion of typical clinical one-dimensional




Figure 5. Separation of computations to three different layers. (1) *Edge layer* – the computations done at the device level which in active acquisition ocular imaging (top) require significant computational power, for example, in the form of an embedded GPU. With wearable intraocular measurement, the contact lens can house only a very low-power microcontroller (MCU), and it needs to let the (2) *Fog layer* to handle most of the signal cleaning, whereas for ocular imaging, the fog device mainly just relays the acquired image to (3) *Cloud layer*. The standardization of the data structure is ensured through FHIR (Fast Healthcare Interoperability Resources) API (application programming interface)¹⁴⁴ before being stored on secure cloud server. This imaging data along with other clinical information can then be accessed via healthcare professionals, patients, and research community.

(1D) biosignal. An illustration of this concept is the fusion of depth and thermal cameras for hand hygiene monitoring,¹⁵⁹ including indoor position tracking sensors to monitor healthcare processes at a hospital level.

Balancing edge and fog computations

For the hardware used in each node, multiple options exist, and in the literature, very heterogeneous architectures are described for the whole system.^{31,160} For example, in the SocialEyes project,²⁵ the diagnostic tests of MARVIN (for mobile autonomous retinal evaluation) are implemented on GPU-powered Android tablet (NVIDIA SHIELD). In their rural visual testing application, the device needs to be transportable and adapted to the limited infrastructure. In this scenario, most of the computations are already done at the tablet level, and the *fog* device could, for example, be a low-cost community smartphone/Wi-Fi link. The data can then be submitted to the cloud holding the centralized electronic health records (EHRs).¹⁶¹ If the local computations required are not very heavy, both the edge and fog functionalities could be combined into one low-cost Raspberry Pi board computer.¹⁶² In

hospital settings with large patient volumes, it would be preferable to explore different task-specific data compression algorithms at the cloud level to reduce storage and bandwidth requirements. In a teleophthalmology setting, the compression could be done already at the edge level before cloud transmission.¹⁶³

In the case of fundus imaging, most of that real-time optimization would be happening at the device level, with multiple different hardware acceleration options.^{164,165} One could rely on a low-cost computer such as Raspberry Pi¹⁶⁶ and allow for limited computations.¹⁶⁷ This can be extended if additional computation power is provided at the cloud level. In many embedded medical applications, GPU options such as the NVIDIA's Tegra/Jetson platform¹⁶⁸ have been increasingly used. The embedded GPU platforms in practice offer a good compromise between ease-of-use and computational power of Raspberry Pi and desktop GPUs, respectively.

In some cases, the general-purpose GPU (GPGPU) option might not be able to provide the energy efficiency needed for the required computation performance. In this case, field-programmable

gate arrays (FPGAs)¹⁶⁹ may be used as an alternative to embedded GPU, as demonstrated for retinal image analysis¹⁷⁰ and real-time video restoration.¹⁷¹ FPGA implementation may, however, be problematic due to increased implementation complexity. Custom-designed accelerator chips¹⁷² and Application-Specific Integrated Circuit (ASIC)¹⁷³ offer even higher performance but at even higher implementation complexity.

In ophthalmology, there are only a limited number of wearable devices, allowing for continuous data acquisition. Although the continuous assessment of intraocular pressure (IOP) is difficult to achieve, or even controversial,¹⁷⁴ commercial products by Triggerfish® (Sensimed AG, Lausanne, Switzerland) and EYEMATE® (Implandata Ophthalmic Products GmbH, Hannover, Germany) have been cleared by the Food and Drug Administration (FDA) for clinical use.

Interesting future direction for this monitoring platform is an integrated MEMS/microfluidics system¹⁷⁵ that could simultaneously monitor the IOP and has a passive artificial drainage system for the treatment of glaucoma.¹⁷⁶ The continuous IOP measurement could be integrated with ‘point structure + function measures’ for individualized deep learning-driven management of glaucoma as suggested for the management of age-related macular degeneration (AMD).¹¹

In addition to pure computational restraints, the size and the general acceptability of the device by the patients can represent a limiting factor, requiring a more patient-friendly approach. For example, devices analyzing eye movements^{177,178} or pupillary light responses¹⁷⁹ can be better accepted and implemented when using more practical portable devices rather than bulky research lab systems. For example, Zhu and colleagues¹⁸⁰ have designed an embedded hardware accelerator for deep learning inference from image sensors of the augmented/mixed reality (AR/MR) glasses.

This could be in future integrated with MEMS-based camera-free eye tracker chip developed by University of Waterloo spin-off company AdHawk Microsystems (Kitchener, ON, Canada)¹⁸¹ for functional diagnostics or to quantify retinal motion. In this example of eye movement diagnostics, most of the computations might be performed at the device level (*edge*), but the patient could carry a smartphone or a dedicated Raspberry

Pi for further postprocessing and transmission to cloud services.

Cloud computing

The *cloud layer* (Figure 5) is used for centralized data storage, allowing both the healthcare professional and patients to access the EHRs, for example, via the FHIR (Fast Healthcare Interoperability Resources) API (application programming interface).¹⁴⁴ Research groups can analyze the records as already demonstrated for deep learning for retinopathy diagnosis.^{9,10} Detailed analysis of different technical options in the *cloud layer* is beyond the scope of this article, and interested readers are referred to the following clinically relevant reviews.^{182,183}

Discussion

Here, we have reviewed the possible applications of deep learning, introduced at the ophthalmic imaging device level. This extends well-known application of deep learning for clinical diagnostics.^{9,10,48} Such an ‘active acquisition’ aims for automatic optimization of imaging parameters, resulting in improved image quality and reduced variability.⁸ This active approach can be added to the existing hardware or can be combined with novel hardware designs.

The main aim of an embedded intelligent deep learning system is to favor acquisition of a high-quality image or recording, without the intervention of a highly skilled operator, in various environments. There are various healthcare delivery models, in which embedded deep learning could be used in future routine eye examination: (1) patients could self-screen themselves, using a shared device located either in a community clinic or at the supermarket, requiring no human supervision; (2) the patients could be imaged by a technician in a ‘virtual clinic’,¹⁸⁴ in a hospital waiting room before an ophthalmologist appointment, or at the optician (<https://www.aop.org.uk/ot/industry/high-street/2017/05/22/oct-rollout-in-every-specsavers-announced>); (3) patients could be scanned in remote areas by a mobile general healthcare practitioner¹⁸⁵; and (4) the patients themselves could do continuous home monitoring for disease progression.^{16,186} Most of the fundus camera and OCT devices come already with some quality metrics probing the operator to re-take the image, but so far no commercial device is offering sufficient automatic reconstruction, for

example, in presence of ocular media opacities and poorly compliant patients.

Healthcare systems experiencing shortage of manpower may benefit from modern automated imaging. Putting more intelligence at the device level will relieve the healthcare professionals from clerical care for actual patient care.¹⁸⁷ With the increased use of artificial intelligence (AI), the role of the clinician will evolve from the medical paternalism of the 19th century and evidence-based medicine of the 20th century to (big) data-driven clinician working more closely with intelligent machines and the patients.¹⁸⁸ The practical-level interaction with AI is not just near-future science fiction, but very much a reality as the recent paper on ‘augmented intelligence’ in radiology demonstrated.¹⁸⁹ A synergy between clinicians and AI system resulted in improved diagnostic accuracy, compared with clinicians’, and was better than AI system’s own performance.

At healthcare systems level, intelligent data acquisition will provide an additional automated data quality verification, resulting in improved management of data volumes. This is required because size of data is reported to double every 12–14 months,¹⁹⁰ addressing the ‘garbage in–garbage out’ problem.^{190,191} Improved data quality will also allow more efficient EHR mining,¹⁹² enabling the healthcare systems to get closer to the long-term goal of learning healthcare systems¹⁹³ leveraging on prior clinical experience in structured data/evidence-based sense along with expert clinical knowledge.^{188,194}

Despite the recent developments of deep learning in ophthalmology, very few prospective clinical trials per se have evaluated its performance in real, everyday life situations. IDx-DR has recently been approved as the first fully autonomous AI-based FDA-approved diagnostic system for diabetic retinopathy,⁴⁸ but the direct benefit of patients, in terms of visual outcome, is still unclear.¹⁹⁵ Future innovations emerging from tech startups, academia, or from established companies will hopefully improve the quality of the data, through cross-disciplinary collaboration of designers, engineers, and clinicians,^{196,197} resulting in improved outcomes of patients with ophthalmic conditions.

Acknowledgements

The authors would like to acknowledge Professor Stephen Burns (Indiana University) for providing

images to illustrate the adaptive optics deep learning correction.

Funding

The authors disclosed receipt of the following financial support for the research, authorship, and/or publication of this article: This work was supported by National Health Innovation Centre Singapore Innovation to Develop (I2D) Grant (NHIC I2D) (NHIC-I2D-1708181).

Conflict of interest statement

The authors declared no potential conflicts of interest with respect to the research, authorship, and/or publication of this article.

ORCID iD

Petteri Teikari  <https://orcid.org/0000-0003-1095-4185>

References

1. Litjens G, Kooi T, Bejnordi BE, *et al.* A survey on deep learning in medical image analysis. *Med Image Anal* 2017; 42: 60–88.
2. Hinton G. Deep learning – a technology with the potential to transform health care. *JAMA* 2018; 320: 1101–1102.
3. Ching T, Himmelstein DS, Beaulieu-Jones BK, *et al.* Opportunities and obstacles for deep learning in biology and medicine. *J R Soc Interface* 2018; 15: 142760.
4. Mandal S, Greenblatt AB and An J. Imaging intelligence: AI is transforming medical imaging across the imaging spectrum. *IEEE Pulse* 2018; 9: 16–24.
5. Schmidt-Erfurth U, Sadeghipour A, Gerendas BS, *et al.* Artificial intelligence in retina. *Prog Retin Eye Res* 2018; 67: 1–29.
6. Ting DSW, Liu Y, Burlina P, *et al.* AI for medical imaging goes deep. *Nat Med* 2018; 24: 539–540.
7. Hogarty DT, Mackey DA and Hewitt AW. Current state and future prospects of artificial intelligence in ophthalmology: a review. *Clin Exp Ophthalmol*. Epub ahead of print 28 August 2018. DOI: 10.1111/ceo.13381.
8. Lee A, Taylor P, Kalpathy-Cramer J, *et al.* Machine learning has arrived! *Ophthalmology* 2017; 124: 1726–1728.
9. DeFauw J, Ledsam JR, Romera-Paredes B, *et al.* Clinically applicable deep learning for diagnosis and referral in retinal disease. *Nat Med* 2018; 24: 1342–1350.

10. Ting DSW, Cheung CYL, Lim G, *et al.* Development and validation of a deep learning system for diabetic retinopathy and related eye diseases using retinal images from multiethnic populations with diabetes. *JAMA* 2017; 318: 2211–2223.
11. Schmidt-Erfurth U, Bogunovic H, Sadeghipour A, *et al.* Machine learning to analyze the prognostic value of current imaging biomarkers in neovascular age-related macular degeneration. *Ophthalmology Retina* 2018; 2: 24–30.
12. Wen JC, Lee CS, Keane PA, *et al.* Forecasting future Humphrey visual fields using deep learning. *Arxiv:180404543 [Cs, Stat]*, 2018, <http://arxiv.org/abs/1804.04543>
13. Poplin R, Varadarajan AV, Blumer K, *et al.* Prediction of cardiovascular risk factors from retinal fundus photographs via deep learning. *Nature Biomed Eng* 2018; 2: 158–164.
14. Rani PK, Bhattarai Y, Sheeladevi S, *et al.* Analysis of yield of retinal imaging in a rural diabetes eye care model. *Indian J Ophthalmol* 2018; 66: 233–237.
15. Tewarie P, Balk L, Costello F, *et al.* The OSCAR-IB consensus criteria for retinal OCT quality assessment. *PLoS ONE* 2012; 7: e34823.
16. Roesch K, Swedish T and Raskar R. Automated retinal imaging and trend analysis – a tool for health monitoring. *Clin Ophthalmol* 2017; 11: 1015–1020.
17. Monroy GL, Won J, Spillman DR, *et al.* Clinical translation of handheld optical coherence tomography: practical considerations and recent advancements. *J Biomed Opt* 2017; 22: 1–30.
18. Chopra R, Mulholland PJ, Dubis AM, *et al.* Human factor and usability testing of a binocular optical coherence tomography system. *Transl Vis Sci Technol* 2017; 6: 16.
19. Kim TN, Myers F, Reber C, *et al.* A smartphone-based tool for rapid, portable, and automated wide-field retinal imaging. *Transl Vis Sci Technol* 2018; 7: 21–21.
20. Katuwal GJ, Kerekes JP, Ramchandran RS, *et al.* Automated fundus image field detection and quality assessment, 2018, <https://patents.google.com/patent/US9905008B2/en>
21. Brea V, Ginhac D, Berry F, *et al.* Special issue on advances on smart camera architectures for real-time image processing. *J Real-Time Image Process* 2018; 14: 635–636.
22. Zhang B, Tang K and Du J. Influence of intelligent unmanned system on the development of intelligent measuring. In: *Proceedings of the global intelligence industry conference (GIIC 2018)*, vol. 10835, 2018, p. 108350Y. International Society for Optics and Photonics, <https://www.spiedigitallibrary.org/conference-proceedings-of-spie/10835/108350Y/Influence-of-intelligent-unmanned-system-on-the-development-of-intelligent/10.1117/12.2503984.short?SSO=1>
23. Gobl GR, Navab N and Hennemersperger C. SUPRA: open source software defined ultrasound processing for real-time applications. *Int J Comput Assist Radiol Surg* 2018; 13: 759–767.
24. Jarosik P and Lewandowski M. WaveFlow-Towards Integration of Ultrasound Processing with Deep Learning. *arXiv preprint arXiv* 2018, <https://arxiv.org/abs/1811.01566>
25. Hansen T. Social eyes uses deep learning to save sight. *NVIDIA Blog*, 2016, <https://blogs.nvidia.com/blog/2016/02/17/deep-learning>
26. Shi W, Cao J, Zhang Q, *et al.* Edge computing: vision and challenges. *IEEE Internet Things J* 2016; 3: 637–646.
27. Cuff J. Getting to the heart of HPC and AI at the edge in healthcare, 2018, <https://goo.gl/F8psgy>
28. Harris S. The next frontier – medical imaging AI in the age of edge computing, 2018, <https://goo.gl/E26sKs>
29. Barik RK, Dubey AC, Tripathi A, *et al.* Mist data: leveraging mist computing for secure and scalable architecture for smart and connected health. *Procedia Comput Sci* 2018; 125: 647–653.
30. Xu J, Liu H, Shao W, *et al.* Quantitative 3-D shape features based tumor identification in the fog computing architecture. *J Amb Intel Hum Comp* 2018; 9: 1–11.
31. Farahani B, Firouzi F, Chang V, *et al.* Towards fog-driven IoT eHealth: promises and challenges of IoT in medicine and healthcare. *Future Gener Comput Syst* 2018; 78: 659–676. <https://doi.org/10.1016/j.future.2017.04.036>
32. Fawzi A, Moosavi-Dezfooli SM, Frossard P, *et al.* Classification regions of deep neural networks. *Arxiv:170509552 [Cs]*, 2017, <https://arxiv.org/abs/1705.09552>
33. Lee CS, Su GL, Baughman DM, *et al.* Disparities in delivery of ophthalmic care: an exploration of public Medicare data. *PLoS ONE* 2017; 12: e0182598.
34. Sommer A, Taylor HR, Ravilla TD, *et al.* Challenges of ophthalmic care in the developing world. *JAMA Ophthalmol* 2014; 132: 640–644.

35. Davila JR, Sengupta SS, Niziol LM, *et al.* Predictors of photographic quality with a handheld nonmydriatic fundus camera used for screening of vision-threatening diabetic retinopathy. *Ophthalmologica* 2017; 238: 89–99.
36. Hassen GW, Chirurugi R, Menoscal JP, *et al.* All eye complaints are not created equal: the value of hand-held retina camera in the emergency department. *Am J Emerg Med* 2018; 36: 1518.
37. Martel JBA, Anders UM and Kravchuk V. Comparative study of teleophthalmology devices: smartphone adapted ophthalmoscope, robotic ophthalmoscope, and traditional fundus camera—The recent advancements in telemedicine. *New Front Ophthalmol* 2015; 1(1): 2–5.
38. Nexy robotic retinal imaging system cleared by the FDA, for the US and Market, 2018, <https://www.prweb.com/releases/2018/06/prweb15554831.htm>
39. Barikian A and Haddock LJ. Smartphone assisted fundus funduscopy/photography. *Curr Ophthalmol Rep* 2018; 6: 46–52.
40. Bifolck E, Fink A, Pedersen D, *et al.* Smartphone imaging for the ophthalmic examination in primary care. *JAAPA* 2018; 31: 34–38.
41. Kim S, Crose M, Eldridge WJ, *et al.* Design and implementation of a low-cost, portable OCT system. *Biomed Opt Express* 2018; 9: 1232–1243.
42. Lin L, Keeler E, Lin LY, *et al.* Progress of MEMS scanning micromirrors for optical bio-imaging. *Micromachines* 2015; 6: 1675–1689.
43. Teikari P, Najjar RP, Malkki H, *et al.* An inexpensive Arduino-based LED stimulator system for vision research. *J Neurosci Methods* 2012; 211: 227–236.
44. Altmann Y, McLaughlin S, Padgett MJ, *et al.* Quantum-inspired computational imaging. *Science* 2018; 361: eaat2298.
45. Liu YZ, South FA, Xu Y, *et al.* Computational optical coherence tomography. *Biomed Opt Express* 2017; 8: 1549–1574.
46. Tang H, Mulligan JA, Untracht GR, *et al.* GPU-based computational adaptive optics for volumetric optical coherence microscopy. In: *Proceedings of the high-speed biomedical imaging and spectroscopy: toward big data instrumentation and management*, vol. 9720. International Society for Optics and Photonics, <https://spie.org/Publications/Proceedings/Volume/9720>
47. Maloca PM, deCarvalho JER, Heeren T, *et al.* High-performance virtual reality volume rendering of original optical coherence tomography point-cloud data enhanced with real-time ray casting. *Transl Vis Sci Technol* 2018; 7: 2.
48. Abràmoff MD, Lavin PT, Birch M, *et al.* Pivotal trial of an autonomous AI-based diagnostic system for detection of diabetic retinopathy in primary care offices. *Npj Digital Medicine* 2018; 1: 39.
49. Samaniego A, Boominathan V, Sabharwal A, *et al.* MobileVision: a face-mounted, voiceactivated, non-mydratic ‘lucky’ ophthalmoscope. In: *Proceedings of the wireless health 2014 on National Institutes of Health WH '14*, pp. 2:1–2:8. New York: ACM, <https://www.ece.rice.edu/~av21/Documents/2014/mobileVision.pdf>
50. Lawson ME and Raskar R. Methods and apparatus for retinal imaging, 2016, <https://patents.google.com/patent/US9295388B2/en>
51. Kinchesh P, Gilchrist S, Beech JS, *et al.* Prospective gating control for highly efficient cardiorespiratory synchronised short and constant TR MRI in the mouse. *Magn Resonan Imag* 2018; 53: 20–27.
52. Ghesu FC, Georgescu B, Grbic S, *et al.* Towards intelligent robust detection of anatomical structures in incomplete volumetric data. *Med Image Anal* 2018; 48: 203–213.
53. Skalic M, Varela-Rial A, Jiménez J, *et al.* LigVoxel: inpainting binding pockets using 3D-convolutional neural networks. *Bioinformatics* 2019; 35: 243–250.
54. Gal Y, Islam R and Ghahramani Z. Deep Bayesian active learning with image data. *Arxiv:170302910 [Cs, Stat]*, 2017, <http://arxiv.org/abs/1703.02910>
55. Eslami SMA, Rezende DJ, Besse F, *et al.* Neural scene representation and rendering. *Science* 2018; 360: 1204–1210.
56. Baghaie A, Yu Z and D’Souza RM. Involuntary eye motion correction in retinal optical coherence tomography: hardware or software solution. *Med Image Anal* 2017; 37: 129–145.
57. Sheehy CK, Yang Q, Arathorn DW, *et al.* High-speed, image-based eye tracking with a scanning laser ophthalmoscope. *Biomed Opt Express* 2012; 3: 2611–2622.
58. Vienola KV, Damodaran M, Braaf B, *et al.* In vivo retinal imaging for fixational eye motion detection using a high-speed digital micromirror device (DMD)-based ophthalmoscope. *Biomed Opt Express* 2018; 9: 591–602.

59. Turpin A, Vishniakou I and Seelig JD. Light scattering control with neural networks in transmission and reflection. *Arxiv:180505602 [Cs]*, 2018, <https://arxiv.org/abs/1805.05602>
60. Carrasco-Zevallos OM, Nankivil D, Viehland C, *et al.* Pupil tracking for real-time motion corrected anterior segment optical coherence tomography. *PLoS ONE* 2016; 11: e0162015.
61. Chen Y, Hong YJ, Makita S, *et al.* Eye-motion-corrected optical coherence tomography angiography using Lissajous scanning. *Biomed Opt Express* 2018; 9: 1111–1129.
62. Zhang K and Kang JU. Real-time 4d signal processing and visualization using graphics processing unit on a regular nonlinear-k Fourier-domain OCT system. *Opt Express* 2010; 18: 11772–11784.
63. Wieser W, Draxinger W, Klein T, *et al.* High definition live 3d-OCT in vivo: design and evaluation of a 4d OCT engine with 1 GVoxel/s. *Biomed Opt Express* 2014; 5: 2963–2977.
64. Eklund A, Dufort P, Forsberg D, *et al.* Medical image processing on the GPU – past, present and future. *Med Image Anal* 2013; 17: 1073–1094.
65. Klein T and Huber R. High-speed OCT light sources and systems. *Biomed Opt Express* 2017; 8: 828–859.
66. Li M, Idoughi R, Choudhury B, *et al.* Statistical model for OCT image denoising. *Biomed Opt Express* 2017; 8: 3903–3917.
67. Liu Y, Liang Y, Mu G, *et al.* Deconvolution methods for image deblurring in optical coherence tomography. *J Opt Soc Am A Opt Image Sci Vis* 2009; 26: 72–77.
68. Bian L, Suo J, Chen F, *et al.* Multi-frame denoising of high speed optical coherence tomography data using inter-frame and intra-frame priors. *Arxiv:13121931*, 2013, <https://arxiv.org/abs/1312.1931>
69. Devalla SK, Subramanian G, Pham TH, *et al.* A deep learning approach to denoise optical coherence tomography images of the optic nerve head. *Arxiv:180910589 [Cs]*, 2018, <http://arxiv.org/abs/1809.10589>
70. Köhler T, Brost A, Mogalle K, *et al.* Multi-frame super-resolution with quality self-assessment for retinal fundus videos. In: *Proceedings of the medical image computing and computer-assisted intervention – MICCAI 2014* (Lecture notes in computer science), Boston, MA, 14–18 September 2014, pp. 650–657. Cham: Springer.
71. Stankiewicz A, Marciniak T, Dabrowski A, *et al.* Matching 3D OCT retina images into superresolution dataset. In: *Proceedings of the 2016 signal processing: algorithms, architectures, arrangements, and applications (SPA)*, Poznan, 21–23 September 2016, pp. 130–137. New York: IEEE.
72. Xu L, Lu C, Xu Y, *et al.* Image smoothing via L0 gradient minimization. In: *Proceedings of the 2011 SIGGRAPH Asia conference – SA '11*, Hong Kong, China, 12–15 December 2015, pp. 174:1–174:12. New York: ACM.
73. Innamorati C, Ritschel T, Weyrich T, *et al.* Decomposing single images for layered photo retouching. *Comput Graph Forum* 2017; 36: 15–25.
74. Balakrishnan G, Zhao A, Sabuncu MR, *et al.* An unsupervised learning model for deformable medical image registration. *Arxiv:180202604 [Cs]*, 2018, <http://arxiv.org/abs/1802.02604>
75. Diamond S, Sitzmann V, Boyd S, *et al.* Dirty pixels: optimizing image classification architectures for raw sensor data. *Arxiv:170106487 [Cs]*, 2017, <http://arxiv.org/abs/1701.06487>
76. Liu D, Wen B, Liu X, *et al.* When image denoising meets high-level vision tasks: a deep learning approach. *Arxiv:170604284 [Cs]*, 2017, <http://arxiv.org/abs/1706.04284>
77. Schwartz E, Giryes R and Bronstein AM. DeepISP: learning end-to-end image processing pipeline. *Arxiv:180106724 [Cs, Eess]*, 2018, <http://arxiv.org/abs/1801.06724>
78. Sitzmann V, Diamond S, Peng Y, *et al.* End-to-end optimization of optics and image processing for achromatic extended depth of field and super-resolution imaging. *ACM Trans Graph* 2018; 37: 1141–11413.
79. Plötz T and Roth S. Benchmarking denoising algorithms with real photographs. *Arxiv:170701313 [Cs]*, 2017, <http://arxiv.org/abs/1707.01313>
80. Burger H, Schuler C and Harmeling S. Image denoising: can plain neural networks compete with BM3d? In: *Proceedings of the 2012 IEEE conference on computer vision and pattern recognition (CVPR)*, Providence, RI, 16–21 June 2012, pp. 2392–2399. New York: IEEE.
81. Mayer MA, Borsdorf A, Wagner M, *et al.* Wavelet denoising of multiframe optical coherence tomography data. *Biomed Opt Express* 2012; 3: 572–589.
82. Köhler T, Batz M, Naderi F, *et al.* Bridging the simulated-to-real gap: benchmarking

- super-resolution on real data. *Arxiv:180906420 [Cs]*, 2018, <http://arxiv.org/abs/1809.06420>
83. Tao X, Gao H, Liao R, et al. Detail-revealing deep video super-resolution. *Arxiv:170402738 [Cs]*, 2017, <http://arxiv.org/abs/1704.02738>
 84. Wang W, Ren C, He X, et al. Video super-resolution via residual learning. *IEEE Access* 2018; 6: 23767–23777.
 85. Marrugo AG, Millán MS, Šorel M, et al. Improving the blind restoration of retinal images by means of point-spread-function estimation assessment. In: *Proceedings of the 10th international symposium on medical information processing and analysis*, vol. 9287. International Society for Optics and Photonics, <https://spie.org/Publications/Proceedings/Paper/10.1117/12.2073820>
 86. Lian J, Zheng Y, Jiao W, et al. Deblurring sequential ocular images from multi-spectral imaging (MSI) via mutual information. *Med Biol Eng Comput* 2018; 56: 1107–1113.
 87. Christaras D, Ginis H, Pennos A, et al. Intraocular scattering compensation in retinal imaging. *Biomed Opt Express* 2016; 7: 3996–4006.
 88. Burns SA, Elsner AE, Sapoznik KA, et al. Adaptive optics imaging of the human retina. *Prog Retin Eye Res*. Epub ahead of print 27 August 2018. DOI: 10.1016/j.preteyeres.2018.08.002.
 89. KyongHwan Jin, McCann MT, Froustey E, et al. Deep convolutional neural network for inverse problems in imaging. *IEEE Trans Image Process* 2017; 26: 4509–4522.
 90. Lee B, Choi W, Liu JJ, et al. Cardiac-gated en face Doppler measurement of retinal blood flow using swept-source optical coherence tomography at 100,000 axial scans per second. *Invest Ophthalmol Vis Sci* 2015; 56: 2522–2530.
 91. Lee CY, Xie S, Gallagher P, et al. Deeply-supervised nets. *Arxiv:14095185 [Cs, Stat]*, 2014, <http://arxiv.org/abs/1409.5185>
 92. Bollepalli SC, Challa SS and Jana S. Robust heartbeat detection from multimodal data via CNN-based generalizable information fusion. *IEEE Trans Biomed Eng*. Epub ahead of print 11 July 2018. DOI: 10.1109/TBME.2018.2854899.
 93. Li S, Deng M, Lee J, et al. Imaging through glass diffusers using densely connected convolutional networks. *Optica* 2018; 5: 803–813.
 94. Fei X, Zhao J, Zhao H, et al. Deblurring adaptive optics retinal images using deep convolutional neural networks. *Biomed Opt Express* 2017; 8: 5675–5687.
 95. Jian Y, Lee S, Ju MJ, et al. Lens-based wavefront sensorless adaptive optics swept source OCT. *Sci Rep* 2016; 6: 27620.
 96. Carpentras D and Moser C. See-through ophthalmoscope for retinal imaging. *J Biomed Opt* 2017; 22: 56006.
 97. DuBose T, Nankivil D, LaRocca F, et al. Handheld adaptive optics scanning laser ophthalmoscope. *Optica* 2018; 5: 1027–1036.
 98. Chou JC, Cousins CC, Miller JB, et al. Fundus densitometry findings suggest optic disc hemorrhages in primary open-angle glaucoma have an arterial origin. *Am J Ophthalmol* 2018; 187: 108–116.
 99. Johnson CA, Nelson-Quigg JM and Morse LS. Wavelength dependent lens transmission properties in diabetics and non-diabetics. In: *Proceedings of the Basic and clinical applications of vision science*, 1997, pp. 217–220, Dordrecht: Springer, <https://www.springer.com/in/book/9780792343486>
 100. Zhang L, Deshpande A and Chen X. Denoising vs. deblurring: HDR imaging techniques using moving cameras. In: *Proceedings of the 2010 IEEE computer society conference on computer vision and pattern recognition*, San Francisco, CA, 13–18 June 2010, pp. 522–529. New York: IEEE.
 101. Ling WA and Ellerbee AK. The effects of reduced bit depth on optical coherence tomography phase data. *Opt Express* 2012; 20: 15654–15668.
 102. Ittarat M, Itthipanichpong R, Manassakorn A, et al. Capability of ophthalmology residents to detect glaucoma using high-dynamic-range concept versus color optic disc photography. *J Ophthalmol* 2017; 2017: 8209270.
 103. Yamashita H, Sugimura D and Hamamoto T. RGB-NIR imaging with exposure bracketing for joint denoising and deblurring of low-light color images. In: *Proceedings of the 2017 IEEE international conference on acoustics, speech and signal processing (ICASSP)*, New Orleans, LA, 5–9 March 2011, pp. 6055–6059. New York: IEEE.
 104. Hernandez-Matas C, Zabulis X, Triantafyllou A, et al. FIRE: fundus image registration dataset. *J Model Ophthalmol* 2017; 1: 16–28.
 105. Xia W and Tao L. Million-pixel computational imaging model. In: *Proceedings of the 2018 25th IEEE international conference on image processing (ICIP)*, Athens, 7–10 October 2018, pp. 425–429. New York: IEEE.

106. Bartczak P, Fält P, Penttinen N, *et al.* Spectrally optimal illuminations for diabetic retinopathy detection in retinal imaging. *Optical Rev* 2017; 24: 105–116.
107. Li H, Liu W, Dong B, *et al.* Snapshot hyperspectral retinal imaging using compact spectral resolving detector array. *J Biophotonics* 2017; 10: 830–839.
108. Ruia S and Saxena S. Spectral domain optical coherence tomography-based imaging biomarkers and hyperspectral imaging. In: Meyer CH, Saxena S and Sadda SR (eds) *Spectral domain optical coherence tomography in macular diseases*. New Delhi, India: Springer, 2017, pp. 109–114.
109. Kaluzny J, Li H, Liu W, *et al.* Bayer filter snapshot hyperspectral fundus camera for human retinal imaging. *Curr Eye Res* 2017; 42: 629–635.
110. Myers JS, Fudenberg SJ and Lee D. Evolution of optic nerve photography for glaucoma screening: a review. *Clin Exp Ophthalmol* 2018; 46: 169–176.
111. Palmer DW, Coppin T, Rana K, *et al.* Glare-free retinal imaging using a portable light field fundus camera. *Biomed Opt Express* 2018; 9: 3178–3192.
112. Rivenson Y, Göröcs Z, Günaydin H, *et al.* Deep learning microscopy. *Optica* 2017; 4: 1437–1443.
113. Cua M, Lee S, Miao D, *et al.* Retinal optical coherence tomography at 1 m with dynamic focus control and axial motion tracking. *J Biomed Optics* 2016; 21: 026007.
114. Fang L, Li S, Cunefare D, *et al.* Segmentation based sparse reconstruction of optical coherence tomography images. *IEEE Trans Med Imaging* 2017; 36: 407–421.
115. Schlemper J, Caballero J, Hajnal JV, *et al.* A deep cascade of convolutional neural networks for dynamic MR image reconstruction. *IEEE Trans Med Imaging* 2018; 37: 491–503.
116. Ju MJ, Heisler M, Athwal A, *et al.* Effective bidirectional scanning pattern for optical coherence tomography angiography. *Biomed Opt Express* 2018; 9: 2336–2350.
117. Köhler T, Budai A, Kraus MF, *et al.* Automatic no-reference quality assessment for retinal fundus images using vessel segmentation. In: *Proceedings of the 2013 IEEE 26th international symposium on computer-based medical systems (CBMS)*, Porto, 20–22 June 2013, pp. 95–100. New York: IEEE.
118. Bendaoudi H, Cheriet F, Manraj A, *et al.* Flexible architectures for retinal blood vessel segmentation in high-resolution fundus images. *J Real-Time Image Process* 2018; 15: 31–42.
119. Saha SK, Fernando B, Cuadros J, *et al.* Automated quality assessment of colour fundus images for diabetic retinopathy screening in telemedicine. *J Digit Imaging* 2018; 31: 869–878.
120. Zhao H, Gallo O, Frosio I, *et al.* Loss functions for image restoration with neural networks. *IEEE T Comput Imag* 2017; 3: 47–57.
121. Wang Z, Simoncelli EP and Bovik AC. Multiscale structural similarity for image quality assessment. In: *Proceedings of the 37th Asilomar conference on signals, systems computers, 2003*, vol. 2, Pacific Grove, CA, 9–12 November 2003, pp. 1398–1402. New York: IEEE.
122. Liba O, Lew MD, SoRelle ED, *et al.* Speckle-modulating optical coherence tomography in living mice and humans. *Nat Commun* 2017; 8: 16131.
123. Li Y, Xue Y and Tian L. Deep speckle correlation: a deep learning approach toward scalable imaging through scattering media. *Optica* 2018; 5: 1181–1190.
124. Chui TYP, Vannasdale DA and Burns SA. The use of forward scatter to improve retinal vascular imaging with an adaptive optics scanning laser ophthalmoscope. *Biomed Opt Express* 2012; 3: 2537–2549.
125. Zhang P, Manna SK, Miller EB, *et al.* Aperture phase modulation with adaptive optics: a novel approach for speckle reduction and structure extraction in optical coherence tomography. *bioRxiv* 2018: 406108.
126. Ling Y, Yao X and Hendon CP. Highly phase-stable 200 kHz swept-source optical coherence tomography based on KTN electro-optic deflector. *Biomed Opt Express* 2017; 8: 3687–3699.
127. Liu Z, Tam J, Saeedi O, *et al.* Trans-retinal cellular imaging with multimodal adaptive optics. *Biomed Opt Express* 2018; 9: 4246–4262.
128. Leitgeb RA, Werkmeister RM, Blatter C, *et al.* Doppler optical coherence tomography. *Prog Retina Eye Res* 2014; 41: 26–43.
129. Dadkhah A, Zhou J, Yeasmin N, *et al.* A multimodal imaging platform with integrated simultaneous photoacoustic microscopy, optical coherence tomography, optical Doppler tomography and fluorescence microscopy. In: *Proceedings of the photons plus ultrasound: imaging and sensing 2018*, vol. 10494. International Society for Optics and Photonics, <https://www.spiedigitallibrary.org/>

- conference-proceedings-of-spie/10494/104940Z/A-multimodal-imaging-platform-with-integrated-simultaneous-photoacoustic-microscopy-optical/10.1117/12.2289211.short?SSO=1
130. Emami H, Dong M, Nejad-Davarani SP, *et al.* Generating synthetic CTs from magnetic resonance images using generative adversarial networks. *Med Phys* 2018; 45: 3627–3636.
 131. Gharbi M, Chen J, Barron JT, *et al.* Deep bilateral learning for real-time image enhancement. *ACM Trans Graph* 2017; 36: 1181–11812.
 132. Kendall A and Gal Y. What uncertainties do we need in Bayesian deep learning for computer vision? In: Guyon I, Luxburg UV, Bengio S, *et al.* (eds) *Advances in neural information processing systems 30*. Red Hook, NY: Curran Associates, Inc., 2017, pp. 5574–5584.
 133. Lundell J, Verdoja F and Kyrki V. Deep network uncertainty maps for indoor navigation. *Arxiv:180904891 [Cs, Eess]*, 2018, <http://arxiv.org/abs/1809.04891>
 134. Leibig C, Allken V, Ayhan MS, *et al.* Leveraging uncertainty information from deep neural networks for disease detection. *Sci Rep* 2017; 7: 17816.
 135. Tanno R, Worrall DE, Ghosh A, *et al.* Bayesian image quality transfer with CNNs: exploring uncertainty in dMRI super-resolution. *Arxiv:170500664 [Cs]*, 2017, <http://arxiv.org/abs/1705.00664>
 136. Eaton-Rosen Z, Bragman F, Bisdas S, *et al.* Towards safe deep learning: accurately quantifying biomarker uncertainty in neural network predictions. *Arxiv:180608640 [Cs]*, 2018, <http://arxiv.org/abs/1806.08640>
 137. Cobb AD, Roberts SJ and Gal Y. Loss-calibrated approximate inference in Bayesian neural networks. *Arxiv:180503901 [Cs, Stat]*, 2018, <http://arxiv.org/abs/1805.03901>
 138. Yuan Y, Su W and Zhu M. Threshold-free measures for assessing the performance of medical screening tests. *Front Public Health* 2015; 3: 57.
 139. Boodhna T and Crabb DP. More frequent, more costly? Health economic modelling aspects of monitoring glaucoma patients in England. *BMC Health Serv Res* 2016; 16: 611.
 140. Li H and Zhang L. Multi-exposure fusion with CNN features. In: *Proceedings of the 2018 25th IEEE international conference on image processing (ICIP)*, Athens, 7–10 October, pp. 1723–1727. New York: IEEE.
 141. Hepp B, Nießner M and Hilliges O. Plan3D: viewpoint and trajectory optimization for aerial multi-view stereo reconstruction. *Arxiv:170509314 [Cs]*, 2017, <http://arxiv.org/abs/1705.09314>
 142. Chen M, Li W, Hao Y, *et al.* Edge cognitive computing based smart healthcare system. *Future Gene Comput Syst* 2018; 86: 403–411.
 143. Davoudi A, Malhotra KR, Shickel B, *et al.* The intelligent ICU pilot study: using artificial intelligence technology for autonomous patient monitoring. *Arxiv:180410201 [Cs, Eess]*, 2018, <http://arxiv.org/abs/1804.10201>
 144. Mandel JC, Kreda DA, Mandl KD, *et al.* SMART on FHIR: a standards-based, interoperable apps platform for electronic health records. *J Am Med Inform Assoc* 2016; 23: 899–908.
 145. Li H, Ota K and Dong M. Learning IoT in edge: deep learning for the internet of things with edge computing. *IEEE Network* 2018; 32: 96–101.
 146. Chang CK and Oyama K. Guest editorial: a roadmap for mobile and cloud services for digital health. *IEEE T Serv Comput* 2018; 11: 232–235.
 147. Bittman T. The edge will eat the cloud, 2017, https://blogs.gartner.com/thomas_bittman/2017/03/06/the-edge-will-eat-the-cloud/
 148. Grand View Research, Inc. Edge computing market size, share & trends analysis report by technology (mobile edge computing, fog computing), by vertical, by organization size, by region, and segment forecasts, 2018–2025, 2018, <https://www.grandviewresearch.com/industry-analysis/edge-computing-market>
 149. Wang Z, Yang Z and Dong T. A review of wearable technologies for elderly care that can accurately track indoor position, recognize physical activities and monitor vital signs in real time. *Sensors* 2017; 17: 341.
 150. NIH. All of us research program, 2018, <https://allofus.nih.gov/>
 151. Barik RK, Priyadarshini R, Dubey H, *et al.* Leveraging machine learning in mist computing telemonitoring system for diabetes prediction. In: Kolhe ML, Trivedi MC, Tiwari S, *et al.* (eds) *Advances in data and information sciences (Lecture notes in networks and systems)*. Singapore: Springer, 2018, pp. 95–104.
 152. Yousefpour A, Fung C, Nguyen T, *et al.* All one needs to know about fog computing and related edge computing paradigms: a complete survey. *Arxiv:180805283 [Csmi]*, 2018, <https://arxiv.org/abs/1808.05283>

153. Chen Z, Lin W, Wang S, *et al.* Intermediate deep feature compression: the next battlefield of intelligent sensing. *Arxiv:180906196 [Cs]*, 2018, <http://arxiv.org/abs/1809.06196>
154. Szydlo T, Sendorek J and Brzoza-Woch R. Enabling machine learning on resource constrained devices by source code generation of the learned models. In: Shi Y, Fu H, Tian Y, *et al.* (eds) *Computational science – ICCS 2018 (Lecture notes in computer science)*. New York: Springer, 2018, pp. 682–694.
155. Nweke HF, Teh YW, Al-garadi MA, *et al.* Deep learning algorithms for human activity recognition using mobile and wearable sensor networks: state of the art and research challenges. *Exp Syst Appl* 2018; 105: 233–261.
156. Yang G, Deng J, Pang G, *et al.* An IoT-enabled stroke rehabilitation system based on smart wearable armband and machine learning. *IEEE J Transl Eng Health Med* 2018; 6: 2100510.
157. Sumi H, Takehara H, Miyazaki S, *et al.* Next-generation Fundus Camera with Full Color Image Acquisition in 0-lx Visible Light by 1.12-micron Square Pixel, 4K, 30-fps BSI CMOS Image Sensor with Advanced NIR Multi-spectral Imaging System, *IEEE Symposium on VLSI Technology*, Honolulu, HI: IEEE; 2018, pp. 163–164. doi: 10.1109/VLSIT.2018.8510698
158. Aggarwal K, Joty S, Luque LF, *et al.* Co-morbidity exploration on wearables activity data using unsupervised pre-training and multi-task learning. *Arxiv: 1712.09527[cs]*, 2017, <http://arxiv.org/abs/1712.09527>
159. Yeung S, Downing NL, Fei-Fei L, *et al.* Bedside computer vision – moving artificial intelligence from driver assistance to patient safety. *N Engl J Med* 2018; 378: 1271–1273.
160. Dubey H, Monteiro A, Constant N, *et al.* Fog computing in medical internet-of-things: architecture, implementation, and applications. In: Khan SU, Zomaya AY and Abbas A (eds) *Handbook of large-scale distributed computing in smart healthcare: scalable computing and communications*. Cham: Springer, 2017, pp. 281–321.
161. Raut A, Yarbrough C, Singh V, *et al.* Design and implementation of an affordable, public sector electronic medical record in rural Nepal. *J Innov Health Inform* 2017; 24: 862.
162. Sahu P, Yu D and Qin H. Apply lightweight deep learning on internet of things for low-cost and easy-to-access skin cancer detection. In: *Proceedings of the medical imaging 2018: imaging informatics for healthcare, research, and applications*, Vol. 10579. International Society for Optics and Photonics, <https://www.spiedigitallibrary.org/conference-proceedings-of-spie/10579/1057912/Apply-lightweight-deep-learning-on-internet-of-things-for-low/10.1117/12.2293350.short>
163. Rippel O and Bourdev L. Real-time adaptive image compression. *Arxiv:170505823 [Cs, Stat]*, 2017, <http://arxiv.org/abs/1705.05823>
164. HajiRassouliha A, Taberner AJ, Nash MP, *et al.* Suitability of recent hardware accelerators (DSPs, FPGAs, and GPUs) for computer vision and image processing algorithms. *Sig Proc Imag Commun* 2018; 68: 101–119.
165. Fey D and Hannig F. Special issue on heterogeneous real-time image processing. *J Real-time Imag Process* 2018; 14: 513–515.
166. Pagnutti MA, Ryan RE, Cazenavette GJ, *et al.* Laying the foundation to use Raspberry Pi 3 V2 camera module imagery for scientific and engineering purposes. *J Electron Imag* 2017; 26: 013014.
167. Shen BY and Mukai S. A portable, inexpensive, nonmydriatic fundus camera based on the raspberry PiR computer. *J Ophthalmol* 2017; 2017: 4526243, <https://doi.org/10.1155/2017/4526243>
168. Pérez J, Rodríguez A, Chico JF, *et al.* Energy-aware acceleration on GPUs: findings on a bioinformatics benchmark. *Sustain Comput Inform Syst* 2018; 20: 88–101.
169. Zhao R, Ng HC, Luk W, *et al.* Towards efficient convolutional neural network for domain-specific applications on FPGA. *Arxiv:180903318 [Cs]*, 2018, <http://arxiv.org/abs/1809.03318>
170. Bendaoudi H. *Flexible hardware architectures for retinal image analysis*. PhD Thesis, École Polytechnique de Montréal, <https://publications.polymtl.ca/2518/>
171. Hung KW, Qiu C and Jiang J. Video restoration using convolution neural networks for low-level FPGAs. In Liu W, Giunchiglia F and Yang B (eds) *Knowledge science, engineering and management (Lecture notes in computer science)*. New York: Springer, 2018, pp. 255–265.
172. Kulkarni A, Page A, Attaran N, *et al.* An energy-efficient programmable manycore accelerator for personalized biomedical applications. *IEEE T VLSI Syst* 2018; 26: 96–109.
173. Jouppi NP, Young C, Patil N, *et al.* In-datacenter performance analysis of a tensor processing unit. *Arxiv:170404760 [Cs]*, 2017, <http://arxiv.org/abs/1704.04760>
174. Vitish-Sharma P, Acheson AG, Stead R, *et al.* Can the Sensimed Triggerfish lens data be used

- as an accurate measure of intraocular pressure? *Acta Ophthalmologica* 2018; 96: e242–e246.
175. Araci IE, Su B, Quake SR, *et al.* An implantable microfluidic device for self-monitoring of intraocular pressure. *Nat Med* 2014; 20: 1074–1078.
176. Molaei A, Karamzadeh V, Safi S, *et al.* Upcoming methods and specifications of continuous intraocular pressure monitoring systems for glaucoma. *J Ophthalmic Vis Res* 2018; 13: 66–71.
177. Najjar RP, Sharma S, Drouet M, *et al.* Disrupted eye movements in preperimetric primary open-angle glaucoma. *Invest Ophthalmol Vis Sci* 2017; 58: 2430–2437.
178. Asfaw DS, Jones PR, Monter VM, *et al.* Does glaucoma alter eye movements when viewing images of natural scenes? A between-eye study. *Invest Ophthalmol Vis Sci* 2018; 59: 3189–3198.
179. Najjar RP, Sharma S, Atalay E, *et al.* Pupillary responses to full-field chromatic stimuli are reduced in patients with early-stage primary open-angle glaucoma. *Ophthalmology* 2018; 125: 1362–1371.
180. Zhu Y, Zuo Y, Zhou T, *et al.* A multi-mode visual recognition hardware accelerator for AR/MR glasses. In: *Proceedings of the 2018 IEEE international symposium on circuits and systems (ISCAS)*, Florence, 27–30 May 2018, pp. 1–5. New York: IEEE.
181. Sarkar N. System and method for resonant eye-tracking, 2018, <https://patents.google.com/patent/US20180210547A1/en>
182. Ping P, Hermjakob H, Polson JS, *et al.* Biomedical informatics on the cloud. *Circ Res* 2018; 122: 1290–1301.
183. Muhammed T, Mehmood R, Albeshri A, *et al.* UbeHealth: a personalized ubiquitous cloud and edge-enabled networked healthcare system for smart cities. *IEEE Access* 2018; 6: 32258–32285.
184. Kotecha A, Brookes J and Foster PJ. A technician-delivered ‘virtual clinic’ for triaging low-risk glaucoma referrals. *Eye* 2017; 31: 899–905.
185. Caffery LJ, Taylor M, Gole G, *et al.* Models of care in tele-ophthalmology: a scoping review. *J Telemed Telecare*. Epub ahead of print 1 January 2017. DOI: 10.1177/1357633X17742182.
186. Hong S, Xiao C, Ma T, *et al.* RDPD: rich data helps poor data via imitation. *Arxiv:180901921 [Cs, Stat]*, 2018, <http://arxiv.org/abs/1809.01921>
187. Verghese A. How tech can turn doctors into clerical workers. *The New York Times*, 2018, <https://goo.gl/6LBm27>
188. Lerner I, Veil R, Nguyen DP, *et al.* Revolution in health care: how will data science impact doctor–patient relationships. *Front Public Health* 2018; 6: 99.
189. Rosenberg L, Willcox G, Halabi S, *et al.* Artificial swarm intelligence employed to amplify diagnostic accuracy in radiology. In: *Proceedings of the EMCON 2018*, Vancouver, BC, Canada: IEEE; p. 6.
190. Kilkenny MF and Robinson KM. Data quality: ‘garbage in – garbage out’. *Health Inf Manag* 2018; 47: 103–105.
191. Feldman M, Even A and Parmet Y. A methodology for quantifying the effect of missing data on decision quality in classification problems. *Commun Stat Theory Methods* 2018; 47: 2643–2663.
192. Shickel B, Tighe PJ, Bihorac A, *et al.* Deep EHR: a survey of recent advances in deep learning techniques for electronic health record (EHR) analysis. *IEEE J Biomed Health Inform* 2018; 22: 1589–1604.
193. Eisenberg RS. Shifting institutional roles in biomedical innovation in a learning healthcare system. *J Inst Econ* 2018: 1–24.
194. Thornton T. Tacit knowledge as the unifying factor in evidence based medicine and clinical judgement. *Philos Ethics Humanit Med* 2006; 1: E2.
195. Keane PA and Topol EJ. With an eye to AI and autonomous diagnosis. *Npj Digital Medicine* 2018; 1: 40.
196. DePasse JW, Carroll R, Ippolito A, *et al.* Less noise, more hacking: how to deploy principles from MIT’s hacking medicine to accelerate health care. *Int J Technol Assess Health Care* 2014; 30: 260–264.
197. Borsci S, Uchegbu I, Buckle P, *et al.* Designing medical technology for resilience: integrating health economics and human factors approaches. *Expert Rev Med Devices* 2018; 15: 15–26.