

SHORT REPORT

The global Alzheimer's Association round robin study on plasma amyloid β methods

Josef Pannee^{1,2} | Leslie M. Shaw³ | Magdalena Korecka³ | Teresa Waligorska³ | Charlotte E. Teunissen⁴ | Erik Stoops⁵ | Hugo M. J. Vanderstichele⁶ | Kimberley Mauroo⁵ | Inge M. W. Verberk⁴ | Ashvini Keshavan⁷ | Pedro Pesini⁸ | Leticia Sarasa⁸ | Maria Pascual-Lucas⁸ | Noelia Fandos⁸ | José-Antonio Allué⁸ | Erik Portelius^{1,2} | Ulf Andreasson^{1,2} | Ritsuko Yoda⁹ | Akinori Nakamura¹⁰ | Naoki Kaneko⁹ | Shieh-Yueh Yang¹¹ | Huei-Chun Liu¹¹ | Stefan Palme¹² | Tobias Bittner¹³ | Kwasi G. Mawuenyega¹⁴ | Vitaliy Ovod¹⁴ | James Bollinger¹⁴ | Randall J. Bateman¹⁴ | Yan Li¹⁴ | Jeffrey L. Dage¹⁵ | Erik Stomrud^{16,17} | Oskar Hansson^{16,17} | Jonathan M. Schott⁷ | Kaj Blennow^{1,2} | Henrik Zetterberg^{1,2,18,19}

¹ Department of Psychiatry and Neurochemistry, Institute of Neuroscience and Physiology, The Sahlgrenska Academy, University of Gothenburg, Mölndal, Sweden

² Clinical Neurochemistry Lab, Sahlgrenska University Hospital, Mölndal, Sweden

³ Perelman School of Medicine, Department of Pathology and Laboratory Medicine, University of Pennsylvania, Philadelphia, Pennsylvania, USA

⁴ Neurochemistry Laboratory, Department of Clinical Chemistry, Amsterdam Neuroscience, Amsterdam UMC, Vrije Universiteit Amsterdam, Amsterdam, the Netherlands

⁵ ADx NeuroSciences, Ghent, Belgium

⁶ Biomarkable, Ghent, Belgium

⁷ Dementia Research Centre, UCL Queen Square Institute of Neurology, London, UK

⁸ Araclon Biotech, Zaragoza, Spain

⁹ Koichi Tanaka Mass Spectrometry Research Laboratory, Shimadzu Corporation, Kyoto, Japan

¹⁰ Department of Biomarker Research, National Center for Geriatrics and Gerontology, Obu, Aichi, Japan

¹¹ MagQu Co., Ltd., New Taipei City, Taiwan

¹² Roche Diagnostics GmbH, Penzberg, Germany

¹³ F. Hoffmann-La Roche Ltd, Basel, Switzerland

¹⁴ Department of Neurology, Washington University School of Medicine, St. Louis, Missouri, USA

¹⁵ Eli Lilly and Company, Indianapolis, Indiana, USA

¹⁶ Clinical Memory Research Unit, Faculty of Medicine, Lund University, Lund, Sweden

¹⁷ Memory Clinic, Skåne University Hospital, Malmö, Sweden

¹⁸ Department of Neurodegenerative Disease, UCL Queen Square Institute of Neurology, London, UK

¹⁹ UK Dementia Research Institute, London, UK

Correspondence

Josef Pannee, Clinical Neurochemistry Lab, Mölndal Hospital, Biskopsbogatan 27, Building V3, floor 2, SE-43180 Mölndal, Sweden.
E-mail: josef.pannee@gu.se

Abstract

Introduction: Blood-based assays to measure brain amyloid beta ($A\beta$) deposition are an attractive alternative to the cerebrospinal fluid (CSF)-based assays currently used in clinical settings. In this study, we examined different blood-based assays to measure $A\beta$ and how they compare among centers and assays.

This is an open access article under the terms of the [Creative Commons Attribution-NonCommercial-NoDerivs](https://creativecommons.org/licenses/by-nc-nd/4.0/) License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

© 2021 The Authors. *Alzheimer's & Dementia: Diagnosis, Assessment & Disease Monitoring* published by Wiley Periodicals, LLC on behalf of Alzheimer's Association

Methods: Aliquots from 81 plasma samples were distributed to 10 participating centers. Seven immunological assays and four mass-spectrometric methods were used to measure plasma A β concentrations.

Results: Correlations were weak for A β 42 while A β 40 correlations were stronger. The ratio A β 42/A β 40 did not improve the correlations and showed weak correlations.

Discussion: The poor correlations for A β 42 in plasma might have several potential explanations, such as the high levels of plasma proteins (compared to CSF), sensitivity to pre-analytical sample handling and specificity, and cross-reactivity of different antibodies. Different methods might also measure different pools of plasma A β 42. We, however, hypothesize that greater correlations might be seen in future studies because many of the methods have been refined during completion of this study.

KEYWORDS

Alzheimer's disease, amyloid beta, biomarkers, method comparison, plasma

1 | INTRODUCTION

In Alzheimer's disease (AD), amyloid beta (A β) deposition in the brain is detectable using the cerebrospinal fluid (CSF) biomarkers A β 42 or A β 42/40 ratio and by using amyloid positron emission tomography (PET).¹ Because CSF sampling is mainly performed at memory clinics and other specialized centers and amyloid PET is costly with limited availability, blood-based assays have long been an attractive alternative, especially in the primary care setting. The ability to reliably distinguish AD dementia from controls using A β in plasma has until 2016 showed poor performance and partially conflicting results.² However, newly developed highly sensitive immunoassays, as well as mass spectrometry (MS) methods, have shown a better and higher concordance of A β in plasma with A β -PET or CSF amyloid status.³⁻⁷

The aim of this study was to examine how different methods that measure plasma A β 42 and A β 40 levels compare, and whether results correlate linearly. Ten centers participated in this study, which included seven immunoassays and four MS methods, each analyzing aliquots of 81 unique ethylenediaminetetraacetic acid (EDTA)-plasma samples.

2 | METHODS

Individual de-identified EDTA-plasma samples ($n = 81$) were measured from the prospective and longitudinal Swedish BioFINDER (Biomarkers for Identifying Neurodegenerative Disorders Early and Reliably) cohort ($n = 48$); the prospective University College London Dementia Research Centre CSF cohort ($n = 24$); and the Clinical Neurochemistry Laboratory at the Sahlgrenska University Hospital, Mölndal, Sweden ($n = 9$). Varied sampling and processing procedures were used across these centers, and for this study the samples were prepared in 250 μ L aliquots, so each underwent one freeze-thaw cycle prior to distribution. These aliquots were kept at -80°C pending distribution to participating centers. The plasma samples were selected based on known

matched CSF A β 42 concentrations previously measured in the original cohorts, to theoretically include samples with a wide range of plasma A β levels. Across the 10 participating centers (Table 1), seven immunological assays and four MS methods were used in this study. All methods measured A β 40 and A β 42 but varied in whether the full-length A β 1-40 and A β 1-42 forms were measured (for simplicity, the terms A β 40 and A β 42 are used throughout), and two methods also measured the APP₆₆₉₋₇₁₁ form. Methods were compared using Passing-Bablok regression⁸ and Spearman's rank correlation coefficient (r_s).

3 | RESULTS

The correlations for pair-wise method comparison (Figure 1) for A β 42 were generally weak to moderate with a median r_s value of 0.24 and highest r_s value of 0.72. The correlations for A β 40 were stronger with a median r_s value of 0.67 and highest r_s value of 0.89. Interestingly, using the ratio A β 42/A β 40 did not improve the correlations (Figure 2) and showed weak correlations (similar to A β 42) with a median r_s value of 0.25 and highest r_s value of 0.65. See [supporting information](#) for full correlation plots between all methods for A β 40, A β 42, and the A β 42/A β 40 ratio.

4 | DISCUSSION

The results in this multicenter study showed acceptable correlations for plasma A β 40, while there were poor correlations for plasma A β 42, as well as for the A β 42/A β 40 ratio. The moderate correlations between the MS assays support comparable measurements but correlations are not ideal (generally < 0.7).

The MagQu method, which uses one antibody to capture A β 40 and A β 42 and immunomagnetic reduction to quantify the protein, does not correlate with the other methods, thus it may measure other forms of

TABLE 1 Participating centers, assay platform, and measured A β species

Center	Technology platform	A β species	Capture antibody	Detection antibody	Calibrant
University of Pennsylvania Biomarker Research Lab	Simoa (commercial)	A β 42	H31L21 12F4	6E10	AnaSpec #24236 & #20276
		A β 40		6E10	
ADX/Euroimmun	ELISA	A β 1-42	21F12 2G3	3D6	rPeptide rPeptide
		A β 1-40		3D6	
Amsterdam UMC/ADX	Simoa (in-house) (Amyblood)	A β 1-42	21F12 2G3	3D6	rPeptide rPeptide
		A β 1-40		3D6	
Araclon Biotech	ELISA (ABtest)	A β 1-42	1F3 (Araclon Biotech)	pAB031 and pAB002 (Araclon Biotech)	Synthetic A β 1-40 and A β 1-42 peptides (Araclon Biotech)
		A β 1-40 and N-truncated species up to N3pE-42			
Araclon Biotech	LC-MS	A β 1-42 A β 1-40	None	None	rPeptide Uniformly labeled 15N, recombinant
Shimadzu Corporation	MALDI-TOF MS	A β 1-42	6E10 (BioLegend)	None	AnaSpec AnaSpec PEPTIDE INSTITUTE
		A β 1-40 APP669-711			
MagQu Taiwan	IMR	A β 42 A β 40	Abcam (ab34376) BAM-10	None	rPeptide
Roche Diagnostics	Elecsys	A β 1-42 A β 1-40	21F12 & 23C2	3D6	Synthetic peptide in artificial matrix
University of Gothenburg	IP-LC-MS	A β 1-42	Biolegend A β 17-24 (4G8) & 1-16 (6E10)	None	rPeptide Uniformly labeled 15N, recombinant
		A β 1-40 APP669-711			
Washington University	IP-LC-MS	A β 42 A β 40	HJ5.1	None	rPeptide Uniformly labeled 15N, recombinant
Eli Lilly	Simoa (in-house)	A β 1-42 A β 1-40	3D6	2G3-A β 1-40 21F12-A β 1-42	Eli Lilly reference standard

Abbreviations: A β , amyloid β ; ADNI, Alzheimer's Disease Neuroimaging Initiative; ELISA, enzyme-linked immunosorbent assay; IMR, ImmunoMagnetic Reduction; IP-LC-MS, immunoprecipitation (IP) coupled to liquid chromatography mass spectrometry (LC-MS); MALDI-TOF-MS, matrix-assisted laser desorption-ionization-time of flight mass spectrometry; Simoa, single molecule array.



FIGURE 1 Amyloid beta (Aβ)1-40 (top), Aβ1-42 (middle), and Aβ1-42/Aβ1-40 (bottom) correlations (Spearman) between the different centers and methods




FIGURE 2 Examples of amyloid beta (Aβ)1-42/Aβ1-40 correlation plots between different centers. The solid line represents the Passing-Bablok regression line and the dashed line denotes the unity line (y = x). See supporting information for complete set of plots for all centers

RESEARCH IN CONTEXT

1. Systematic review: The authors reviewed the literature using PubMed and conference presentations. While blood-based assays until recently have shown conflicting results in the ability to distinguish Alzheimer's disease from controls compared to cerebrospinal fluid (CSF) biomarker profiles (amyloid beta [$A\beta$] and tau) and amyloid positron emission tomography (PET), newly developed methods to measure $A\beta$ in plasma have shown results with improved diagnostic performance for specific applications. Citations directly relevant to the included assays and their contexts are cited.
2. Interpretation: The findings in this study show correlations among 11 methods that measured ethylenediaminetetraacetic acid–plasma $A\beta_{42}$ and $A\beta_{40}$. Further standardization, qualification, and validation work is needed to obtain a more harmonized outcome among detection methods.
3. Future directions: Since completion of this study, many of the methods have undergone additional refinement by the vendors. Future method comparison studies will show if this will result in higher correlations between the methods or improved clinical performance; if not, an in-depth analysis of method differences needs to be undertaken.

$A\beta$, which might explain the increased (not decreased) levels of $A\beta_{42}$ and $A\beta_{42}/A\beta_{40}$ ratio in plasma of AD patients compared to controls.⁹ Based on previous studies, this method may require special sample preparation procedures to obtain consistent results.¹⁰

There might be several potential explanations for the discrepancies between the measurements obtained by the different methods used in this study. First, plasma is a much more complex matrix compared to CSF, with very high levels of albumin, immunoglobulin G, and other plasma proteins (approximately 200 times higher in plasma than in CSF), and also lipoprotein particles containing apolipoprotein E (apoE) and other apolipoproteins that may form complexes with $A\beta$. This makes plasma a difficult matrix for $A\beta$ measurements. These proteins may block the binding of antibodies to their respective analytes in the assays. In contrast, CSF has a less complicated matrix, and round robin studies on CSF $A\beta_{42}$ and $A\beta_{40}$ show very tight correlations across different assays, with a median correlation coefficient of 0.98.¹¹ It is also possible that different methods measure different pools of plasma $A\beta_{42}$ but these may still show diagnostic utility as reported by different groups^{12,13} and, as exemplified also by the inverse correlations for the MagQu assay. Different methods might also be differentially sensitive to method-specific pre-analytical sample handling in the local analysis laboratories, which might have been different in the originating cohorts, but aliquots distributed to the different centers were identical in the present study. Specificity of the used antibodies, cross-reactivity

with other $A\beta$ isoforms, sample dilution before analysis, additives and pre-incubation procedures are other factors that might influence the sensitivities. In addition, $A\beta_{42}$ concentrations are still at or close to the lower limit of quantification of most methods in plasma samples, which also may explain the higher correlations between assays for the more abundant $A\beta_{40}$ compared to $A\beta_{42}$. Furthermore, several studies reported similar findings comparing enzyme-linked immunosorbent assays and Simoa platforms for plasma $A\beta_{40}$ and $A\beta_{42}$.^{7,14,15} Spearman coefficients were 0.68 and 0.71 for, respectively, $A\beta_{40}$ and $A\beta_{42}$, which corroborates the findings in this article for the same assays. Since completion of this study, many of the methods have undergone additional refinement and new method comparison studies are underway. We hypothesize that greater correlations will now be seen; if not, an in-depth analysis of method differences will need to be undertaken.

ACKNOWLEDGMENTS

SP thanks Lisa Protzer for her technical assistance.

CONFLICTS OF INTEREST

OH has acquired research support (for the institution) from Avid Radiopharmaceuticals, Biogen, Eli Lilly, Eisai, GE Healthcare, Pfizer, and Roche. In the past 2 years, he has received consultancy/speaker fees from AC Immune, Alzpath, Biogen, Cerveau, and Roche. SP is a full-time employee of Roche Diagnostics GmbH and holds shares in Roche. TB is a full-time employee of and owns stock in F. Hoffmann-La Roche Ltd. KB has served as a consultant, on advisory boards, or on data monitoring committees for Abcam, Axon, Biogen, JOMDD/Shimadzu, Julius Clinical, Lilly, MagQu, Novartis, Prothena, Roche Diagnostics, and Siemens Healthineers, and is a co-founder of Brain Biomarker Solutions in Gothenburg AB (BBS), which is a part of the GU Ventures Incubator Program. HZ has served on scientific advisory boards for Alektor, Eisai, Denali, Roche Diagnostics, Wave, Samumed, Siemens Healthineers, Pinteon Therapeutics, Nervgen, AZTherapies and CogRx. JLD is an employee and stockholder of Eli Lilly and Company. ELECSYS is a trademark of Roche. E Stoops and KM are full-time paid employees of ADx NeuroSciences. JMS has received research funding from Avid Radiopharmaceuticals (a wholly owned subsidiary of Eli Lilly and Company); has consulted for Roche Pharmaceuticals, Biogen, Merck, and Eli Lilly; given educational lectures sponsored by GE Healthcare, Eli Lilly, and Biogen; and serves on a Data Safety Monitoring Committee for Axon Neuroscience SE. RJB cofounded C2N Diagnostics. Washington University and Dr. Bateman have equity ownership interest in C2N Diagnostics and receive royalty income based on technology (stable isotope labeling kinetics and blood plasma assay) licensed by Washington University to C2N Diagnostics. He receives income from C2N Diagnostics for serving on the scientific advisory board. Washington University, with RJB as co-inventor, has submitted the US provisional patent application "Plasma Based Methods for Detecting CNS Amyloid Deposition." He has received consultant fees from Roche, C2N Diagnostics, Genentech, AbbVie, Pfizer, Boehringer-Ingelheim, Eisai, AC Immune, Janssen, and Merck. He serves as principal investigator of the DIAN-TU, which is supported by the Alzheimer's Association, GHR Foundation, Eisai, an anonymous organization, and the DIAN-TU

Pharma Consortium. HV is a founder of Biomarkable and a co-founder of ADx NeuroSciences. RY and NK are full-time employees of Shimadzu Corporation. NK holds stock in Shimadzu Corporation and has received payment for manuscript writing from Rinshohoushasen. CET has a collaboration contract with ADx Neurosciences and Quanterix; performed contract research or received grants from AC-Immune, Axon Neurosciences, Biogen, Brainstorm Therapeutics, Celgene, EIP Pharma, Eisai, PeopleBio, Roche, Toyama, Vivoryon; received honoraria from Medidact Neurologie. LMS has received honorarium from Biogen for teaching. LS and JA have submitted patents for "Methods for quantification of amyloid beta peptides in plasma by mass spectrometry." AN received honoraria from The Educational Program for Dementia Experts in Hokuriku (NINPRO), The Japan Society for the Promotion of Science (JSPS), Translational Research Center for Medical Innovation (TRI), Eisai Co. Ltd. SY is an employee and shareholder of MagQu Co., Ltd. KGM, VO, and JB have submitted patent application "Plasma Based Methods for Detecting CNS Amyloid Deposition" and may receive royalties based on blood plasma assay technology licensed to C2N Diagnostics. ES has received payment (to institution) from Roche Diagnostics for medical writing.

FUNDING INFORMATION

Work at Lund University was supported by the Swedish Research Council (2016-00906), the Knut and Alice Wallenberg foundation (2017-0383), the Marianne and Marcus Wallenberg foundation (2015.0125), the Swedish Alzheimer Foundation (AF-939932), the Swedish Brain Foundation (FO2019-0326), and the Swedish federal government under the ALF agreement (2018-Projekt0279). AK was supported by a Wolfson Clinical Research Fellowship and a Weston Brain Institute and Selfridges Group Foundation award (UB17005). JMS is supported by University College London Hospitals Biomedical Research Centre, Engineering and Physical Sciences Research Council (EP/J020990/1), British Heart Foundation (PG/17/90/33415), and EU's Horizon 2020 research and innovation program (666992), the National Institute for Health Research Queen Square Dementia Biomedical Research Unit and the Leonard Wolfson Experimental Neurology Centre. HZ is a Wallenberg Scholar supported by grants from the Swedish Research Council (#2018-02532), the European Research Council (#681712), Swedish State Support for Clinical Research (#ALFGBG-720931), the Alzheimer Drug Discovery Foundation (ADDF), USA (#201809-2016862), the AD Strategic Fund and the Alzheimer's Association (#ADSF-21-831376-C, #ADSF-21-831381-C and #ADSF-21-831377-C), the Olav Thon Foundation, the Erling-Persson Family Foundation, Stiftelsen för Gamla Tjänarinnor, Hjärfonden, Sweden (#FO2019-0228), the European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement No 860197 (MIRIADE), and the UK Dementia Research Institute at UCL. KB is supported by the Swedish Research Council (#2017-00915), the Swedish Alzheimer Foundation (#AF-742881), Hjärfonden, Sweden (#FO2017-0243), and the Swedish state under the agreement between the Swedish government and the County Councils, the ALF-agreement (#ALFGBG-715986). Work at Washington University in St. Louis was supported by institutional

gift funds (RJB) and the NIH National Institute on Aging grants R56AG061900 and RF1AG061900 (PI: RJB). RJB receives lab research funding from the National Institutes of Health, Alzheimer's Association, BrightFocus Foundation, Rainwater Foundation Tau Consortium, Association for Frontotemporal Degeneration, the Cure Alzheimer's Fund, Centene Corporation, Tau SILK Consortium (AbbVie, Biogen, and Eli Lilly and Company), and an anonymous organization. LMS receives support from the NIA for his work in the ADNI Biomarker Core including provision of QC oversight for analyses of biomarkers in biofluids; and for AD biomarker analyses within the UPenn ADRC; he has received research support from Michael J. Fox Foundation for Parkinson's Disease Research and IIS support from Roche. CET receives support by the European Commission (Marie Curie International Training Network, grant agreement No 860197 [MIRIADE], and JPND), Health Holland, the Dutch Research Council (ZonMW), Alzheimer Drug Discovery Foundation, The Selfridges Group Foundation, Alzheimer Netherlands, Alzheimer Association. CT is are recipients of ABOARD, which is a public-private partnership receiving funding from ZonMW (#73305095007) and Health~Holland, Topsector Life Sciences & Health (PPP-allowance; #LSHM20106). ABOARD also receives funding from Edwin Bouw Fonds and Gieskes-Strijbisfonds. IV is appointed on a research grant by Alzheimer Nederland (NL-17004). UA has received funding from Stiftelsen för Gamla Tjänarinnor. AN has received Research and Development Grants for Dementia from the Japan Agency for Medical Research and Development (AMED), payments were made to the National Center for Geriatrics and Gerontology.

REFERENCES

1. Cohen AD, Landau SM, Snitz BE, Klunk WE, Blennow K, Zetterberg H. Fluid and PET biomarkers for amyloid pathology in Alzheimer's disease. *Mol Cell Neurosci*. 2019;97:3-17.
2. Olsson B, Lautner R, Andreasson U, et al. CSF and blood biomarkers for the diagnosis of Alzheimer's disease: a systematic review and meta-analysis. *Lancet Neurol*. 2016;15:673-684.
3. Ovod V, Ramsey KN, Mawuenyega KG, et al. Amyloid beta concentrations and stable isotope labeling kinetics of human plasma specific to central nervous system amyloidosis. *Alzheimers Dement*. 2017;13:841-849.
4. Nakamura A, Kaneko N, Villemagne VL, et al. High performance plasma amyloid-beta biomarkers for Alzheimer's disease. *Nature*. 2018;554:249-254.
5. Palmqvist S, Janelidze S, Stomrud E, et al. Performance of fully automated plasma assays as screening tests for Alzheimer's disease-related beta-amyloid status. *JAMA Neurol*. 2019;76:1060-1069.
6. Schindler SE, Bollinger JG, Ovod V, et al. High-precision plasma beta-amyloid 42/40 predicts current and future brain amyloidosis. *Neurology*. 2019;93:e1647.
7. Keshavan A, Pannee J, Karikari TK, et al. Population-based blood screening for preclinical Alzheimer's disease in a British birth cohort at age 70. *Brain*. 2021;144:434-449.
8. Passing H, Bablok A. A new biometrical procedure for testing the equality of measurements from two different analytical methods. Application of linear regression procedures for method comparison studies in clinical chemistry, Part I. *J Clin Chem Clin Biochem*. 1983;21:709-720.
9. Lue LF, Kuo YM, Sabbagh M. Advance in plasma AD core biomarker development: current findings from immunomagnetic reduction-based SQUID technology. *Neurol Ther*. 2019;8:95-111.

10. Tsai CL, Liang CS, Lee JT, et al. Associations between plasma biomarkers and cognition in patients with Alzheimer's disease and amnesic mild cognitive impairment: a cross-sectional and longitudinal study. *J Clin Med*. 2019;8(11):1893.
11. Andreasson U, Kuhlmann J, Pannee J, et al. Commutability of the certified reference materials for the standardization of beta-amyloid 1-42 assay in human cerebrospinal fluid: lessons for tau and beta-amyloid 1-40 measurements. *Clin Chem Lab Med*. 2018;56:2058-2066.
12. Doecke JD, Perez-Grijalba V, Fandos N, et al. Total Abeta42/Abeta40 ratio in plasma predicts amyloid-PET status, independent of clinical AD diagnosis. *Neurology*. 2020;94:e1580-e91.
13. Perez-Grijalba V, Romero J, Pesini P, et al. Plasma Abeta42/40 ratio detects early stages of Alzheimer's disease and correlates with CSF and neuroimaging biomarkers in the AB255 study. *J Prev Alzheimers Dis*. 2019;6:34-41.
14. De Meyer S, Schaevebeke JM, Verberk IMW, et al. Comparison of ELISA- and SIMOA-based quantification of plasma Abeta ratios for early detection of cerebral amyloidosis. *Alzheimers Res Ther*. 2020;12:162.
15. Thijssen EH, Verberk IMW, Vanbrabant J, et al. Highly specific and ultrasensitive plasma test detects Abeta(1-42) and Abeta(1-40) in Alzheimer's disease. *Sci Rep*. 2021;11:9736.

SUPPORTING INFORMATION

Additional supporting information may be found in the online version of the article at the publisher's website.

How to cite this article: Pannee J, Shaw LM, Korecka M, et al.

The global Alzheimer's Association round robin study on plasma amyloid β methods. *Alzheimer's Dement*.

2021;13:e12242. <https://doi.org/10.1002/dad2.12242>