



A patient-specific multi-modality abdominal aortic aneurysm imaging phantom

Callum D. Little^{1,2,3} · Eleanor C. Mackle^{1,2} · Efthymios Maneas^{1,2} · Debra Chong^{1,4} · Daniil Nikitichev¹ · Jason Constantinou⁴ · Janice Tsui^{1,4} · George Hamilton^{1,4} · Roby D. Rakhit³ · Tara M. Mastracci⁵ · Adrien E. Desjardins^{1,2}

Received: 10 January 2022 / Accepted: 15 March 2022
© The Author(s) 2022

Abstract

Purpose Multimodality imaging of the vascular system is a rapidly growing area of innovation and research, which is increasing with awareness of the dangers of ionizing radiation. Phantom models that are applicable across multiple imaging modalities facilitate testing and comparisons in pre-clinical studies of new devices. Additionally, phantom models are of benefit to surgical trainees for gaining experience with new techniques. We propose a temperature-stable, high-fidelity method for creating complex abdominal aortic aneurysm phantoms that are compatible with both radiation-based, and ultrasound-based imaging modalities, using low cost materials.

Methods Volumetric CT data of an abdominal aortic aneurysm were acquired. Regions of interest were segmented to form a model compatible with 3D printing. The novel phantom fabrication method comprised a hybrid approach of using 3D printing of water-soluble materials to create wall-less, patient-derived vascular structures embedded within tailored tissue-mimicking materials to create realistic surrounding tissues. A non-soluble 3-D printed spine was included to provide a radiological landmark.

Results The phantom was found to provide realistic appearances with intravascular ultrasound, computed tomography and transcatheter ultrasound. Furthermore, the utility of this phantom as a training model was demonstrated during a simulated endovascular aneurysm repair procedure with image fusion.

Conclusion With the hybrid fabrication method demonstrated here, complex multimodality imaging patient-derived vascular phantoms can be successfully fabricated. These have potential roles in the benchtop development of emerging imaging technologies, refinement of novel minimally invasive surgical techniques and as clinical training tools.

Keywords Ultrasound · Imaging phantoms · Tissue-mimicking material · Vascular · Abdominal aortic aneurysm

Callum D. Little and Eleanor C. Mackle contributed equally to the manuscript as first authors.

✉ Adrien E. Desjardins
a.desjardins@ucl.ac.uk

¹ Wellcome Trust-EPSCRC Centre for Interventional and Surgical Sciences, London W1W 7TS, UK

² Department of Medical Physics and Bioengineering, University College London, London WC1E 6BT, UK

³ Department of Cardiology, Royal Free Hospital, London NW3 2QG, UK

⁴ Department of Vascular Surgery, Royal Free Hospital, London NW3 2QG, UK

⁵ Division of Surgery and Interventional Science, University College London, London W1W 7TY, UK

Introduction

Abdominal aortic aneurysms are highly complex three dimensional anatomical structures of particular interest within the field of Vascular Surgery, due to the risk of a potential aneurysmal rupture [1]. Endovascular intervention is often the preferred treatment option over an open surgical approach due to decreased patient recovery time and risk of wound complications [2].

Traditionally, surgical training has been undertaken on human and animal cadavers, before progression to supervised participation on patients. Endovascular aortic repair (EVAR) procedures are particularly challenging to learn due to the need for radiological familiarisation and the requirement for

complex psychomotor skills to manipulate the endovascular devices. Furthermore, shortened post-graduate training programmes and increasing clinical commitments can limit the opportunities that trainees have to gain experience and competency with these procedures [3].

Phantom models are powerful tools for trainee surgeons learning to perform these procedures. They can provide both a realistic representation of patient anatomy and an opportunity to gain familiarity with surgical equipment and techniques, without risking harm to a patient. Simulation training on phantom models has been shown to result in decreased fluoroscopy and procedural times [4, 5] as well as overall volume of contrast used [6], as compared to traditional teaching methods. Additionally, patient-specific phantom models can be used in the preoperative planning of interventions [7, 8]. Phantom models can also provide opportunity for benchtop characterization of emerging procedural devices and techniques, such as in-situ fenestration of EVAR [9], prior to in-vivo testing.

Current simulation tools include both virtual reality (VR) platforms and physical phantoms. Although VR training platforms have shown promising outcomes in EVAR, with decreased procedural time and post-procedure endoleaks observed [10], physical phantoms are also important tools for surgeons seeking to develop the complex motor skills required for EVAR procedures [11]. However, most commercially available models have many limitations. Currently, phantoms that are low-cost often tend to have a low level of anatomical accuracy and are not typically patient specific, which limits their use for surgical planning or in-depth training. Others that are more anatomically realistic tend to be very expensive, and even those may only be compatible with one imaging modality. Many methods for creating AAA models in the literature include the use of high cost or difficult-to-source materials and techniques, such as stereolithography with UV-curing resins [12] 13, so they cannot readily be reproduced for clinical training. Additionally, most current phantoms include only the vessel structures [14] without surrounding tissues, so that their realism and use with multimodality imaging are limited.

Image fusion (IF) is a technique where data from one imaging modality are combined with another to produce an overlay in real time. Within the field of vascular surgery, it has an increasing role in the endovascular treatment of complex aortic aneurysms, and therefore, it is important that EVAR training phantoms are designed to be compatible with IF. Using IF, preoperative CT data can be overlaid onto intra-procedural fluoroscopic images, to facilitate fenestrated stent deployment [9, 15]. This requires fixed radiological landmarks, such as the spine, to be visible on both of the imaging modalities.

Here, we present a novel technique based on 3D printing for creating anatomically accurate AAA phantoms, which

can be used in the simulation of EVAR procedures. The methods and materials resulted in a model that is compatible with computerized tomography (CT) and ultrasound (US), and with IF. The geometry of the vascular structures was derived from patient data. The utility of this phantom is then demonstrated during a simulated EVAR procedure.

Methods

Phantom fabrication

There are several key steps in the fabrication method presented, which are illustrated in Fig. 1. Firstly, patient CT data are acquired allowing segmentation of areas of interest. Next, the segmentations are converted into models for 3D printing; the vasculature is then printed in a water-soluble material, whilst bone structures are printed in a non-water-soluble material. Finally, the phantom is constructed by pouring a tissue-mimicking material (TMM), which has the radiological appearance of human soft tissue, over the 3D printed model. Once the TMM has set, the phantom is submerged in water to dissolve out the vascular structures, producing a hollow, or “wall-less” phantom.

Model generation

For both models, volumetric patient-specific data were obtained from arterial phase contrast CT acquisitions (Fig. 1a). The data set was segmented manually, using an open source medical image analysis software platform (3D Slicer), according to vascular regions of interest (Fig. 1b). A 3D model of the segmented data was exported to Meshmixer (Autodesk Inc, San Rafael, CA, USA) to allow for smoothing and refinement (Fig. 1c), before generation of an STL-file compatible with 3D extrusion printing software. Unsupported structures with an overhead angle to the base plate of greater than 70° were segmented into separate models, to prevent the requirement for support material during the printing process, before later amalgamation (Fig. 1d). The wall thickness of the model was 1 mm with a layer height of 0.15 mm.

3D printing

The models were printed using a dual extrusion, commercially available 3D printer (Ultimaker S5, Ultimaker B.V., Utrecht, Netherlands) (Fig. 1e). The material used for printing was Poly(vinyl alcohol) (PVA) (RS components Ltd, Northants, UK) which is a water soluble, non-toxic filament commonly used as a support material for 3D printing. Once printed, the separate segments were combined together to form the final model using PVA glue (Elmer's Products Inc,




Fig. 1 Phantom model fabrication. **(a)** Volumetric data are acquired from a patient CT scan and **(b)** regions of interest from this data-set are manually segmented; **(c)** these segments are then exported as 3D computer aided design models and the mesh is refined to create an STL file compatible with 3D printers; **(d)** the models are orientated to negate the requirement for support material; **(e)** large overhanging structures

are separated and printed individually using PVA before **(f)** assembly into the completed model; **(g)** the printed Aorta and Spine are secured in an acrylic box with both ends of the PVA Aorta exposed; **(h)** tissue mimicking material is integrated into the box and the inner PVA Aorta is dissolved out using water

NC, USA), a water-soluble adhesive (Fig. 1f). The aorta model was secured in a custom laser-cut acrylic open-topped box, with either end of the main vascular structure exposed (Fig. 1g). An additional insoluble spine, 3D printed using Poly-lactic acid (PLA), was included within the model. This provided an anatomical landmark, allowing for image fusion (IF) guidance techniques and facilitating benchtop endovascular procedures.

Tissue-mimicking material

Gel-wax has recently emerged as a promising TMM and has been used to create heterogenous anatomical ultrasound tissue phantoms [16, 17]. It is an insoluble, mineral-oil based material that is solid at room temperature. It has a low melting point of 70°C, above which it becomes highly mobile. Ultrasound properties can be tuned through the addition of glass spheres. Gel-wax offers several advantages compared to other TMMs such as PVA or silicone. Firstly, it is both mechanically and thermally stable; secondly, it has a low viscosity, facilitating integration into phantom models; lastly it is insoluble in water which facilitates generation of wall-less phantom models when using a water-soluble inner model.

The gel-wax (Mindsets UK, Essex, UK) was heated to a temperature of 150°C for a period of 4 h to ensure uniform consistency. Air removal was facilitated using a vacuum chamber before glass microspheres were added in a concentration of 0.5% for acoustic scattering. The mixture was

sonicated for 5 min, to guarantee uniform particle distribution and achieve acoustic homogeneity. The mixture was introduced into the acrylic open-topped box, allowing the gel-wax to surround the PVA aorta and PLA spine models, cooling into a solid state. Finally, the box was immersed in de-ionised water for 24 h. The exposed ends of the main vascular structure provided an entry point for the water resulting in the dissolution of the PVA aorta, leaving behind a negative, wall-less vascular structure, surrounded by tissue-mimicking material (Fig. 1h).

Imaging

Transcutaneous ultrasound

Transcutaneous ultrasound imaging was obtained using a commercially available clinical system (HS40, Samsung Healthcare, Seoul, South Korea) with a curved array transducer probe (C2-8) operating with a frequency range of 2–6 MHz. The phantom was submerged in water to remove acoustic shadowing from air.

Intravascular ultrasound

Intravascular ultrasound was performed using an intracoronary imaging system (Opticross, Boston Scientific, Marlborough, MA, USA) operating with a transducer frequency

of 40 MHz. A 0.018-in. diameter coronary guidewire (Hi-Torque Balance Middleweight, Abbott Cardiovascular, MN, USA) was passed through the water-filled phantom and fixed at either end. The imaging catheter was then loaded onto the coronary guidewire and positioned in the region of an iliac vessel. Images were acquired with an automated pullback speed of 1 mm/s.

Computerised tomography

CT acquisition was acquired using a 320-slice scanner (Aquillion One, Canon Medical Systems Corporation, Tochigi, Japan), with a CT angiography protocol (thickness 0.5 mm; 100 kV; rotation time 0.5 s; mA R 388). To provide realistic contrast enhancement for the vascular structures, the phantom was submerged in a mixture of water and Iodixanol (Visipaque, GE Healthcare, Chicago, IL, USA), a water soluble, isosmolar, non-ionic, iodinated radiographic contrast agent.

Results

With transcutaneous ultrasound imaging various anatomical areas of interest were clearly delineated (Fig. 2a). The transcutaneous axial view (Fig. 2a) demonstrates the main aorta (A), renal (LR/RR), and superior mesenteric (SMA) arteries, visible as hypoechoic areas. The TMM appears as a homogenous hyperechoic medium surrounding these structures. In the longitudinal view (Fig. 2b) the abdominal aortic aneurysm sac (AAA) and neck are visible. Intravascular ultrasound of the phantom (Fig. 2c) demonstrates an echobright cylindrical structure consistent with the appearance of vascular endothelium, with underlying low attenuation material consistent with the appearance of adventitia.

The CT acquisition of the phantom was compared to the original patient CT dataset from which the phantom was derived (Fig. 3). Three different radiological planes, at similar slices, are demonstrated; axial (Fig. 3a, b), Sagittal (Fig. 3c, d) and coronal (Fig. 3e, f). In both acquisition sets, there is contrast clearly visible within the vascular structures, allowing for delimitation from surrounding material. The aneurysmal sac of the complex abdominal aortic aneurysm (AAA) is seen in two orthogonal views (Fig. 3d, f), with orientation of visceral side branches, such as Iliacs (IL), Renals (RR/LR), Superior Mesenteric (SMA) and Coeliac (C), preserved. The PLA spine (S) included in the phantom model is visible as a radiopaque structure (Fig. 3b) and corresponds with the appearance of the spine from the patient's CT scan (Fig. 3a).

The phantom model was utilised successfully as part of an EVAR training simulation. A trainee with no prior experience of EVAR deployed a AAA endovascular bifurcated stent

(Zenith Flex, Cook Medical, Bloomington, IL, USA) under the guidance of a Consultant Vascular Surgeon (Fig. 4a). Access to the aorta was obtained through the right iliac. Fluoroscopy was used to confirm positioning of the device (Fig. 4b).

Discussion

The method of fabrication that we have presented here can be used to create a high-fidelity patient specific vascular phantom of an AAA, compatible with both US and CT imaging. The novelty of this study stems from several factors. Firstly, the use of 3D printed dissolvable vascular models embedded within TMM. Secondly, we have demonstrated the phantom's utility as a training tool during simulation of an EVAR procedure. The materials used in the fabrication process are inexpensive, stable at room temperature and have no specific limitations on shelf-life. The total material cost of the phantom was less than €100. Similar phantoms with material costs of up to €2000 are quoted in the literature as reaching [13]. The use of 3D printed dissolvable models allows for complex geometries to be generated, which not achievable using current techniques. The inclusion of radiopaque landmarks within the model ensures compatibility with image fusion guidance techniques, which help facilitate detection of visceral side-branches during complex EVAR procedures [15]. A further advantage of this fabrication method is that the process of dissolving the printed structures out of the model leaves behind a hollow, wall-less vasculature. This then has a realistic ultrasound appearance, and does not have the highly reflective signals that are present using traditional silicone tubing or resin-based imaging phantoms.

Whilst the potential utility of the fabricated AAA phantom model has been demonstrated in the simulation of an EVAR procedure, further evaluation in the form of surgeon feedback and objective outcome data is required to determine the efficacy of this as a training tool. It is, however, envisaged that other complex vascular structures, such as coronary or cerebral vessels, can be recreated using the same method. This may allow fabrication of suitable phantoms for simulation of procedures relevant to trainee Interventional Cardiologists or Radiologists, provided accurate luminal casts can be made of small vessels. In addition to compatibility with transcutaneous ultrasound, the phantom model produces realistic imaging when IVUS is used. IVUS has a role in EVAR procedures allowing accurate sizing of vessels and reducing contrast volume [18]. IVUS is also frequently used to guide coronary and peripheral vascular angioplasty and therefore compatibility with this imaging modality broadens the applicability of this phantom. Compatibility with both IVUS and transcutaneous US could allow for EVAR procedures to be performed with reduced ionising radiation. For instance,





Fig. 2 Ultrasound imaging of the phantom model. (a) External ultrasound demonstrating an axial view of the Aorta (A) surrounded by an ultrasonically homogenous tissue mimicking material. Left (LR)/ Right (RR) Renal arteries and the Superior Mesenteric Artery (SMA) origins

are visible. (b) A longitudinal view demonstrating the Abdominal Aortic Aneurysm sac (AAA). (c) Intravascular ultrasound with the imaging catheter (IVUS) located centrally within the lumen of the vessel. W represents acoustic artefact created by the coronary guidewire used to insert the imaging probe

Fig. 3 Comparison of CT acquisitions from both the patient with an abdominal aortic aneurysm and the patient-derived phantom model. (a) Axial view of patient CT at the level of the renal artery bifurcation with the left renal artery (LR), spine (S) and aorta (A) demonstrated; (b) Corresponding axial view from the phantom model CT demonstrating the same structures surrounded by tissue mimicking material (TMM); (c) Sagittal view of patient CT demonstrating abdominal aortic aneurysm (AAA) sac, coeliac artery (C), superior mesenteric artery (SMA) and iliac artery (IL); (d) Corresponding sagittal view from the phantom model CT; (e) Coronal view of patient CT demonstrating the abdominal aortic aneurysm (AAA) and right renal artery (RR); (f) Corresponding coronal slice of phantom model CT




emerging ultrasonic tracking technologies with fibre optic receivers integrated into devices [19] could be used to accurately position the stent graft; likewise, emerging rotational intravascular technologies could be used to align fenestrations with side branches [20].

The smallest vessel size fabricated in this model was the right renal artery (ca. 5 mm diameter); however depending on the resolution of the 3D printer system used, it may be possible to include vessels as small as 1 mm in diameter [21]. This would allow reproduction of complex arteriole vascular beds, such as those seen with arterio-venous malformations [22] if

a good luminal model could be created. Indeed, the utility of dissolvable structures is not limited to recreation of vascular structures but can also potentially be used to simulate other hollow branching systems such as lymphatics, ureters or the biliary tree. Surrounding these structures with tissue mimicking material, cast in organ-specific shapes, could allow for realistic imaging phantoms for visceral organs such as the liver, kidney or prostate. The TMM used was specific to ultrasound imaging, however inclusion of alternative materials with differing imaging contrast could potentially allow for compatibility with further imaging modalities such as

Fig. 4 Simulated EVAR procedure. (a) Insertion of catheter (Ca) into the aorta (A) during the simulated EVAR procedure with the spine (S) visible within the tissue mimicking material (TMM); (b) anterior–posterior fluoroscopic view of the phantom model post EVAR stent deployment. The spine (S) is visible as a hyperdense structure with the left renal artery (LR), right renal artery (RR) and aorta (A) visible as hypodense structures



magnetic resonance imaging (MRI). Furthermore, the elastic properties of the tissue mimicking material used in this study can be altered by the addition of paraffin oil [17], allowing further tailoring of vascular phantoms for use with flow simulations in a wide range of clinical contexts [23, 24].

Conclusions

Through a hybrid approach including 3D printing vascular structures in water-soluble material, and integration of tissue mimicking materials, anatomically complex multimodality patient-specific imaging phantoms can be fabricated, which are of use in simulation training of EVAR.

Funding This work was supported by the National Institute for Health Research UCL Biomedical Research Centre and the Wellcome/EPSRC Centre for Interventional and Surgical Sciences (WEISS) (203145Z/16/Z; NS/A000050/1).

Declarations

Conflict of interest The authors declare that they have no conflict of interest.

Research involving human participants and/or animals All procedures performed in studies involving human participants were in accordance with the ethical standards of the institutional and/or national research committee and with the 1964 Helsinki Declaration and its later amendments or comparable ethical standards. This article does not contain any studies with animals performed by any of the authors.

Informed consent For this study, an anonymised computed tomography (CT) dataset was provided by T.M.M. (co-author). Informed consent was obtained from all individual participants included in the study.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

1. Brown PM, Zelt DT, Sobolev B (2003) The risk of rupture in untreated aneurysms: the impact of size, gender, and expansion rate. *J Vasc Surg* 37:280–284. <https://doi.org/10.1067/mva.2003.119>
2. Chaikof EL, Dalman RL, Eskandari MK, Jackson BM, Lee WA, Mansour MA, Mastracci TM, Mell M, Murad MH, Nguyen LL, Oderich GS (2018) The Society for vascular surgery practice guidelines on the care of patients with an abdominal aortic aneurysm. *J Vasc Surg* 67:2–77. <https://doi.org/10.1016/j.jvs.2017.10.044>
3. Dean B, Pereira E (2011) Surgeons and training time. *BMJ*. <https://doi.org/10.1136/bmj.d6724>
4. Torres IO, De Luccia N (2017) A simulator for training in endovascular aneurysm repair: the use of three dimensional printers. *Eur J Vasc Endovasc Surg* 54:247–253. <https://doi.org/10.1016/j.ejvs.2017.05.011>
5. Kim AH, Kendrick DE, Moorehead PA, Nagavalli A, Miller CP, Liu NT, Wang JC, Kashyap VS (2016) Endovascular aneurysm repair simulation can lead to decreased fluoroscopy time and accurately

- delineate the proximal seal zone. *J Vasc Surg* 64:251–258. <https://doi.org/10.1016/j.jvs.2016.01.050>
6. Vento V, Cercenelli L, Mascoli C, Gallitto E, Ancetti S, Faggioli G, Freyre A, Marcelli E, Gargiulo M, Stella A (2018) The role of simulation in boosting the learning curve in EVAR procedures. *J Surg Educ* 75:534–540. <https://doi.org/10.1016/j.jsurg.2017.08.013>
 7. Bortman J, Mahmood F, Schermerhorn M, Lo R, Swerdlow N, Mahmood F, Matyal R (2019) Use of 3-dimensional printing to create patient-specific abdominal aortic aneurysm models for preoperative planning. *J Cardiothorac Vasc Anesth* 33(5):1442–1446. <https://doi.org/10.1053/j.jcva.2018.08.011>
 8. Coles-Black J, Chuen J (2019) 3D printed AAA phantoms for presurgical evar simulation—a single center experience. *Eur J Vasc Endovasc Surg* 33:1442–1446. <https://doi.org/10.1016/j.ejvs.2019.09.296>
 9. Leger T, Tacher V, Majewski M, Touma J, Desgranges P, Kobeiter H (2019) Image fusion guidance for in situ laser fenestration of aortic stent graft for endovascular repair of complex aortic aneurysm: feasibility, efficacy and overall functional success. *Cardiovasc Intervent Radiol* 42:1371–1379. <https://doi.org/10.1007/s00270-019-02231-8>
 10. Saratzis A, Calderbank T, Sidloff D, Bown MJ, Davies RS (2017) Role of simulation in endovascular aneurysm repair (EVAR) training: a preliminary study. *Eur J Vasc Endovasc Surg* 53:193–198. <https://doi.org/10.1016/j.ejvs.2016.11.016>
 11. Kärkkäinen JM, Sandri G, Tenorio ER, Alexander A, Bjellum K, Matsumoto J, Morris J, Mendes BC, DeMartino RR, Oderich GS (2019) Simulation of endovascular aortic repair using 3D printed abdominal aortic aneurysm model and fluid pump. *Cardiovasc Intervent Radiol* 42:1627–1634. <https://doi.org/10.1007/s00270-019-02257-y>
 12. Kim T, Hong D, Ock J, Park SJ, Rhee Y, Lee S, Kim GB, Yang DH, Kim JB, Kim N (2021) Utilizing patient-specific 3D printed guides for graft reconstruction in thoracoabdominal aortic repair. *Sci Rep*. <https://doi.org/10.1038/s41598-021-97541-8>
 13. Tam CHA, Chan YC, Law Y, Cheng SWK (2018) The role of three-dimensional printing in contemporary vascular and endovascular surgery: a systematic review. *Ann Vasc Surg*
 14. Meess KM, Izzo RL, Dryjski ML, Curl RE, Harris LM, Springer M, Siddiqui AH, Rudin S, Ionita CN (2017) 3D printed abdominal aortic aneurysm phantom for image guided surgical planning with a patient specific fenestrated endovascular graft system. *Proc SPIE Int Soc Opt Eng*. <https://doi.org/10.1117/12.2253902>
 15. Sailer AM, De Haan MW, Peppelenbosch AG, Jacobs MJ, Wildberger JE, Schurink GW (2014) CTA with fluoroscopy image fusion guidance in endovascular complex aortic aneurysm repair. *Eur J Vasc Endovasc Surg* 47:349–356
 16. Maneas E, Xia W, Nikitichev DI, Daher B, Manimaran M, Wong RY, Chang CW, Rahmani B, Capelli C, Schievano S, Burriesci G (2018) Anatomically realistic ultrasound phantoms using gel wax with 3D printed moulds. *Phys Med Biol* 63:015033. <https://doi.org/10.1088/1361-6560/aa9e2c>
 17. Maneas E, Xia W, Ogunlade O, Fonseca M, Nikitichev DI, David AL, West SJ, Ourselin S, Hebden JC, Vercauteren T, Desjardins AE (2018) Gel wax-based tissue-mimicking phantoms for multispectral photoacoustic imaging. *Biomed Opt Express* 9:1151–1163. <https://doi.org/10.1364/boe.9.001151>
 18. Illuminati G, Pacilè MA, Ceccanei G, Ruggeri M, La Torre G, Ricco JB (2020) Peroperative intravascular ultrasound for endovascular aneurysm repair versus peroperative angiography: a pilot study in fit patients with favorable anatomy. *Ann Vasc Surg* 64:54–61. <https://doi.org/10.1016/j.avsg.2019.11.013>
 19. Mathews SJ, Shakir DI, Mosse CA, Xia W, Zhang EZ, Beard PC, West SJ, David A, Ourselin S, Vercauteren T, Desjardins AE (2022) Ultrasonic needle tracking with dynamic electronic focusing. *Ultrasound Med Biol* 48:520–529. <https://doi.org/10.1016/j.ultrasmedbio.2021.11.008>
 20. Colchester RJ, Little C, Dwyer G, Noimark S, Alles EJ, Zhang EZ, Loder CD, Parkin I, Papakonstantinou I, Beard P, Finlay M, Rakhit RD, Desjardins AE (2019) All-Optical Rotational Ultrasound Imaging. *Sci Rep* 9:1–8. <https://doi.org/10.1038/s41598-019-41970-z>
 21. Mackle EC, Maneas E, Little C, Carr E, Xia W, Nikitichev D, Rakhit RD, Finlay MC, Desjardins AE (2019) Wall-less vascular poly(vinyl) alcohol gel ultrasound imaging phantoms using 3D printed vessels. *Proc SPIE BiOS*. <https://doi.org/10.1117/12.2510033>
 22. Conti A, Pontoriero A, Iati G, Marino D, La Torre D, Vinci S, Germanò A, Pergolizzi S, Francesco T (2016) 3D-printing of arteriovenous malformations for radiosurgical treatment: pushing anatomy understanding to real boundaries. *Cureus* 8:594. <https://doi.org/10.7759/cureus.594>
 23. Tabaczynski JR, Stoll T, Shepard L, Siddiqui MI, Karkhanis NV, Sommer K, Siddiqui AH, Ionita CN (2018) Use of patient specific 3D printed (3DP) neurovascular phantoms for mechanical assessment of devices used in image guided minimally invasive procedures. *Proc SPIE Med Im*. <https://doi.org/10.1117/12.2293370>
 24. Wang Y, Joannic D, Juillion P, Monnet A, Delassus P, Lalande A, Fontaine JF (2018) Validation of the strain assessment of a phantom of abdominal aortic aneurysm: comparison of results obtained from magnetic resonance imaging and stereovision measurements. *J Biomech Eng* 140:031001. <https://doi.org/10.1115/1.4038743>

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.