

Sign determination methods for the respiratory signal in data-driven PET gating

This content has been downloaded from IOPscience. Please scroll down to see the full text.

2017 Phys. Med. Biol. 62 3204

(<http://iopscience.iop.org/0031-9155/62/8/3204>)

View [the table of contents for this issue](#), or go to the [journal homepage](#) for more

Download details:

IP Address: 128.41.61.19

This content was downloaded on 01/06/2017 at 18:34

Please note that [terms and conditions apply](#).

You may also be interested in:

[Extension of a data-driven gating technique to 3D, whole body PET studies](#)

Paul J Schleyer, Michael J O'Doherty and Paul K Marsden

[Data-driven event-by-event respiratory motion correction using TOF PET list-mode centroid of distribution](#)

Silin Ren, Xiao Jin, Chung Chan et al.

[Retrospective data-driven respiratory gating for PET/CT](#)

Paul J Schleyer, Michael J O'Doherty, Sally F Barrington et al.

[Joint PET-MR respiratory motion models for clinical PET motion correction](#)

Richard Manber, Kris Thielemans, Brian F Hutton et al.

[Impact of respiratory motion on tumor quantification and delineation in static PET/CT imaging](#)

Chi Liu, Larry A Pierce II, Adam M Alessio et al.

[Fully 4D list-mode reconstruction applied to respiratory-gated PET scans](#)

N Grotus, A J Reader, S Stute et al.

[Elastic respiratory motion correction in PET list-mode reconstruction](#)

F Lamare, M J Ledesma Carbayo, T Cresson et al.

3D Geometric Distortion of MRI

Analyze large 3D volumes in <10 minutes

[Learn More](#)



Sign determination methods for the respiratory signal in data-driven PET gating

Ottavia Bertolli¹, Simon Arridge², Scott D Wollenweber³, Charles W Stearns³, Brian F Hutton¹ and Kris Thielemans¹

¹ Institute of Nuclear Medicine, UCL, London, United Kingdom

² Department of Computer Science, UCL, London, United Kingdom

³ GE Healthcare, Waukesha, WI, United States of America

E-mail: ottavia.bertolli.13@ucl.ac.uk

Received 11 February 2016, revised 9 January 2017

Accepted for publication 14 February 2017

Published 24 March 2017



CrossMark

Abstract

Patient respiratory motion during PET image acquisition leads to blurring in the reconstructed images and may cause significant artifacts, resulting in decreased lesion detectability, inaccurate standard uptake value calculation and incorrect treatment planning in radiation therapy. To reduce these effects data can be regrouped into (nearly) ‘motion-free’ gates prior to reconstruction by selecting the events with respect to the breathing phase. This gating procedure therefore needs a respiratory signal: on current scanners it is obtained from an external device, whereas with data driven (DD) methods it can be directly obtained from the raw PET data. DD methods thus eliminate the use of external equipment, which is often expensive, needs prior setup and can cause patient discomfort, and they could also potentially provide increased fidelity to the internal movement. DD methods have been recently applied on PET data showing promising results. However, many methods provide signals whose direction with respect to the physical motion is uncertain (i.e. their sign is arbitrary), therefore a maximum in the signal could refer either to the end-inspiration or end-expiration phase, possibly causing inaccurate motion correction. In this work we propose two novel methods, *CorrWeights* and *CorrSino*, to detect the correct direction of the motion represented by the DD signal, that is obtained by applying principal component analysis (PCA) on the acquired data. They only require the PET raw data, and they rely on the assumption that one of the major causes of change in the acquired data related to the chest is respiratory motion in the axial direction, that generates a cranio-caudal motion of the internal organs. We also implemented two versions of a



Original content from this work may be used under the terms of the [Creative Commons Attribution 3.0 licence](https://creativecommons.org/licenses/by/3.0/). Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.

published registration-based method, that require image reconstruction. The methods were first applied on XCAT simulations, and later evaluated on cancer patient datasets monitored by the Varian Real-time Position Management™ (RPM) device, selecting the lower chest bed positions. For each patient different time intervals were evaluated ranging from 50 to 300s in duration. The novel methods proved to be generally more accurate than the registration-based ones in detecting the correct sign of the respiratory signal, and their failure rates are lower than 3% when the DD signal is highly correlated with the RPM. They also have the advantage of faster computation time, avoiding reconstruction. Moreover, *CorrWeights* is not specifically related to PCA and considering its simple implementation, it could easily be applied together with any DD method in clinical practice.

Keywords: PET imaging, respiratory motion, data-driven gating

(Some figures may appear in colour only in the online journal)

1. Introduction

For a diagnostic PET/CT exam a typical acquisition duration for the PET scan is 2 min per bed position, therefore making the patient respiratory motion unavoidable and the ‘breath-hold’ technique often applied to CT acquisitions (Fin *et al* 2008) unfeasible. The acquired data are therefore averaged over several breathing cycles and, if no correction is applied, respiratory motion will result in a blurring of the reconstructed images and in an overall degradation of the contrast (Nehmeh *et al* 2002b, Liu *et al* 2009). Moreover, in PET/CT respiratory motion may lead to mismatch between CT and PET datasets, possibly generating artifacts in the reconstructed PET images due to errors in the attenuation correction (Erdi *et al* 2004). There are therefore several reasons why accounting for respiratory motion when reconstructing PET data can improve PET image quality

In order to reduce the artifacts induced by the patient breathing, PET data can be acquired in discrete bins in correlation with the respiratory cycle, therefore minimizing the motion within a single bin. As a result, multiple PET images are reconstructed, each related to a section of the respiratory cycle, and motion artefacts and image blurring are expected to decrease compared to the non-gated image. Studies have shown, both in phantom experiments and with lung cancer patient data, that the application of respiratory gating reduces the degrading effect of breathing motion in PET images, producing a reduction in the lesion volume up to 34% and an increase in SUV_{max} up to 156.16% compared to the non-gated images (Nehmeh *et al* 2002a, 2002b).

In clinical practice respiratory gating techniques rely on external devices to measure the respiratory motion and several solutions have been implemented on commercial imaging systems (Visvikis *et al* 2006, Rahmim *et al* 2007, Pepin *et al* 2014). Common utilized devices are the Anzai AZ-733V system (Anzai Medical Corp., Tokyo, Japan), which is a pressure belt that measures the expansion of the patient’s chest (Riedel *et al* 2006), and the Real-time Position Management (RPM) system (Varian Medical Systems, Palo Alto, California, USA), which tracks with a video camera the motion of an infrared marker placed on the patient’s abdomen (Nehmeh *et al* 2002b, Otani *et al* 2010). Alternatively other methods have been proposed that measure the temperature (Boucher *et al* 2004) or flow of the air (Guivarc’h *et al* 2004) during

patient respiration, via a thermistor and a spirometer respectively, then relating those changes to respiratory motion.

Considering the extra costs related to external tracking devices, the discomfort patients could experience because of them and also because of some evidence that the internal organ motion might differ from patient external surface motion (Ozhasoglu and Murphy 2002, Fayad *et al* 2011), data-driven (DD) post-acquisition approaches have been suggested, see Kesner *et al* (2014) for a review. In Nehmeh *et al* (2003) an ^{18}F -FDG point source is set on the patient's abdomen and its position used to track respiratory motion through the consecutive dynamic frames; in Visvikis *et al* (2003) time activity curves are obtained from ROIs selected on the series of dynamic images and a Fourier transform in time is performed in order to estimate the frequency of the motion, assuming periodicity; in Bundschuh *et al* (2007) a volume of interest (VOI) is manually defined in a summed image around the lesion, the center of mass (COM) of the activity distribution inside the VOI is determined and the z-coordinate of the COM as a function of time delivers the respiratory signal. These methods require image reconstruction whereas others only rely on the acquired raw data, and do not need any manual input (e.g. for drawing ROIs/VOIs) or the use of external radioactive sources: in Schleyer *et al* (2009) the variation of the counts within regions subject to respiratory motion is used to estimate the respiratory signal (Spectral Analysis Method, SAM); in Thielemans *et al* (2011) principal component analysis (PCA) is applied to the listmode PET data, and the respiratory signal is obtained as the principal component weight factor whose frequency spectrum has the highest peak in the respiratory band; in Wachinger *et al* (2012) an alternative dimensionality reduction technique is utilized, i.e. laplacian eigenmaps (LE) decomposition, in order to obtain the respiratory signal from the raw data; in He *et al* (2008) the non-uniformity of the geometric sensitivity of the scanner is exploited as a means of obtaining information on motion, since the count rate for a given organ will depend on the axial location of the organ within the scanner, the respiratory phase is determined from count rate changes in the listmode data (SENS method); in Kesner and Kuntner (2010) a respiratory trace is obtained from the rebinned sinogram by combining together information from the time activity curves extracted from all the voxels of the sinogram (sinogram region fluctuation method (SRF)).

However, many of these methods, and in particular SAM and PCA, provide respiratory signals that are only determined up to an arbitrary scale factor. As this scale factor can also be negative, this means that the direction of the represented motion is undefined and the maxima in the signal could refer either to the end-inspiration or end-expiration phase of the breathing cycle. The possibility of obtaining a signal that represents motion in the opposite direction compared to the truth could lead to inaccurate motion correction, inconsistencies between adjacent beds and an incorrect matching between PET and CT gates when performing attenuation correction, e.g. when using an end-expiration CT. In order to correctly identify the gates, the correspondence between the sign of the signal and the direction of motion has to be determined and fixed. This issue has been previously addressed in Schleyer *et al* (2011), where a method requiring the registration in the axial direction of the gated PET images (reconstructed without attenuation correction) is implemented.

In this work we propose two novel methods, *CorrWeights* and *CorrSino*, for the determination of the direction of motion of PCA respiratory signal, we implement a published method (Schleyer *et al* 2011) and compare the performance of the methods on simulation data and FDG lung cancer patients' data.

2. Methods

2.1. Extraction of respiratory signal

The data-driven (DD) method implemented in this work is PCA, as in Thielemans *et al* (2011), and in this section it will be described in detail as it is necessary for the later introduction of the proposed novel methods. To describe the method we will assume *single slice rebinning* (SSRB) (Daube-Witherspoon and Muehllehner 1987) has been performed on 3D PET data. The following coordinate system will be used (Bailey 2005): the radial position r is the radial displacement of the line of response (LOR) in the transaxial plane, the axial coordinate z , the azimuthal angle ϕ is the angle formed between the LOR (projected onto the transaxial plane) and the y -axis of the scanner, and the time index t . All coordinates are discretized. The dynamic sinogram data are therefore described as the vector $d(r, z, \phi, t)$, and the empirical temporal mean on the total number of time frames T is given by:

$$\bar{d}(r, z, \phi) = \frac{1}{T} \sum_{t=1}^T d(r, z, \phi, t). \quad (1)$$

PCA is a ‘dimensionality reduction’ technique that attempts to find a linear transformation between the original data and a space of lower dimension (Pearson 1901), where the structure of the original data can be better observed. PCA represents the data as a combination of orthogonal basis vectors that are ordered with respect to descending variance, such that the first vector has the largest variance and thus represents the maximum observed variation. Given the data $d(r, z, \phi, t)$, PCA provides the following linear expansion:

$$d(r, z, \phi, t) \approx \bar{d}(r, z, \phi) + \sum_{k=1}^K p_k(r, z, \phi) w_k(t) \quad (2a)$$

$$w_k(t) = \sum_{r, z, \phi} p_k(r, z, \phi) (d(r, z, \phi, t) - \bar{d}(r, z, \phi)) \quad (2b)$$

where $p_k(r, z, \phi)$ are the principal components and $w_k(t)$ are the weight factors, that are the coordinates of the sinogram data described in the new basis, that has K vectors. The equality in (2b) derives from the orthonormality of the Principal Components, as they are basis vectors. Usually only the first three components contain motion related information so in our implementation $K = 3$.

PCA can be directly applied to PET dynamic sinograms as they are acquired by scanners but there are some issues related to this, e.g. the noise in the data and also the high spatial resolution, that provides more data than is actually needed by PCA to detect the biggest variations in time. In our implementation we therefore unlist the PET listmode data into dynamic sinograms (with time frame duration of 500 ms) with reduced spatial resolution. Furthermore, these are processed using the Freeman–Tukey variance stabilisation technique (Freeman and Tukey 1950) to obtain approximately normally distributed samples, as described in (Thielemans *et al* 2011):

$$d(r, z, \phi, t)_{\text{FT}} = \sqrt{d(r, z, \phi, t)} + \sqrt{d(r, z, \phi, t) + 1} \quad (3)$$

then spatial filtering is performed on $d(r, z, \phi, t)_{\text{FT}}$ and finally PCA is applied. Note that this variance stabilization process effectively pre-whitens the sinogram data before the application of PCA.

The most respiratory-like signal is obtained by selecting the weight factor $w_k(t)$ whose main frequency most closely corresponds to respiration according to the following procedure (see figure 1): the power spectrum $\mathcal{S}_k(f) = |\hat{w}_k(f)|$ (where \hat{w} is the FFT of the signal at frequency f) is computed for the three weight factors, the ratio R_k between the peak in the respiratory frequency band ([0.1, 0.4] Hz) and the mean power above 0.4 Hz is evaluated, the $w_k(t)$ that provides the highest R_k is selected (see figure 1(b)). The PC corresponding to the selected weight factor is defined as the respiratory PC (RPC) and contains information on the respiratory motion.

The sign of the respiratory signal thus obtained is not unequivocally determined: in (2a) the result of the product $p_k(r, z, \phi)w_k(t)$ will not change if, at the same time, $w_k(t) \rightarrow -w_k(t)$ and $p_k(r, z, \phi) \rightarrow -p_k(r, z, \phi)$, therefore making the sign of the signal returned by PCA arbitrary, i.e. independent of noise and implementation details. Since the RPC weight factor (hereinafter referred to as RPC signal) needs to represent the physical respiratory motion, where maxima and minima should correspond to the end-expiration and end-inspiration phase respectively, the correct sign of $w_k(t)$ has to be determined prior to using it as a gating signal. Furthermore, consistency in the relationship between the signal and the direction of motion is required when combining adjacent bed positions.

As we are considering chest acquisitions, cardiac motion will also be present in the acquired data. Data-driven methods have proven to be capable of detecting cardiac motion (Büther *et al* 2009, Thielemans *et al* 2014), however the temporal resolution of the unlisted sinograms has to be high enough in order to be able to detect changes with cardiac frequencies, that are generally in the range of 1–2 Hz. In this work, the temporal resolution of 500 ms is designed to reduce the sensitivity of PCA to the cardiac contraction.

2.2. Sign determination methods

In order to determine the direction of the motion represented by the DD signal, Schleyer *et al* (2011) assume that respiration causes a cranio-caudal motion of the internal organs. The method analyses reconstructed images of the gated data. In this paper, we developed two new methods based on the same assumption that axial motion is the major cause of changes in the acquired data. However, these methods only require the PET sinograms and neither gating nor reconstruction are needed. All the following described methods take as input PET listmode data, that are then unlisted into dynamic sinograms $d(r, z, \phi, t)$ with a temporal resolution of 500 ms. The DD respiratory signal is obtained by applying PCA on sinograms that have undergone SSRB (see section 2.1), and radial and angular mashing (i.e. the summation of adjacent sinogram elements with respect to r and ϕ to reduce the dimension).

2.2.1. Registration-based method. The method as proposed in Schleyer *et al* (2011) performs registration via 1D rigid translation on gated images reconstructed without attenuation correction (to avoid issues with misalignment with the CT). The acquired data are gated with displacement gating based on the DD signal, normalized for detection efficiencies and reconstructed with FBP (with ramp filter). The registration process is performed on the sagittal and coronal Maximum Intensity Projection images (MIP) of the gates, in order to reduce noise. For each gate the displacement in axial direction (z_{coronal} or z_{sagittal}) with respect to a reference gate is estimated. The final registration displacement is given as the average of the two.

We implemented the method with minor modifications. Each of the obtained gates is in turn taken as the reference for the registration and an average of the obtained displacements is obtained, rather than choosing as single reference the gate with most counts. We did this





Figure 1. (a) The weights w_1 , w_2 and w_3 related to the first three PCs, obtained from one patient dataset. (b) Power spectrum of the 3 weights. The peak within the respiratory frequency band is highlighted in red, the mean power above 0.4 Hz is displayed in light-blue and the value of their ratio for each weight is: $R_1 = 15.0$, $R_2 = 2.9$ and $R_3 = 3.7$. In this case, w_1 would be chosen as the respiratory signal.

because our displacement gating was set to distribute the counts among the gates as evenly as possible to avoid large differences in noise between the gates. In addition, we have minimised the ℓ^1 norm of the sum of the absolute differences in pixel values, rather than the ℓ^2 norm, to increase robustness. This process is performed both with sagittal and coronal MIPs (obtaining $z_{\text{coronal,avg}}$ and $z_{\text{sagittal,avg}}$) and the direction of motion is defined by looking at:

$$\text{dir} = \text{sign}(z_{\text{coronal,avg}} + z_{\text{sagittal,avg}}) \quad (4)$$

if $\text{sign} > 0$ the sign of the PC respiratory signal is kept as it is, if $\text{sign} < 0$ then the respiratory signal is reversed. We also implemented a variation of this method by registering, as opposed to the MIPs, the images obtained by collapsing the 3D images in the sagittal or coronal plane (equivalent to summing together all the 2D sagittal or coronal planes). The two implementations are referred to as *MIP* and *SUM*.

2.2.2. Sinogram-based methods. Under the assumption that motion occurs uniquely in the z -direction, we derived an approximation for the dynamic sinograms.

If we assume that motion occurs only in the z -direction, the change in the dynamic sinogram can be represented by $d(z, t)$ (we omit the dependencies on r and ϕ in the following steps). Since the sensitivity of the scanner is constant in time and depends on z , in order to determine the time dependence of the sinograms, we will first consider how the sinograms normalised for the axial pattern in the sensitivity change over time:

$$d(z, t) = s(z)n(z, t) \quad \implies n(z, t) = \frac{d(z, t)}{s(z)} \quad (5)$$

where $s(z)$ is the sensitivity of the scanner (determined from the number of rings and axial acceptance of the scanner) and $n(z, t)$ the normalized sinograms, that are going to be used in the following steps. The normalization of the sinograms is performed after SSRB. Assuming small translations $\delta z(t)$ with respect to the normalized sinogram in $t = 0$, the normalized dynamic sinogram can be described as follows:

$$n(z, t) = n_0(z + \delta z(t)) \quad \text{where} \quad \delta z(0) = 0, \quad n_0(z) = n(z, 0) \quad (6)$$

and writing the Taylor expansion up to the first order we get:

$$n(z, t) \sim n_0(z) + \delta z(t)n'_0(z) \quad (7)$$

with $n'(z)$ the gradient of the sinograms along the z -direction. Subtracting the temporal mean we get:

$$n(z, t) - \bar{n}(z) \sim (\delta z(t) - \overline{\delta z})n'_0(z) \quad (8)$$

where $\bar{n}(z)$ and $\overline{\delta z}$ are the temporal mean of $n(z, t)$ and $\delta z(t)$. The left-hand side term of (8) can be obtained directly from the dynamic sinogram, whereas if $n'_0(z)$ were extracted from the initial temporal frame ($t = 0$) it would be too noisy. We therefore chose to approximate $n_0(z)$ with the temporal mean of the entire sinograms $\bar{n}(z)$, consequently decreasing the noise. The gradient along z can be approximated using finite differences and is defined by GradSino(z):

$$\text{GradSino}(z) = \frac{\partial}{\partial z} \bar{n}(z) \sim \frac{\bar{n}(z+1) - \bar{n}(z-1)}{2}. \quad (9)$$

As a result the 'component' $n'_0(z)$ in (8), that is multiplied by the function describing the axial motion, is approximated by GradSino:

$$n(z, t) - \bar{n}(z) \sim (\delta z(t) - \overline{\delta z})\text{GradSino}(z). \quad (10)$$

In order to reproduce the process performed on the sinograms that is ultimately utilised for PCA, we approximate the Freeman–Tukey transformation on the original dynamic sinogram $d(z, t)$ and apply a Taylor expansion, given the assumption that the motion in the z -direction




Figure 2. (a) Example of RPC for a time interval of the dynamic sinogram; (b) temporal mean of the dynamic sinogram; (c) WeightedGradSino obtained from the mean. Grey scale range for (a) is $[-0.05, 0.03]$ and for (c) is $[-0.65, 0.40]$ and for (b) is $(0.07, 0.86)$. RPC, mean and WeightedGradSino are sinograms and what is shown here are the projections in the anterior–posterior direction. The horizontal and vertical coordinates in these images refer to the r and z of the sinograms. Truncation in the WeightedGradSino is performed to avoid issues with border conditions.

is small (see appendix for detailed description). The final result is that the dynamic sinogram that is given to PCA can be represented by the following (the temporal mean is subtracted):

$$d(z, t)_{\text{FT}} - \bar{d}(z)_{\text{FT}} \simeq (\delta z(t) - \bar{\delta z}) \text{WeightedGradSino}(z). \quad (11)$$

where:

$$\text{WeightedGradSino}(z) = \frac{\text{GradSino}(z) \sqrt{s(z)}}{\sqrt{\bar{n}(z)}}. \quad (12)$$

This description of the temporal changes of the sinograms from its mean resembles the result expected from PCA: it represents the dynamic sinogram with a product of a temporal one-dimensional signal with a component in the sinogram space. This similarity could be exploited to compare the assumed motion along the axial direction to the PCA output. The respiratory signal of interest is represented by the term $\delta z(t) - \bar{\delta z}$, and can be described as the ‘weight’ of the component represented by WeightedGradSino.

Figure 2(a) shows an example of the RPC obtained from the application of PCA on a time interval (i.e. temporal portion) of the acquisition of a patient, figure 2(b) shows the temporal mean of the sinograms and figure 2(c) shows the resulting WeightedGradSino (where it can be seen that 4 planes are set to zero at the top and at the bottom of the projection, in order to avoid issues with boundary conditions when applying (9)).

The above approximations motivate two new methods that only require the PET raw data (listmode data that are consequently unlisted into sinograms) to determine the sign of the RPC signal, and these are presented in the following sections.

CorrWeights method. The correspondence between the WeightedGradSino and the direction of motion is known as it derives from the application of the gradient in the z -direction, see (9), therefore its corresponding weight is related to the correct direction of the motion. An increase in the signal would correspond to motion towards the head (‘up’) while a decrease corresponds to motion towards the feet (‘down’), thus maxima would correspond to the end-expiration phase and minima to the end-inspiration phase.

As a consequence of (11), the WeightedGradSino weight can be obtained making use of (2b). Therefore evaluating the inner-product between the WeightedGradSino and the dynamic sinogram on the left-hand side of (11):

$$u(t) = \sum_{r,z,\phi} \text{WeightedGradSino}(r, z, \phi) (d(r, z, \phi, t)_{\text{FT}} - \bar{d}(r, z, \phi)_{\text{FT}}) \quad (13)$$




Figure 3. On the left the RPC signal, $w(t)$; on the right the WeightedGradSino signal $u(t)$, corresponding to the same time interval of the dynamic sinogram.

where the sums are over all sinogram elements. Examples of an RPC signal w and the WeightedGradSino weight u corresponding to the same time interval of PET data is shown in figure 3, where high similarity between the two signals can be observed.

In practice, the signal u cannot usually be used as a good respiratory signal since (13) is based on various approximations (*viz.* that the changes in the dynamic sinogram are small and only in the axial direction, see (6) and (7)). Its importance lies in its known relationship with the direction of the axial motion, that makes it a comparison tool to determine whether w correctly represents the direction of the physical motion of the patient.

These arguments lead us to propose the following method to determine the direction of the DD respiratory signal: if the correlation between the two signals w and u is calculated, the sign of the correlation gives an indication about whether the DD respiratory signal and WeightedGradSino signal represent motion in the same direction. The Pearson correlation between the two signals is given by:

$$\text{corr}(w, u) = \frac{\sum_i (w(t) - \bar{w})(u(t) - \bar{u})}{\sqrt{\sum_i (w(t) - \bar{w})^2 (u(t) - \bar{u})^2}}. \quad (14)$$

If the correlation is negative it is an indicator that w represents a motion in the opposite direction compared to the motion related to u , thus the sign of w (i.e. the respiratory signal used for gating) has to be reversed. For the example shown in figure 3 (where only a 60 s interval of the total 300 s signals is displayed) the correlation is 0.98, while in other cases it can be lower.

This method is not specifically related to PCA, therefore could be utilized to determine the correct direction of the motion of respiratory signals obtained with any other DD method.

CorrSino method. Considering the similarity between (2b) and (13), where the latter is expected to hold when the assumptions of small motion along the z -direction are met, it could be argued that WeightedGradSino should be similar to the RPC to a certain extent (see figure 2). WeightedGradSino could therefore be expected to be representative of the biggest changes that occur in the dynamic sinogram because of respiration, while additionally having a known relationship with the direction of axial motion. For this reason, to check if the sign of the RPC signal corresponds to axial motion in the desired direction, the RPC itself can be directly compared to the WeightedGradSino, evaluating the correlation between the two sinograms:




Figure 4. (a) and (b) are the RPC and WeightedGradSino as in figure 2; (c) is the opposite of (a) and its grey scale is $[-0.05, 0.03]$. It can be seen that (a) is much more similar to (b) (correlation equal to $+0.57$) than (c) is (correlation equal to -0.57).

$$\text{corr}(\text{RPC}, \text{WeightedGradSino}) = \frac{\sum_{r,z,\phi} \text{RPC}(r, z, \phi) \text{WeightedGradSino}(r, z, \phi)}{\sqrt{\sum_{r,z,\phi} \text{RPC}(r, z, \phi)^2 \text{WeightedGradSino}(r, z, \phi)^2}} \quad (15)$$

The RPC and WeightedGradSino undergo a masking process prior to the evaluation of the correlation, in order to consider only non-zero bins. The absolute value of the correlation relates to the degree of agreement between the RPC and the WeightedGradSino sinogram, taking higher values in the case the gradient sinogram adequately represents the motion detected by the respiratory PC. If the correlation is negative the sign of the RPC signal has to be reversed, as in the *CorrWeights* method.

In figure 4 an example is shown of the appearance of the RPC when it has the wrong sign: the features and intensity patterns look similar in WeightedGradSino (4(b)) and in RPC (4(a)) and their correlation, from equation 15, is equal to 0.57, whereas 4(c) is the negative version of 4(a) ($-$ RPC) and would be the result of the PCA related to the ‘wrong’ direction of motion. The correlation between 4(b) and 4(c) would result in -0.57 and the method would in this case reverse the corresponding respiratory signal w .

This method is specific to PCA, as it relies on the comparison of the RPC to the WeightedGradSino.

2.3. Data

2.3.1. Phantom simulations. As an initial test of the performance of the proposed methods, they were applied on simulated data produced with the XCAT phantom Segars *et al* (2010). In order to make the simulation realistic, the RPM respiratory trace of a normal-breathing patient was used as a template for the phantom motion. Diaphragm maximum displacement was set to 2 cm and anterior-posterior maximum displacement to 1.2 cm (default in the XCAT parameters). The duration was set to 50 s and the temporal sampling to 500 ms. Activity and attenuation images were produced. The activity images were projected through the use of STIR utilities (Thielemans *et al* 2012), simulating 3D projections in a GE Discovery STE PET/CT scanner (Teräs *et al* 2007), and their projections were attenuated using attenuation coefficients obtained from the attenuation images. Poisson noise was added to the projections, and no random nor scattered events were simulated (total number of counts per frame $\sim 1.2 \times 10^5$, which is similar to half the counts in clinical data, considering there are no random nor scatter events).

2.3.2. Patient data. To test the performance of the four sign-determination methods we used FDG data of oncology patients. Patient datasets were acquired in 3D listmode on GE Discovery STE (21 datasets) and GE Discovery 690 (Bettinardi *et al* 2011) (16 datasets)

PET/CT scanners using activity levels for routine clinical protocols. The acquisition times ranged from 240 s to 720 s per bed position and only the lower chest bed positions were selected for this study. The acquisitions were monitored by the Varian Real-time Position Management™ (RPM) device, whose signal was used as the comparative standard for the evaluation process. For each patient the selected bed acquisition was also subdivided to independent shorter time intervals of 50, 100, 200 and 300 s in order to assess the performance of the methods on smaller amounts of data (i.e. for a 360 s acquisition, we selected 7 intervals of 50 s, 3 of 100 s, 1 of 200 s and 1 of 300 s).

2.3.3. Data processing. The 3D sinograms provided by the scanner (and those obtained from the XCAT simulation) consist of 553 2D sinograms of 329×280 in r, ϕ for Discovery STE and 553 2D sinograms of 381×288 in r, ϕ for Discovery 690, the axial field-of-view is for both scanners 157 mm and the listmode data were unlisted into sinograms with time frames of 500 ms. Before the application of PCA and consequently the sinogram-based sign determination methods, the sinograms underwent SSRB and radial and angular mashing: the number of summed angles was 40 for the Discovery STE and 32 for Discovery 690 and the number of summed radial bins was 4 in both cases, so that the final data format was $82 \times 7 \times 47$ (radial positions \times angles \times transaxial planes) for Discovery STE datasets and $95 \times 9 \times 47$ for Discovery 690 datasets.

For the registration-based method, the procedure of section 2.2.1 was used and images were reconstructed with voxel dimensions of $5.99 \times 5.99 \times 3.27 \text{ mm}^3$ for Discovery STE datasets and $5.33 \times 5.33 \times 3.27 \text{ mm}^3$ for Discovery 690 datasets.

2.4. Evaluation

The evaluation followed the following steps:

- (i) for the selected interval of the patient listmode file and for the projections of the simulation data, the PCA-method described in section 2.1 was used to produce the RPC and its respiratory signal, with a time frame duration of 500 ms;
- (ii) the four methods (*CorrSino*, *CorrWeights*, *MIP* and *SUM*) were then applied on the selected intervals and the obtained RPC signal was corrected according to the method's response;
- (iii) the correlation was evaluated between each of the obtained respiratory signals and the corresponding RPM signal (downsampled to the same time resolution as the RPC signal). When the correlation is positive, the respiratory signal is correct, when the correlation is negative the respiratory signal is opposite to the RPM, therefore the method failed in correcting the sign.

3. Results

The application of the presented sign determination methods on XCAT data resulted in each case in the correct determination of the signal, and the correlation values obtained from the application of *CorrWeights* and *CorrSino* were 0.81 and 0.3 respectively (see (14) and (15)).

For the patient data, examples of the appearance of the RPC and *WeightedGradSino* sinograms, and of the RPC signal and $u(t)$ have been shown in figures 2 and 3. The application of the methods resulted in the failure rates reported in table 1, that includes all the acquired patients and all the different time duration intervals.

Table 1. Percentage of failures that were obtained for each group of intervals (that vary in time duration). The numbers in brackets refer to the amount of evaluated intervals for each group.

| Method | Duration (n.intervals) | | | | All (431) |
|-------------|------------------------|-------------|------------|------------|-----------|
| | 50 s (242) | 100 s (113) | 200 s (44) | 300 s (32) | |
| CorrSino | 5.8 | 4.4 | 6.8 | 9.4 | 5.8 |
| CorrWeights | 4.9 | 4.4 | 6.8 | 9.4 | 5.3 |
| MIP | 21.9 | 15.0 | 9.0 | 12.5 | 18.0 |
| SUM | 14.9 | 12.4 | 6.8 | 6.3 | 12.8 |

Table 1 shows that methods *CorrSino* and *CorrWeights* have a more stable behavior with respect to the duration of the time intervals, compared to the registration-based methods *MIP* and *SUM* that show much higher failure rates in the 50 s and 100 s cases. The reduced amount of data that is utilized in such cases affects the quality of the reconstructed gated images, therefore the registration process is more likely to fail.

The DD output depends on the actual presence of respiratory motion and on the assumption that it represents one of the main causes of variation in the acquired data. The quality of the provided respiratory trace, considered as degree of agreement with the RPM signal, is therefore not predictable and can vary between different patient acquisitions and also between different time intervals of the same study. The same issues exist for the sign determination methods, as they too rely on the same assumptions as the DD method, and their outcome might vary with respect to different acquisitions and different respiratory signals.


In order to better understand the trend of the results with respect to the ‘respiratory-likeness’ of the PCA signal $w(t)$, the failure rates have been analysed with respect to the correlation between $w(t)$ and the RPM signal. The majority of the obtained signals $w(t)$ (401 cases, equal to 93%) are highly similar to the RPM trace, showing correlations ranging from 0.75 to a maximum of 0.99, while in the remaining 7% of the intervals (30 cases) the quality of the PCA signal is sub-optimal, with 21 intervals with correlation included in $[0.5, 0.75]$, 3 cases in $[0.25, 0.5]$ and 6 cases in $[0, 0.25]$. The failure rates corresponding to the 401 cases with DD-RPM correlation values above 0.75 and the 30 cases with DD-RPM correlation values below 0.75 are shown in table 2.

To compare in more detail the performance and the reliability of *CorrSino* and *CorrWeights*, we examined the numerical results for each patient. In figure 5 the results obtained by the application of both methods are displayed per patient (for the Discovery STE datasets): each point in the plot corresponds to one evaluated interval and its absolute value corresponds to the absolute value of the methods’ outcome (see (14) for *CorrWeights* and (15) for *CorrSino*). Each point is assigned a positive or negative sign, depending on whether the method succeeded or failed in detecting the correct direction of motion in that specific interval. For each patient, the displayed values correspond to the evaluated intervals ordered from the shortest to the longest in duration, and intervals of the same duration are ordered chronologically (i.e. $[250, 300]$, $[300, 350]$, $[350, 400]$,...).

Figure 5 shows that the sinogram-based methods *CorrWeights* and *CorrSino* fail in the same cases (apart from two 50s intervals where *CorrSino* fails and *CorrWeights* succeeds). This could be expected considering that both methods rely on similar assumptions and therefore are potentially vulnerable to cases where the assumptions are not met, e.g. when the biggest changes in the acquired data are caused by bulk motion. Nevertheless it is clearly visible that *CorrWeights* values are considerably higher than those provided by *CorrSino*, therefore suggesting *CorrWeights* might be considered more reliable. For Patient 1, Patient 4

Table 2. Failure rates (%) over all intervals with respect to the correlation with the RPM signal.

| Method | Correlation range (n.intervals) | |
|-------------|---------------------------------|--------------------|
| | [0, 0.75] (30) | [0.75, 1.00] (401) |
| CorrSino | 46.0 | 2.7 |
| CorrWeights | 36.6 | 2.9 |
| MIP | 36.7 | 16.6 |
| SUM | 46.7 | 10.1 |

**Figure 5.** Correlation values for *CorrWeights*, see (14), and *CorrSino*, see (15), for all interval durations, for Discovery STE datasets. When values are negative, the method failed in detecting the correct direction of the RPC signal.

and Patient 5 both methods failed in several time intervals, possibly due to the patient moving during the acquisition. The plot also suggests that the results can vary considerably from patient to patient. This can be expected considering that the breathing patterns can be noticeably different between patients and also that interference due to other types of motion can be quite variable.

4. Discussion

When using the simulation data, all the presented sign-determination methods correctly detected the direction of axial motion, even in the presence of anterior-posterior motion. This suggests that the rationale for the two sinogram-based methods is reasonable. Moreover, the motion in the XCAT simulation data is non-rigid and relatively large (2 cm diaphragm movement and 1.2 cm chest expansion), therefore we have shown that the methods, that suppose motion is in axial direction, small and rigid, perform well even in this instance. Nevertheless, the structure and motion of the XCAT phantom is relatively simple and the evaluation of the methods on the patient data is therefore essential.

When applying the methods on the available patients datasets, some failures occurred. *CorrSino* and *CorrWeights* proved to be generally more accurate than *SUM* and *MIP* in detecting the correct sign of the respiratory signal, providing an overall failure rate of 5.8% and 5.3% on the 431 evaluated intervals, compared to 18.0% and 12.8%. The former two methods

also have the advantage of avoiding the reconstruction step making use of the raw PET data only. Table 1 also shows that the sinogram-based methods are more stable compared to the registration-based ones with respect to the amount of data that is used.

In the majority of the cases in our cohort, which includes the lower chest bed position of FDG scans of oncology patients, the RPC signal was highly correlated with the external RPM signal (93% had correlation higher than 0.75).

The failure rates displayed in table 2 show that all sign-determination methods are more likely to fail as the correlation between RPC signal and RPM decreases, i.e. when the RPC signal is not adequately representing the external motion of the chest, although there is insufficient data (only 30 of the total 431 evaluated intervals) to draw significant conclusions about which method is best in these circumstances. For the cases with correlation higher than 0.75, the sinogram-based methods are clearly performing better compared to the registration-based methods, both failing in less than 3% of the cases.

The more detailed analysis carried out on *CorrWeights* and *CorrSino* shows that the correlation between the WeightedGradSino weight and the RPC signal is generally higher than the correlation between their corresponding sinograms (see figure 5). This indicates that *CorrWeights* could be a more reliable technique than *CorrSino*. Furthermore, the majority of the failures of *CorrWeights* arise when its result has an absolute value below 0.2, suggesting that a threshold could be defined to only use the method when it produces high values and is therefore highly likely to succeed. The determination of such a threshold will be the subject of future investigation.

In addition, it is worth noting that the *CorrWeights* method can be applied to fix the sign of the signal obtained from any DD method as it does not rely specifically on the use of PCA.

In general, failures in determining the sign of the DD respiratory motion with the presented techniques could be due to several reasons. Low contrast and noise in the data hinder the extraction of a good DD respiratory signal. Minimal motion along the z -direction contradicts the assumption at the core of the investigated methods. Also other types of patient movements during the acquisition might contaminate the RPC signal. Nevertheless figure 5 clearly shows that for the sinogram-based methods the strongest variability in the outcome is found between different patients and also that the duration of the interval of data taken into account does not have a noticeable impact.

In our implementation all of the data in the sinograms are used both for the extraction of the respiratory signal and for the determination of its sign. Selecting regions of interest (ROI) known to include respiratory motion might potentially improve both the quality of the signal and the performance of the methods. However, defining ROIs would require reconstruction of the data, as anatomical features are difficult to detect in sinogram space, and would therefore require additional processing time and external input from trained personnel. As the goal of our work is to provide a methodology which is exclusively dependent on the acquired PET raw data, we have not put this into practice.

The evaluation on patient data of the proposed methods has been performed via comparison with the RPM signal. This has been shown to potentially exhibit a time lag when compared to the internal organ motion (Ionascu *et al* 2007), while DD methods are expected to provide signals that are closely representative of the latter. However, as this time lag is expected to be small (below 0.2 s in Ionascu *et al* (2007)), we do not expect it to undermine the conclusions of this work, as it would only decrease the absolute value of the evaluated correlation between RPM and RPC signal but not its sign. Nevertheless a limitation of this study is that it relies on the RPM signal correctly representing the respiratory motion of the patient. A poor correlation between RPC signal and RPM could therefore be due to a deficient RPM rather than

inaccurate RPC signal. However, the RPM signal is currently utilized in clinical practice and is usually considered to be a reliable method.

As for potential irregularity of the breathing in time, we expect the methods not to be affected by it, as long as the respiratory motion is similar throughout the acquisition. What could potentially change is the value of the ratio shown in figure 1 for an example of PCs: if the breathing is irregular the peak in the frequency band would be expected to decrease, and so would the resulting ratio value.

Future work will include a further analysis of the failures of *CorrSino* and *CorrWeights* and, in addition, the design of an informative metric for the goodness of the obtained signals, in order to provide confidence in the RPC signal in the absence of an external device (in our work RPC signals were compared to the RPM signal). The presented work is limited to the use of the PCA and the implemented methods were applied only to static studies in FDG oncology patients, therefore future developments and applications could include the extension to other DD algorithms and also to different types of acquisitions.

5. Conclusion

We presented two new methods, *CorrWeights* and *CorrSino*, for the determination of the direction of the data-driven respiratory signal obtained with PCA. The new methods provided lower failure rates compared to the previously published registration-based methods. When the extracted respiratory signal is highly correlated with the RPM they fail in less than 3% of the cases, compared to more than 10%. Among the two proposed methods, *CorrWeights* is not specifically related to PCA, therefore it could be utilized to determine the correct direction of the motion of respiratory signals obtained with any other DD method. Considering its simple implementation, it could easily be applied together with any DD method in clinical practice.

Acknowledgments

This project is supported by EPSRC Industrial CASE studentship 13220093, co-sponsored by GE Healthcare, and by the National Institute for Health Research, University College London Hospitals Biomedical Research Centre.

We wish to thank Valentino Bettinardi (Scientific Institute Ospedale San Raffaele, Milan, Italy), Michael O'Doherty (Guy's and St Thomas' Hospital, London) and François Benard (BCCA, Vancouver, CA) for the anonymized data, and Paul Schleyer for advice and collaboration.

Appendix

The detailed steps to obtain (11) are reported here. The Freeman–Tukey transformation is approximated as follows:

$$d(z, t)_{\text{FT}} \simeq 2\sqrt{d(z, t)} = 2\sqrt{n(z, t)s(z)}$$

and $\sqrt{n(z, t)}$ can be written using (10):

$$\sqrt{n(z, t)} \simeq \sqrt{\bar{n}(z) + (\delta z(t) - \bar{\delta z})\text{GradSino}(z)}.$$

Assuming that the motion is small, the variation over time in the dynamic sinogram will be small as well and consequently the Taylor expansion of $\sqrt{x + \epsilon}$ can be utilised, yielding:

$$\sqrt{n(z, t)} \simeq \sqrt{\bar{n}(z)} + \frac{1}{2} \frac{(\delta z(t) - \bar{\delta z}) \text{GradSino}(z)}{\sqrt{\bar{n}(z)}}.$$

We are ultimately interested in the result of $d(z, t)_{\text{FT}} - \bar{d}(z)_{\text{FT}}$, and using the equivalence $d(z, t)_{\text{FT}} \simeq 2\sqrt{n(z, t)s(z)}$ and applying the time average to the latter, we obtain:

$$d(z, t)_{\text{FT}} - \bar{d}(z)_{\text{FT}} \simeq (\delta z(t) - \bar{\delta z}) \frac{\text{GradSino}(z)\sqrt{s(z)}}{\sqrt{\bar{n}(z)}}.$$

References

- Bailey D L 2005 *Positron Emission Tomography* (Berlin: Springer) pp 41–62
- Bettinardi V, Presotto L, Rapisarda E, Picchio M, Gianolli L and Gilardi M 2011 Physical performance of the new hybrid PET/CT discovery-690 *Med. Phys.* **38** 5394–411
- Boucher L, Rodrigue S, Lecomte R and Bénard F 2004 Respiratory gating for 3-dimensional PET of the thorax: feasibility and initial results *J. Nucl. Med.* **45** 214–9
- Bundschuh R A, Martínez-Moeller A, Essler M, Martínez M J, Nekolla S G, Ziegler S I and Schwaiger M 2007 Postacquisition detection of tumor motion in the lung and upper abdomen using list-mode PET data: a feasibility study *J. Nucl. Med.* **48** 758–63
- Büther F, Dawood M, Stegger L, Wübbeling F, Schäfers M, Schober O and Schäfers K P 2009 List mode-driven cardiac and respiratory gating in PET *J. Nucl. Med.* **50** 674–81
- Daube-Witherspoon M E and Muehllehner G 1987 Treatment of axial data in three-dimensional PET *J. Nucl. Med.* **28** 1717–24
- Erdi Y E et al 2004 The CT motion quantitation of lung lesions and its impact on PET-measured SUVs *J. Nucl. Med.* **45** 1287–92
- Fayad H, Pan T, Clement J F and Visvikis D 2011 Technical note: correlation of respiratory motion between external patient surface and internal anatomical landmarks *Med. Phys.* **38** 3157–64
- Fin L, Daouk J, Morvan J, Bailly P, El Esper I, Saidi L and Meyer M E 2008 Initial clinical results for breath-hold CT-based processing of respiratory-gated pet acquisitions *Eur. J. Nucl. Med. Mol. Imaging* **35** 1971–80
- Freeman M F and Tukey J W 1950 Transformations related to the angular and the square root *Ann. Math. Stat.* **21** 607–11
- Guivarc'h O, Turzo A, Visvikis D and Bizais Y 2004 Synchronization of pulmonary scintigraphy by respiratory flow and by impedance plethysmography *Proc. SPIE* **5370** 1166–75
- He J, Keefe G J, Gong S J, Jones G, Saunder T, Scott A M and Geso M 2008 A novel method for respiratory motion gated with geometric sensitivity of the scanner in 3D PET *IEEE Trans. Nucl. Sci.* **55** 2557–65
- Ionascu D, Jiang S B, Nishioka S, Shirato H and Berbeco R I 2007 Internal-external correlation investigations of respiratory induced motion of lung tumors *Med. Phys.* **34** 3893–903
- Kesner A L and Kuntner C 2010 A new fast and fully automated software based algorithm for extracting respiratory signal from raPET data and its comparison to other methods *Med. Phys.* **37** 5550–9
- Kesner A L, Schleyer P J, Büther F, Walter M A, Schäfers K P and Koo P J 2014 On transcending the impasse of respiratory motion correction applications in routine clinical imaging—a consideration of a fully automated data driven motion control framework *Eur. J. Nucl. Med. Mol. Imaging Phys.* **1** 8
- Liu C, Pierce L A II, Alessio A M and Kinahan P E 2009 The impact of respiratory motion on tumor quantification and delineation in static pet/ct imaging *Phys. Med. Biol.* **54** 7345
- Nehmeh S A, Erdi Y E, Ling C C, Rosenzweig K E, Schoder H, Larson S M, Macapinlac H A, Squire O D and Humm J L 2002a Effect of respiratory gating on quantifying PET images of lung cancer *J. Nucl. Med.* **43** 876–81
- Nehmeh S A, Erdi Y E, Rosenzweig K E, Schoder H, Larson S M, Squire O D and Humm J L 2003 Reduction of respiratory motion artifacts in PET imaging of lung cancer by respiratory correlated dynamic PET: methodology and comparison with respiratory gated PET *J. Nucl. Med.* **44** 1644–8

- Nehmeh S *et al* 2002b Effect of respiratory gating on reducing lung motion artifacts in PET imaging of lung cancer *Med. Phys.* **29** 366–71
- Otani Y, Fukuda I, Tsukamoto N, Kumazaki Y, Sekine H, Imabayashi E, Kawaguchi O, Nose T, Teshima T and Dokiya T 2010 A comparison of the respiratory signals acquired by different respiratory monitoring systems used in respiratory gated radiotherapy *Med. Phys.* **37** 6178–86
- Ozhasoglu C and Murphy M J 2002 Issues in respiratory motion compensation during external-beam radiotherapy *Int. J. Radiat. Oncol.* **52** 1389–99
- Pearson K 1901 On lines and planes of closest fit to system of points in space *Phil. Mag.* **2** 559–72
- Pepin A, Daouk J, Bailly P, Hapdey S and Meyer M E 2014 Management of respiratory motion in PET/computed tomography: the state of the art *Nucl. Med. Commun.* **35** 113
- Rahmim A, Rousset O and Zaidi H 2007 Strategies for motion tracking and correction in PET *PET Clin.* **2** 251–66
- Riedel M, Navab N and Moller A 2006 Respiratory motion estimation: tests and comparison of different sensors *Interdisciplinary Project Technische Universität München Fakultät für Informatik, Germany*
- Schleyer P J, O'Doherty M J and Marsden P K 2011 Extension of a data-driven gating technique to 3D, whole body pet studies *Phys. Med. Biol.* **56** 3953
- Schleyer P J, O'Doherty M J, Barrington S F and Marsden P K 2009 Retrospective data-driven respiratory gating for PET/CT *Phys. Med. Biol.* **54** 1935
- Segars W P, Sturgeon G, Mendonca S, Grimes J and Tsui B M W 2010 4D XCAT phantom for multimodality imaging research *Med. Phys.* **37** 4902–15
- Teräs M, Tolvanen T, Johansson J, Williams J and Knuuti J 2007 Performance of the new generation of whole-body PET/CT scanners: discovery STE and discovery VCT *Eur. J. Nucl. Med. Mol. Imaging* **34** 1683–92
- Thielemans K, Rathore S, Engbrant F and Razifar P 2011 Device-less gating for PET/CT using PCA *IEEE Nuclear Science Symp. and Medical Imaging Conf. Record* pp 3904–10
- Thielemans K *et al* 2014 Data-driven dual-gating for cardiac PET *IEEE Nuclear Science Symp. and Medical Imaging Conf. Record* (IEEE) pp 1–4
- Thielemans K, Tsoumpas C, Mustafovic S, Beisel T, Aguiar P, Dikaivos N and Jacobson M W 2012 STIR: software for tomographic image reconstruction release 2 *Phys. Med. Biol.* **57** 867
- Visvikis D, Barret O, Fryer T, Turzo A, Lamare F, Rest L, Cheze C and Bizais Y 2003 A posteriori respiratory motion gating of dynamic PET images *IEEE Nuclear Science Symp. and Medical Imaging Conf. Record* vol 5 pp 3276–80
- Visvikis D, Lamare F, Bruyant P, Boussion N and Le Rest C C 2006 Respiratory motion in positron emission tomography for oncology applications: problems and solutions *Nucl. Instrum. Methods A* **569** 453–7
- Wachinger C, Yigitsoy M, Rijkhorst E J and Navab N 2012 Manifold learning for image-based breathing gating in ultrasound and MRI *Med. Image Anal.* **16** 806–18